

Inorganic Standards



3 Year
Minimum Shelf Life on
Single Element
ICP, ICP/MS and AA
Standards



AccuTrace™ Documentation includes:

- Traceability to NIST SRM by Wet Chemical / Gravimetric Assay
- Traceability to NIST SRM by Instrumental Analysis
- Reference to NIST Traceability during product preparation

Inorganic products containing acid generally
require a hazardous shipping fee.

Inorganic products in water generally do not.



Single Element	348-364
ICP	348-351
ICP/MS	352-353
AA	354-356
Matrix Modifiers	357
Ion Chrom	358-361
Wet Chemicals	362-363
TPH, Oil and Grease (EPA Methods)	364
ICP Multi-Element Standards	366-392
Quality Control and Second Source QC	366
Instrument Check and Screening	367
SDWA Drinking Water	368
Groundwater & Wastewater	369
TCLP	369
MISA Test Group 29 Analysis	370
Contract Laboratory Program (CLP)	371-373
Calibration Check Standards	371
Verification and Spiking	372
Interference Check	373
Detection Limit	373
SW-846, EPA Method 200.7	374-377
Calibration Standards	374
Instrument Performance Check Standards	375
LPCS and LFSS	375
Instrument Fortifying Standards	376
Spiking Standards	377
Interference Check Standards	377
EPA Method 6010	378-379
Merck Multi-Element	381-382
Alternate Source (Agilent/Varian, Perkin Elmer, Jobin Yvon, Teledyne)	383-391
EU Formulation	392
ASTM Methods D5184, D5600	392
ICP/MS Multi-Element Standards	393-396
Calibration Standards and Blanks	393
Tuning Solutions	394
Interference Check	394
Memory Check	394
Spiking Standards for Water & Soil	395
Quality Control	395
Internal Standards	395
EPA Method 200.8	396
EPA Method 6020	396
Organometallic Standards	397-404
Single Element Wear Metals	397
Metals Additives	397
Multi-Element Wear Metals	398
Sulfur-Free Single Element	399
Sulfur-Free Multi-Element	400
Sulfur and Metals in Oil	401
Vanadium and Nickel	401
Lube Oil Standards	402-404
Inorganic ASTM Methods	405
Solid Matrix Standards	406-408

Documentation

AccuTrace™



Density included for easy conversion to weight/weight applications.

Most Single element standards have a minimum 3 Year expiration period.

Accreditations

- ☒ Included on ISO/IEC 17025 Scope of Accreditation
- ☒ Included on ISO Guide 34 Scope of Accreditation

Impurity Scan for 68 elements in final solution.

Elements in µg/mL											
Ag	nd<0.02	Cr	nd<0.02	Ho	nd<0.02	Nd	nd<0.02	S	N/A	Ti	nd<0.02
Al	nd<0.02	Cs	N/A	In	N/A	Ni	nd<0.02	Sb	nd<0.2	Tl	nd<0.2
As	nd<0.2	Cu	*	Ir	N/A	Os	N/A	Se	nd<0.02	Tm	nd<0.02
Au	nd<0.02	Dy	nd<0.02	K	nd<0.2	P	N/A	Si	nd<0.2	U	nd<0.2
B	nd<0.2	Er	nd<0.02	La	nd<0.02	Pb	N/A	Sm	nd<0.2	V	nd<0.02
Ba	nd<0.02	Eu	nd<0.02	Li	nd<0.02	Pd	nd<0.2	Sn	N/A	W	nd<0.2
Be	nd<0.02	Fe	0.02	Lu	nd<0.02	Pr	nd<0.2	Sr	nd<0.02	Y	0.02
Bi	N/A	Ga	nd<0.02	Mg	nd<0.02	Pt	nd<0.2	Ta	nd<0.02	Yb	nd<0.02
Ca	0.06	Gd	nd<0.02	Mn	nd<0.02	Rb	N/A	Tb	nd<0.02	Zn	nd<0.02
Cd	nd<0.02	Ge	nd<0.2	Mo	nd<0.02	Re	nd<0.2	Te	nd<0.2	Zr	nd<0.02
Ce	nd<0.2	Hf	nd<0.02	Na	0.04	Rh	nd<0.2				
Co	nd<0.02	Hg	nd<0.2	Nb	nd<0.2	Ru	nd<0.02	Th	nd<0.02		

Uncertainty reported for statistical confidence.

Concentration verified by two independent methods for added assurance.

Highest purity starting materials & matrices used.

The gravimetric uncertainty for this product is $\pm 0.24\%$. The CRM uncertainty is $\pm 5\%$. See reverse side for details.
This solution was assayed titrimetrically, using EDTA which was standardized against NIST SRM #928 (lead nitrate.)

The concentration of the final solution was verified by ICP against NIST SRM# 3114

RESULTS: This solution standard was certified for accuracy of major elemental constituency via methodology traceable to primary or well characterized secondary standards. All trace level elements and impurities were determined via plasma emission spectroscopy on the concentrate.

This standard was prepared gravimetrically to contain the elemental concentrations shown above. Balances, used in the preparation, are calibrated regularly using NIST-traceable weights. All glassware used in preparation is Class A.

We use the highest purity raw materials available, including high purity acids, ASTM type I 18 megohm deionized water, and typically 99.999%+ starting materials to minimize impurity levels in the final solution. All bottles are acid leached and then triple rinsed with deionized water prior to use.

Use good laboratory procedure when diluting this product. Shake bottle prior to use and do not pipette directly out of the bottle. Use only cleaned Class A volumetric glassware.

We certify the accuracy of this standard to be $\pm 0.5\%$ of the stated value until the expiration date listed above, provided it is kept tightly capped and stored under normal laboratory conditions.

For use in routine laboratory analysis.

AccuStandard is accredited to ISO Guide 34, ISO/IEC 17025 and certified to ISO 9001

L. Snyder
Lydia Snyder
Inorganic QC Supervisor

OR-ORG/IND-001
Rev. 7/11

Standards prepared under the guidelines of ISO Guide 34, ISO 17025 and ISO 9001 Quality System.

Signature approval indicating product integrity to specifications.


Quality Control

AccuStandard's Inorganic Department combines the best of state of the art instrumentation with classical stoichiometric analysis, to bring you the most reliable Standards available.

Why choose AccuStandard for your Inorganic Analytical Needs?

- Standards are our only product.
- Thousands of Inorganic Standards - the widest selection available.
 - ICP
 - ICP/MS
 - AA
 - IC
 - Wet Chemical
 - Wear Metals
- No across the board price increase in 25 years
- 3-year minimum Shelf Life on single ICP, AA and ICP/MS Standards
- Custom Formulations
- Small Batch Sizes for Fresher Lots
- Technical Service from people with Real World Laboratory Experience
- 24/7 Online Access to MSDSs, COAs, Literature and Product Support
- Accredited to ISO Guide 34, ISO 17025 and certified to ISO 9001
- On-going stability program to ensure product integrity

State of the Art Instrumentation




Perkin Elmer Optima 5300 Dual View ICP/OES



Perkin Elmer Elan 9000 ICP/MS

Quality System

ISO 9001 • ISO/IEC 17025 • ISO Guide 34





ICP

Single Element

- Traceable to NIST Reference Materials
- Formulated from Ultra High Purity Starting Materials and Acids
- 18 Megohm de-ionized Water

- Concentration verified by Wet Chemical and Instrumental Analysis
- Packaged in specially prepared Acid leached bottles
- Actual Lot Analysis provided on label and C of A

Single Element ICP

NoHaz 20 mL Size

Element Starting Material Matrix	Unit	1000 µg/mL Cat. No.	10,000 µg/mL Cat. No.	20 mL (NoHaz) 10,000 µg/mL Cat. No.
Aluminum	50 mL	-----	--	ICP-01N-10X-0.5
Al(NO ₃) ₃ • 9H ₂ O	100 mL	ICP-01N-1	ICP-01N-10X-1	ICP-01N-10X-20ML
2-5% Nitric acid	500 mL	ICP-01N-5	ICP-01N-10X-5	
Antimony	50 mL	-----	--	ICP-02N-10X-0.5
Sb Dilute HNO ₃ tr	100 mL	ICP-02N-1	ICP-02N-10X-1	ICP-02N-10X-20ML
Tartaric acid	500 mL	ICP-02N-5	ICP-02N-10X-5	
Arsenic	50 mL	-----	--	ICP-03N-10X-0.5
As	100 mL	ICP-03N-1	ICP-03N-10X-1	ICP-03N-10X-20ML
2-5% Nitric acid	500 mL	ICP-03N-5	ICP-03N-10X-5	
Barium	50 mL	-----	--	ICP-04N-10X-0.5
Ba(NO ₃) ₂	100 mL	ICP-04N-1	ICP-04N-10X-1	ICP-04N-10X-20ML
2-5% Nitric acid	500 mL	ICP-04N-5	ICP-04N-10X-5	
Beryllium	50 mL	-----	--	ICP-05N-10X-0.5
BeO(C ₂ H ₃ O ₂) ₆	100 mL	ICP-05N-1	ICP-05N-10X-1	ICP-05N-10X-20ML
2-5% Nitric acid	500 mL	ICP-05N-5	ICP-05N-10X-5	
Bismuth	50 mL	-----	--	ICP-06N-10X-0.5
Bi	100 mL	ICP-06N-1	ICP-06N-10X-1	ICP-06N-10X-20ML
2-10% Nitric acid	500 mL	ICP-06N-5	ICP-06N-10X-5	
Boron	50 mL	-----	--	ICP-07W-10X-0.5 ▼
H ₃ BO ₃	100 mL	ICP-07W-1 ▼	ICP-07W-10X-1 ▼	ICP-07W-10X-20ML
Water tr. NH ₄ OH	500 mL	ICP-07W-5 ▼	ICP-07W-10X-5 ▼	
Cadmium	50 mL	-----	--	ICP-08N-10X-0.5
Cd	100 mL	ICP-08N-1	ICP-08N-10X-1	ICP-08N-10X-20ML
2-5% Nitric acid	500 mL	ICP-08N-5	ICP-08N-10X-5	
Calcium	50 mL	-----	--	ICP-09N-10X-0.5
CaCO ₃	100 mL	ICP-09N-1	ICP-09N-10X-1	ICP-09N-10X-20ML
2-5% Nitric acid	500 mL	ICP-09N-5	ICP-09N-10X-5	
Cerium	50 mL	-----	--	ICP-11N-10X-0.5
Ce(NO ₃) ₃	100 mL	ICP-11N-1	ICP-11N-10X-1	ICP-11N-10X-20ML
2-5% Nitric acid	500 mL	ICP-11N-5	ICP-11N-10X-5	
Cesium	50 mL	-----	--	ICP-12N-10X-0.5
CsNO ₃	100 mL	ICP-12N-1	ICP-12N-10X-1	ICP-12N-10X-20ML
2-5% Nitric acid	500 mL	ICP-12N-5	ICP-12N-10X-5	
Chromium reduced to (+3) state	50 mL	-----	--	ICP-13N-10X-0.5
(NH ₄) ₂ Cr ₂ O ₇	100 mL	ICP-13N-1	ICP-13N-10X-1	ICP-13N-10X-20ML
2-5% Nitric acid	500 mL	ICP-13N-5	ICP-13N-10X-5	
Cobalt	50 mL	-----	--	ICP-14N-10X-0.5
Co	100 mL	ICP-14N-1	ICP-14N-10X-1	ICP-14N-10X-20ML
2-5% Nitric acid	500 mL	ICP-14N-5	ICP-14N-10X-5	
Copper	50 mL	-----	--	ICP-15N-10X-0.5
Cu	100 mL	ICP-15N-1	ICP-15N-10X-1	ICP-15N-10X-20ML
2-5% Nitric acid	500 mL	ICP-15N-5	ICP-15N-10X-5	
Dysprosium	50 mL	-----	--	ICP-16N-10X-0.5
Dy ₂ O ₃	100 mL	ICP-16N-1	ICP-16N-10X-1	ICP-16N-10X-20ML
2-5% Nitric acid	500 mL	ICP-16N-5	ICP-16N-10X-5	
Erbium	50 mL	-----	--	ICP-17N-10X-0.5
Er ₂ O ₃	100 mL	ICP-17N-1	ICP-17N-10X-1	ICP-17N-10X-20ML
2-5% Nitric acid	500 mL	ICP-17N-5	ICP-17N-10X-5	
Europium	50 mL	-----	--	ICP-18N-10X-0.5
Eu ₂ O ₃	100 mL	ICP-18N-1	ICP-18N-10X-1	ICP-18N-10X-20ML
2-5% Nitric acid	500 mL	ICP-18N-5	ICP-18N-10X-5	
Gadolinium	50 mL	-----	--	ICP-19N-10X-0.5
Gd ₂ O ₃	100 mL	ICP-19N-1	ICP-19N-10X-1	ICP-19N-10X-20ML
2-5% Nitric acid	500 mL	ICP-19N-5	ICP-19N-10X-5	
Gallium	50 mL	-----	--	ICP-20N-10X-0.5
Ga	100 mL	ICP-20N-1	ICP-20N-10X-1	ICP-20N-10X-20ML
2-5% Nitric acid	500 mL	ICP-20N-5	ICP-20N-10X-5	

▼ Hazardous fee not required.

ICP

Single Element



**3 Year
Minimum Shelf Life on
Single Element ICP
Standards**



Save with NoHaz Option on ICP Single Elements

- No Hazardous Shipping Fees
- Lower Shipping Costs (less weight)
- Yields More - 200 mL from 20 mL concentrate

Includes empty pre-washed, pre-labeled HDPE 250 mL bottle

Single Element ICP				NoHaz 20 mL Size
Element Starting Material Matrix	Unit	1000 µg/mL Cat. No.	10,000 µg/mL Cat. No.	20 mL (NoHaz) 10,000 µg/mL Cat. No.
Germanium	50 mL	-----	--	ICP-21W-10X-0.5 ▼
(NH ₄) ₂ GeF ₆	100 mL	ICP-21W-1 ▼	ICP-21W-10X-1 ▼	ICP-21W-10X-20ML
Water tr. HF	500 mL	ICP-21W-5 ▼	ICP-21W-10X-5 ▼	
Gold	50 mL	-----	--	ICP-22H-10X-0.5
Au	100 mL	ICP-22H-1	ICP-22H-10X-1	ICP-22H-10X-20ML
10% HCl (min.)	500 mL	ICP-22H-5	-----	--
Hafnium	50 mL	-----	--	ICP-23N-10X-0.5
HfO ₂	100 mL	ICP-23N-1	ICP-23N-10X-1	ICP-23N-10X-20ML
2-5% Nitric acid tr. HF	500 mL	ICP-23N-5	ICP-23N-10X-5	
Holmium	50 mL	-----	--	ICP-24N-10X-0.5
Ho ₂ O ₃	100 mL	ICP-24N-1	ICP-24N-10X-1	ICP-24N-10X-20ML
2-5% Nitric acid	500 mL	ICP-24N-5	ICP-24N-10X-5	
Indium	50 mL	-----	--	ICP-25N-10X-0.5
In	100 mL	ICP-25N-1	ICP-25N-10X-1	ICP-25N-10X-20ML
2-5% Nitric acid	500 mL	ICP-25N-5	ICP-25N-10X-5	
Iridium	50 mL	-----	--	ICP-26H-10X-0.5
IrCl ₃ • 3H ₂ O	100 mL	ICP-26H-1	ICP-26H-10X-1	ICP-26H-10X-20ML
10% HCl (min.)	500 mL	ICP-26H-5	-----	--
Iron	50 mL	-----	--	ICP-27N-10X-0.5
Fe	100 mL	ICP-27N-1	ICP-27N-10X-1	ICP-27N-10X-20ML
2-5% Nitric acid	500 mL	ICP-27N-5	ICP-27N-10X-5	
Lanthanum	50 mL	-----	--	ICP-28N-10X-0.5
La ₂ O ₃	100 mL	ICP-28N-1	ICP-28N-10X-1	ICP-28N-10X-20ML
2-5% Nitric acid	500 mL	ICP-28N-5	ICP-28N-10X-5	
Lead	50 mL	-----	--	ICP-29N-10X-0.5
Pb(NO ₃) ₂	100 mL	ICP-29N-1	ICP-29N-10X-1	ICP-29N-10X-20ML
2-5% Nitric acid	500 mL	ICP-29N-5	ICP-29N-10X-5	
Lithium	50 mL	-----	--	ICP-30N-10X-0.5
Li ₂ CO ₃	100 mL	ICP-30N-1	ICP-30N-10X-1	ICP-30N-10X-20ML
2-5% Nitric acid	500 mL	ICP-30N-5	ICP-30N-10X-5	
Lutetium	50 mL	-----	--	ICP-31N-10X-0.5
Lu ₂ O ₃	100 mL	ICP-31N-1	ICP-31N-10X-1	ICP-31N-10X-20ML
2-5% Nitric acid	500 mL	ICP-31N-5	ICP-31N-10X-5	
Magnesium	50 mL	-----	--	ICP-32N-10X-0.5
MgO	100 mL	ICP-32N-1	ICP-32N-10X-1	ICP-32N-10X-20ML
2-5% Nitric acid	500 mL	ICP-32N-5	ICP-32N-10X-5	
Manganese	50 mL	-----	--	ICP-33N-10X-0.5
Mn(C ₂ H ₃ O ₂) ₂	100 mL	ICP-33N-1	ICP-33N-10X-1	ICP-33N-10X-20ML
2-5% Nitric acid	500 mL	ICP-33N-5	ICP-33N-10X-5	
Mercury	50 mL	-----	--	ICP-34N-10X-0.5
Hg	100 mL	ICP-34N-1	ICP-34N-10X-1	ICP-34N-10X-20ML
10% Nitric acid	500 mL	ICP-34N-5	ICP-34N-10X-5	
Molybdenum	50 mL	-----	--	ICP-35W-10X-0.5 ▼
(NH ₄) ₂ MoO ₄	100 mL	ICP-35W-1 ▼	ICP-35W-10X-1 ▼	ICP-35W-10X-20ML
Water tr. NH ₄ OH	500 mL	ICP-35W-5 ▼	ICP-35W-10X-5 ▼	
Neodymium	50 mL	-----	--	ICP-36N-10X-0.5
Nd ₂ O ₃	100 mL	ICP-36N-1	ICP-36N-10X-1	ICP-36N-10X-20ML
2-5% Nitric acid	500 mL	ICP-36N-5	ICP-36N-10X-5	
Nickel	50 mL	-----	--	ICP-37N-10X-0.5
Ni	100 mL	ICP-37N-1	ICP-37N-10X-1	ICP-37N-10X-20ML
2-5% Nitric acid	500 mL	ICP-37N-5	ICP-37N-10X-5	

▼ Hazardous fee not required.



ICP

Single Element

- Traceable to NIST Reference Materials
- Formulated from Ultra High Purity Starting Materials and Acids
- 18 Megohm de-ionized Water

- Concentration verified by Wet Chemical and Instrumental Analysis
- Packaged in specially prepared Acid leached bottles
- Actual Lot Analysis provided on label and COA

Single Element ICP				NoHaz 20 mL Size
Element Starting Material Matrix	Unit	1000 µg/mL Cat. No.	10,000 µg/mL Cat. No.	20 mL (NoHaz) 10,000 µg/mL Cat. No.
Niobium	50 mL	-----	--	ICP-38W-10X-20ML
Nb ₂ O ₅	100 mL	ICP-38W-1 ▼	ICP-38W-10X-1 ▼	
Water tr. HF	500 mL	ICP-38W-5 ▼	ICP-38W-10X-5 ▼	
Palladium	50 mL	-----	--	ICP-40H-10X-20ML
Pd	100 mL	ICP-40H-1	ICP-40H-10X-1	
10% HCl (min.)	500 mL	ICP-40H-5	-----	
Phosphorus	50 mL	-----	--	ICP-41W-10X-20ML
NH ₄ H ₂ PO ₄	100 mL	ICP-41W-1 ▼	ICP-41W-10X-1 ▼	
Water	500 mL	ICP-41W-5 ▼	ICP-41W-10X-5 ▼	
Platinum	50 mL	-----	--	ICP-42H-10X-20ML
Pt	100 mL	ICP-42H-1	ICP-42H-10X-1	
10% HCl (min.)	500 mL	ICP-42H-5	-----	
Potassium	50 mL	-----	--	ICP-43N-10X-20ML
KNO ₃	100 mL	ICP-43N-1	ICP-43N-10X-1	
2-5% Nitric acid	500 mL	ICP-43N-5	ICP-43N-10X-5	
Praseodymium	50 mL	-----	--	ICP-44N-10X-20ML
Pr ₆ O ₁₁	100 mL	ICP-44N-1	ICP-44N-10X-1	
2-5% Nitric acid	500 mL	ICP-44N-5	ICP-44N-10X-5	
Rhenium	50 mL	-----	--	ICP-45W-10X-20ML
Re	100 mL	ICP-45W-1 ▼	ICP-45W-10X-1 ▼	
Water tr. Nitric acid	500 mL	ICP-45W-5 ▼	ICP-45W-10X-5 ▼	
Rhodium	50 mL	-----	--	ICP-46H-10X-20ML
RhCl ₃ • 3H ₂ O	100 mL	ICP-46H-1	ICP-46H-10X-1	
10% HCl (min.)	500 mL	ICP-46H-5	-----	
Rubidium	50 mL	-----	--	ICP-47N-10X-20ML
RbNO ₃	100 mL	ICP-47N-1	ICP-47N-10X-1	
2-5% Nitric acid	500 mL	ICP-47N-5	ICP-47N-10X-5	
Ruthenium	50 mL	-----	--	ICP-48H-10X-20ML
RuCl ₃ • 3H ₂ O	100 mL	ICP-48H-1	ICP-48H-10X-1	
10% HCl	500 mL	ICP-48H-5	-----	
Samarium	50 mL	-----	--	ICP-49N-10X-20ML
Sm ₂ O ₃	100 mL	ICP-49N-1	ICP-49N-10X-1	
2-5% Nitric acid	500 mL	ICP-49N-5	ICP-49N-10X-5	
Scandium	50 mL	-----	--	ICP-50N-10X-20ML
Sc ₂ O ₃	100 mL	ICP-50N-1	ICP-50N-10X-1	
2-5% Nitric acid	500 mL	ICP-50N-5	ICP-50N-10X-5	
Selenium	50 mL	-----	--	ICP-51N-10X-20ML
Se	100 mL	ICP-51N-1	ICP-51N-10X-1	
2-5% Nitric acid	500 mL	ICP-51N-5	ICP-51N-10X-5	
Silicon	50 mL	-----	--	ICP-52W-10X-20ML
(NH ₄) ₂ SiF ₆	100 mL	ICP-52W-1 ▼	ICP-52W-10X-1 ▼	
Water tr. HF	500 mL	ICP-52W-5 ▼	ICP-52W-10X-5 ▼	
Silver	50 mL	-----	--	ICP-53N-10X-20ML
AgNO ₃	100 mL	ICP-53N-1	ICP-53N-10X-1	
2-5% Nitric acid	500 mL	ICP-53N-5	ICP-53N-10X-5	
Sodium	50 mL	-----	--	ICP-54N-10X-20ML
NaNO ₃	100 mL	ICP-54N-1	ICP-54N-10X-1	
2-5% Nitric acid	500 mL	ICP-54N-5	ICP-54N-10X-5	
Strontium	50 mL	-----	--	ICP-55N-10X-20ML
Sr(NO ₃) ₂	100 mL	ICP-55N-1	ICP-55N-10X-1	
2-5% Nitric acid	500 mL	ICP-55N-5	ICP-55N-10X-5	
Sulfur	50 mL	-----	--	ICP-56W-10X-20ML
(NH ₄) ₂ SO ₄	100 mL	ICP-56W-1 ▼	ICP-56W-10X-1 ▼	
Water	500 mL	ICP-56W-5 ▼	ICP-56W-10X-5 ▼	
Tantalum	50 mL	-----	--	ICP-57W-10X-20ML
Ta	100 mL	ICP-57W-1 ▼	ICP-57W-10X-1 ▼	
Water tr. HF	500 mL	ICP-57W-5 ▼	ICP-57W-10X-5 ▼	

▼ Hazardous fee not required.

ICP

Single Element



**3 Year
Minimum Shelf Life on
Single Element ICP
Standards**



Save with NoHaz Option on ICP Single Elements

- No Hazardous Shipping Fees
- Lower Shipping Costs (less weight)
- Yields More - 200 mL from 20 mL concentrate

Includes empty pre-washed, pre-labeled HDPE 250 mL bottle

Single Element ICP				NoHaz 20 mL Size	
Element		1000 µg/mL	10,000 µg/mL	20 mL (NoHaz)	
Starting Material	Unit	Cat. No.	Cat. No.	10,000 µg/mL	Cat. No.
Tellurium	50 mL	-----	--	ICP-58H-10X-0.5	ICP-58H-10X-20ML
Te	100 mL	ICP-58H-1	ICP-58H-10X-1		
20% HCl (min.)	500 mL	ICP-58H-5	ICP-58H-10X-5		
Terbium	50 mL	-----	--	ICP-59N-10X-0.5	ICP-59N-10X-20ML
Tb ₄ O ₇	100 mL	ICP-59N-1	ICP-59N-10X-1		
2-5% Nitric acid	500 mL	ICP-59N-5	ICP-59N-10X-5		
Thallium	50 mL	-----	--	ICP-60N-10X-0.5	ICP-60N-10X-20ML
Tl	100 mL	ICP-60N-1	ICP-60N-10X-1		
2-5% Nitric acid	500 mL	ICP-60N-5	ICP-60N-10X-5		
Thorium		-----	--	-----	--
Th(NO ₃) ₄ • 4H ₂ O	100 mL	ICP-61N-1			
2-5% Nitric acid	500 mL	ICP-61N-5			
Thulium	50 mL	-----	--	ICP-62N-10X-0.5	ICP-62N-10X-20ML
Tm ₂ O ₃	100 mL	ICP-62N-1	ICP-62N-10X-1		
2-5% Nitric acid	500 mL	ICP-62N-5	ICP-62N-10X-5		
Tin	50 mL	-----	--	ICP-63N-10X-0.5	ICP-63N-10X-20ML
Sn	100 mL	ICP-63N-1	ICP-63N-10X-1		
2-5% Nitric acid tr. HF	500 mL	ICP-63N-5	ICP-63N-10X-5		
Titanium	50 mL	-----	--	ICP-64W-10X-0.5 ▼	-----
(NH ₄) ₂ TiF ₆	100 mL	ICP-64W-1 ▼	ICP-64W-10X-1 ▼		
Water tr. HF	500 mL	ICP-64W-5 ▼	ICP-64W-10X-5 ▼		
Tungsten	50 mL	-----	--	ICP-65W-10X-0.5 ▼	ICP-65W-10X-20ML
(NH ₄) ₂ WO ₄	100 mL	ICP-65W-1 ▼	ICP-65W-10X-1 ▼		
Water tr. NH ₄ OH	500 mL	ICP-65W-5 ▼	ICP-65W-10X-5 ▼		
Uranium	----	-----	--	-----	--
U ₃ O ₈	100 mL	ICP-66N-1			
2-5% Nitric acid	500 mL	ICP-66N-5			
Vanadium	50 mL	-----	--	ICP-67N-10X-0.5	ICP-67N-10X-20ML
NH ₄ VO ₃	100 mL	ICP-67N-1	ICP-67N-10X-1		
2-5% Nitric acid	500 mL	ICP-67N-5	ICP-67N-10X-5		
Ytterbium	50 mL	-----	--	ICP-68N-10X-0.5	ICP-68N-10X-20ML
Yb ₂ O ₃	100 mL	ICP-68N-1	ICP-68N-10X-1		
2-5% Nitric acid	500 mL	ICP-68N-5	ICP-68N-10X-5		
Yttrium	50 mL	-----	--	ICP-69N-10X-0.5	ICP-69N-10X-20ML
Y ₂ O ₃	100 mL	ICP-69N-1	ICP-69N-10X-1		
2-5% Nitric acid	500 mL	ICP-69N-5	ICP-69N-10X-5		
Zinc	50 mL	-----	--	ICP-70N-10X-0.5	ICP-70N-10X-20ML
Zn	100 mL	ICP-70N-1	ICP-70N-10X-1		
2-5% Nitric acid	500 mL	ICP-70N-5	ICP-70N-10X-5		
Zirconium	50 mL	-----	--	ICP-71N-10X-0.5	ICP-71N-10X-20ML
ZrO(NO ₃) ₂	100 mL	ICP-71N-1	ICP-71N-10X-1		
2-5% Nitric acid	500 ml	ICP-71N-5	ICP-71N-10X-5		

Calibration and Matrix Blanks

Nitric Acid Blank

CLP-BLN-5 500 mL
CLP-BLN-L-VAP 1L
(2 x 500 mL)

5% HNO₃ in 18 Megohm ASTM
Type I deionized Water

Hydrochloric Acid Blank

CLP-BLH-5 500 mL
CLP-BLH-L-VAP 1L
(2 x 500 mL)

5% HCl in 18 Megohm ASTM Type I
deionized Water

Mixed Acid Blank

CLP-BLMA-5 500 mL
CLP-BLMA-L-VAP 1L
(2 x 500 mL)

5% HCl + 1% HNO₃ in 18 Megohm
ASTM Type I deionized Water

Water Blank

CLP-BLW-5 ▼ 500 mL
CLP-BLW-L-VAP ▼ 1L
(2 x 500 mL)

18 Megohm ASTM Type I deionized
Water

▼ Hazardous fee not required.



ICP/MS

Single Element

AccuStandard's ICP/MS Standards are formulated to meet the needs of this very special instrument. As matrix effect is of utmost concern, each standard is formulated in specially purified 18 megohm de-ionized water and ultra pure acids. After both wet chemical and instrumental analysis, the standards are packaged in acid leached FLPE containers to provide required protection.

**3 Year
Minimum Shelf Life on
Single Element ICP/MS
Standards**

Single Element ICP/MS

Element	Matrix	Unit	100 µg/mL Cat. No.	1,000 µg/mL Cat. No.	10,000 µg/mL Cat. No.
Al (Aluminum)	2-5% HNO ₃	100 mL	ICP-MS-01N-0.01X-1	ICP-MS-01N-0.1X-1	ICP-MS-01N-1
Sb (Antimony)	2-5% HNO ₃	100 mL	ICP-MS-02N-0.01X-1	ICP-MS-02N-0.1X-1	ICP-MS-02N-1
As (Arsenic)	2-5% HNO ₃	100 mL	ICP-MS-03N-0.01X-1	ICP-MS-03N-0.1X-1	ICP-MS-03N-1
Ba (Barium)	2-5% HNO ₃	100 mL	ICP-MS-04N-0.01X-1	ICP-MS-04N-0.1X-1	ICP-MS-04N-1
Be (Beryllium)	2-5% HNO ₃	100 mL	ICP-MS-05N-0.01X-1	ICP-MS-05N-0.1X-1	ICP-MS-05N-1
Bi (Bismuth)	2-10% HNO ₃	100 mL	ICP-MS-06N-0.01X-1	ICP-MS-06N-0.1X-1	ICP-MS-06N-1
B (Boron) ▼	H ₂ O tr. NH ₄ OH	100 mL	ICP-MS-07W-0.01X-1	ICP-MS-07W-0.1X-1	ICP-MS-07W-1
Cd (Cadmium)	2-5% HNO ₃	100 mL	ICP-MS-08N-0.01X-1	ICP-MS-08N-0.1X-1	ICP-MS-08N-1
Ca (Calcium)	2-5% HNO ₃	100 mL	ICP-MS-09N-0.01X-1	ICP-MS-09N-0.1X-1	ICP-MS-09N-1
Ce (Cerium)	2-5% HNO ₃	100 mL	ICP-MS-11N-0.01X-1	ICP-MS-11N-0.1X-1	ICP-MS-11N-1
Cs (Cesium)	2-5% HNO ₃	100 mL	ICP-MS-12N-0.01X-1	ICP-MS-12N-0.1X-1	ICP-MS-12N-1
Cr (Chromium)	2-5% HNO ₃	100 mL	ICP-MS-13N-0.01X-1	ICP-MS-13N-0.1X-1	ICP-MS-13N-1
Co (Cobalt)	2-5% HNO ₃	100 mL	ICP-MS-14N-0.01X-1	ICP-MS-14N-0.1X-1	ICP-MS-14N-1
Cu (Copper)	2-5% HNO ₃	100 mL	ICP-MS-15N-0.01X-1	ICP-MS-15N-0.1X-1	ICP-MS-15N-1
Dy (Dysprosium)	2-5% HNO ₃	100 mL	ICP-MS-16N-0.01X-1	ICP-MS-16N-0.1X-1	ICP-MS-16N-1
Er (Erbium)	2-5% HNO ₃	100 mL	ICP-MS-17N-0.01X-1	ICP-MS-17N-0.1X-1	ICP-MS-17N-1
Eu (Europium)	2-5% HNO ₃	100 mL	ICP-MS-18N-0.01X-1	ICP-MS-18N-0.1X-1	ICP-MS-18N-1
Gd (Gadolinium)	2-5% HNO ₃	100 mL	ICP-MS-19N-0.01X-1	ICP-MS-19N-0.1X-1	ICP-MS-19N-1
Ga (Gallium)	2-5% HNO ₃	100 mL	ICP-MS-20N-0.01X-1	ICP-MS-20N-0.1X-1	ICP-MS-20N-1
Ge (Germanium) ▼	H ₂ O tr. HF	100 mL	ICP-MS-21W-0.01X-1	ICP-MS-21W-0.1X-1	ICP-MS-21W-1
Au (Gold)	10% HCl	100 mL	ICP-MS-22H-0.01X-1	ICP-MS-22H-0.1X-1	ICP-MS-22H-1
Hf (Hafnium)	2-5% HNO ₃ tr. HF	100 mL	ICP-MS-23N-0.01X-1	ICP-MS-23N-0.1X-1	ICP-MS-23N-1
Ho (Holmium)	2-5% HNO ₃	100 mL	ICP-MS-24N-0.01X-1	ICP-MS-24N-0.1X-1	ICP-MS-24N-1
In (Indium)	2-5% HNO ₃	100 mL	ICP-MS-25N-0.01X-1	ICP-MS-25N-0.1X-1	ICP-MS-25N-1
Ir (Iridium)	10% HCl	100 mL	ICP-MS-26H-0.01X-1	ICP-MS-26H-0.1X-1	ICP-MS-26H-1
Fe (Iron)	2-5% HNO ₃	100 mL	ICP-MS-27N-0.01X-1	ICP-MS-27N-0.1X-1	ICP-MS-27N-1
La (Lanthanum)	2-5% HNO ₃	100 mL	ICP-MS-28N-0.01X-1	ICP-MS-28N-0.1X-1	ICP-MS-28N-1
Pb (Lead)	2-5% HNO ₃	100 mL	ICP-MS-29N-0.01X-1	ICP-MS-29N-0.1X-1	ICP-MS-29N-1
Li (Lithium)	2-5% HNO ₃	100 mL	ICP-MS-30N-0.01X-1	ICP-MS-30N-0.1X-1	ICP-MS-30N-1
Lu (Lutetium)	2-5% HNO ₃	100 mL	ICP-MS-31N-0.01X-1	ICP-MS-31N-0.1X-1	ICP-MS-31N-1
Mg (Magnesium)	2-5% HNO ₃	100 mL	ICP-MS-32N-0.01X-1	ICP-MS-32N-0.1X-1	ICP-MS-32N-1
Mn (Manganese)	2-5% HNO ₃	100 mL	ICP-MS-33N-0.01X-1	ICP-MS-33N-0.1X-1	ICP-MS-33N-1
Hg (Mercury)	10% HNO ₃	100 mL	ICP-MS-34N-0.01X-1	ICP-MS-34N-0.1X-1	ICP-MS-34N-1
Mo (Molybdenum) ▼	H ₂ O tr. NH ₄ OH	100 mL	ICP-MS-35W-0.01X-1	ICP-MS-35W-0.1X-1	ICP-MS-35W-1
Nd (Neodymium) ▼	2-5% HNO ₃	100 mL	ICP-MS-36N-0.01X-1	ICP-MS-36N-0.1X-1	ICP-MS-36N-1



▼ Hazardous fee not required.

ICP/MS

Single Element



- Traceable to NIST Reference Materials
- Formulated from Ultra High Purity Starting Materials and Acids
- 18 Megohm de-ionized Water

- Concentration verified by Wet Chemical and Instrumental Analysis
- Packaged in specially prepared Acid leached FLPE bottles
- Actual Lot Analysis provided on label and COA

Single Element ICP/MS

Element	Matrix	Unit	100 µg/mL Cat. No.	1,000 µg/mL Cat. No.	10,000 µg/mL Cat. No.
Ni (Nickel)	2-5% HNO ₃	100 mL	ICP-MS-37N-0.01X-1	ICP-MS-37N-0.1X-1	ICP-MS-37N-1
Nb (Niobium)	H ₂ O tr. HF	100 mL	ICP-MS-38W-0.01X-1	ICP-MS-38W-0.1X-1	ICP-MS-38W-1
Pd (Palladium)	10% HCl	100 mL	ICP-MS-40H-0.01X-1	ICP-MS-40H-0.1X-1	ICP-MS-40H-1
P (Phosphorous) ▼	H ₂ O	100 mL	ICP-MS-41W-0.01X-1	ICP-MS-41W-0.1X-1	ICP-MS-41W-1
Pt (Platinum)	10% HCl	100 mL	ICP-MS-42H-0.01X-1	ICP-MS-42H-0.1X-1	ICP-MS-42H-1
K (Potassium)	2-5% HNO ₃	100 mL	ICP-MS-43N-0.01X-1	ICP-MS-43N-0.1X-1	ICP-MS-43N-1
Pr (Praseodymium)	2-5% HNO ₃	100 mL	ICP-MS-44N-0.01X-1	ICP-MS-44N-0.1X-1	ICP-MS-44N-1
Re (Rhenium) ▼	H ₂ O tr. HNO ₃	100 mL	ICP-MS-45W-0.01X-1	ICP-MS-45W-0.1X-1	ICP-MS-45W-1
Rh (Rhodium)	10% HCl	100 mL	ICP-MS-46H-0.01X-1	ICP-MS-46H-0.1X-1	ICP-MS-46H-1
Rb (Rubidium)	2-5% HNO ₃	100 mL	ICP-MS-47N-0.01X-1	ICP-MS-47N-0.1X-1	ICP-MS-47N-1
Ru (Ruthenium)	10% HCl	100 mL	ICP-MS-48H-0.01X-1	ICP-MS-48H-0.1X-1	ICP-MS-48H-1
Sm (Samarium)	2-5% HNO ₃	100 mL	ICP-MS-49N-0.01X-1	ICP-MS-49N-0.1X-1	ICP-MS-49N-1
Sc (Scandium)	2-5% HNO ₃	100 mL	ICP-MS-50N-0.01X-1	ICP-MS-50N-0.1X-1	ICP-MS-50N-1
Se (Selenium)	2-5% HNO ₃	100 mL	ICP-MS-51N-0.01X-1	ICP-MS-51N-0.1X-1	ICP-MS-51N-1
Si (Silicon) ▼	H ₂ O tr. HF	100 mL	ICP-MS-52W-0.01X-1	ICP-MS-52W-0.1X-1	ICP-MS-52W-1
Ag (Silver)	2-5% HNO ₃	100 mL	ICP-MS-53N-0.01X-1	ICP-MS-53N-0.1X-1	ICP-MS-53N-1
Na (Sodium)	2-5% HNO ₃	100 mL	ICP-MS-54N-0.01X-1	ICP-MS-54N-0.1X-1	ICP-MS-54N-1
Sr (Strontium)	2-5% HNO ₃	100 mL	ICP-MS-55N-0.01X-1	ICP-MS-55N-0.1X-1	ICP-MS-55N-1
S (Sulfur) ▼	H ₂ O	100 mL	ICP-MS-56W-0.01X-1	ICP-MS-56W-0.1X-1	ICP-MS-56W-1
Ta (Tantalum) ▼	H ₂ O tr. HF	100 mL	ICP-MS-57W-0.01X-1	ICP-MS-57W-0.1X-1	ICP-MS-57W-1
Te (Tellurium)	10% HCl (min.)	100 mL	ICP-MS-58H-0.01X-1	ICP-MS-58H-0.1X-1	ICP-MS-58H-1
Tb (Terbium)	2-5% HNO ₃	100 mL	ICP-MS-59N-0.01X-1	ICP-MS-59N-0.1X-1	ICP-MS-59N-1
Tl (Thallium)	2-5% HNO ₃	100 mL	ICP-MS-60N-0.01X-1	ICP-MS-60N-0.1X-1	ICP-MS-60N-1
Th (Thorium)	2-5% HNO ₃	100 mL	ICP-MS-61N-0.01X-1	ICP-MS-61N-0.1X-1	----- --
Tm (Thulium)	2-5% HNO ₃	100 mL	ICP-MS-62N-0.01X-1	ICP-MS-62N-0.1X-1	ICP-MS-62N-1
Sn (Tin)	2-5% HNO ₃ tr. HF	100 mL	ICP-MS-63N-0.01X-1	ICP-MS-63N-0.1X-1	ICP-MS-63N-1
Ti (Titanium) ▼	H ₂ O tr. HF	100 mL	ICP-MS-64W-0.01X-1	ICP-MS-64W-0.1X-1	ICP-MS-64W-1
W (Tungsten) ▼	H ₂ O tr. NH ₄ OH	100 mL	ICP-MS-65W-0.01X-1	ICP-MS-65W-0.1X-1	ICP-MS-65W-1
U (Uranium)	2-5% HNO ₃	100 mL	ICP-MS-66N-0.01X-1	ICP-MS-66N-0.1X-1	----- --
V (Vanadium)	2-5% HNO ₃	100 mL	ICP-MS-67N-0.01X-1	ICP-MS-67N-0.1X-1	ICP-MS-67N-1
Yb (Ytterbium)	2-5% HNO ₃	100 mL	ICP-MS-68N-0.01X-1	ICP-MS-68N-0.1X-1	ICP-MS-68N-1
Y (Yttrium)	2-5% HNO ₃	100 mL	ICP-MS-69N-0.01X-1	ICP-MS-69N-0.1X-1	ICP-MS-69N-1
Zn (Zinc)	2-5% HNO ₃	100 mL	ICP-MS-70N-0.01X-1	ICP-MS-70N-0.1X-1	ICP-MS-70N-1
Zr (Zirconium)	2-5% HNO ₃	100 mL	ICP-MS-71N-0.01X-1	ICP-MS-71N-0.1X-1	ICP-MS-71N-1

▼ Hazardous fee not required.



AA Single Element

Each standard is prepared from high purity starting materials, 18 megohm de-ionized water and high purity acids. Every standard is instrumentally assayed to verify concentration of specified element. Actual Lot Analysis is provided on the label and a Certificate of Analysis is included for ease of record keeping and availability at audits.

- Traceable to NIST Reference Materials
- Certificate of Analysis included
- Actual Concentration Analysis
- High Purity Starting Materials and Acids
- 18 megohm de-ionized Water
- 36 Month Shelf Life



Single Element AA		NoHaz 20 mL Size	
Element Starting Material Matrix	Unit	1000 µg/mL Cat. No.	20 mL (NoHaz) 10,000 µg/mL Cat. No.
Aluminum	100 mL	AA01N-1	AA01N-10X-20ML
2-5% Nitric acid	500 mL	AA01N-5	
Antimony	100 mL	AA02N-1	AA02W-10X-20ML
2-5% HNO ₃ tr. Tartaric acid	500 mL	AA02N-5	
Arsenic	100 mL	AA03N-1	AA03N-10X-20ML
2-5% Nitric acid	500 mL	AA03N-5	
Barium	100 mL	AA04N-1	AA04N-10X-20ML
2-5% Nitric acid	500 mL	AA04N-5	
Beryllium	100 mL	AA05N-1	AA05N-10X-20ML
2-5% Nitric acid	500 mL	AA05N-5	
Bismuth	100 mL	AA06N-1	AA06N-10X-20ML
10% Nitric acid	500 mL	AA06N-5	
Boron	100 mL	AA07W-1 ▼	AA07W-10X-20ML
Water tr. NH ₄ OH	500 mL	AA07W-5 ▼	
Cadmium	100 mL	AA08N-1	AA08N-10X-20ML
2-5% Nitric acid	500 mL	AA08N-5	
Calcium	100 mL	AA09N-1	AA09N-10X-20ML
2-5% Nitric acid	500 mL	AA09N-5	
Cerium	100 mL	AA11N-1	AA11N-10X-20ML
2-5% Nitric acid	500 mL	AA11N-5	
Cesium	100 mL	AA12N-1	AA12N-10X-20ML
2-5% Nitric acid	500 mL	AA12N-5	
Chromium	100 mL	AA13N-1	AA13N-10X-20ML
2-5% Nitric acid	500 mL	AA13N-5	
Cobalt	100 mL	AA14N-1	AA14N-10X-20ML
2-5% Nitric acid	500 mL	AA14N-5	
Copper	100 mL	AA15N-1	AA15N-10X-20ML
2-5% Nitric acid	500 mL	AA15N-5	
Dysprosium	100 mL	AA16N-1	AA16N-10X-20ML
2-5% Nitric acid	500 mL	AA16N-5	
Erbium	100 mL	AA17N-1	AA17N-10X-20ML
2-5% Nitric acid	500 mL	AA17N-5	
Europium	100 mL	AA18N-1	AA18N-10X-20ML
2-5% Nitric acid	500 mL	AA18N-5	
Gadolinium	100 mL	AA19N-1	AA19N-10X-20ML
2-5% Nitric acid	500 mL	AA19N-5	
Gallium	100 mL	AA20N-1	AA20N-10X-20ML
2-5% Nitric acid	500 mL	AA20N-5	
Germanium	100 mL	AA21W-1 ▼	AA21W-10X-20ML
Water tr. HF	500 mL	AA21W-5 ▼	
Gold	100 mL	AA22H-1	AA22H-10X-20ML
5% HCl (min.)	500 mL	AA22H-5	
Hafnium	100 mL	AA23N-1	AA23N-10X-20ML
2-5% Nitric acid tr. HF	500 mL	AA23N-5	
Holmium	100 mL	AA24N-1	AA24N-10X-20ML
2-5% Nitric acid	500 mL	AA24N-5	
Indium	100 mL	AA25N-1	AA25N-10X-20ML
2-5% Nitric acid	500 mL	AA25N-5	
Iridium	100 mL	AA26H-1	AA26H-10X-20ML
10% HCl (min.)	500 mL	AA26H-5	
Iron	100 mL	AA27N-1	AA27N-10X-20ML
2-5% Nitric acid	500 mL	AA27N-5	

▼ Hazardous fee not required.

AA Single Element



3 Year
Minimum Shelf Life on
Single Element AA
Standards

Save with NoHaz Option on AA Single Elements

- No Hazardous Shipping Fees
- Lower Shipping Costs (less weight)
- Yields More - 200 mL from 20 mL concentrate

Includes empty pre-washed, pre-labeled HDPE 250 mL bottle



Single Element AA		NoHaz 20 mL Size	
Element Starting Material Matrix	Unit	1000 µg/mL Cat. No.	20 mL (NoHaz) 10,000 µg/mL Cat. No.
Lanthanum	100 mL	AA28N-1	AA28N-10X-20ML
2-5% Nitric acid	500 mL	AA28N-5	
Lead	100 mL	AA29N-1	AA29N-10X-20ML
2-5% Nitric acid	500 mL	AA29N-5	
Lithium	100 mL	AA30N-1	AA30N-10X-20ML
2-5% Nitric acid	500 mL	AA30N-5	
Lutetium	100 mL	AA31N-1	AA31N-10X-20ML
2-5% Nitric acid	500 mL	AA31N-5	
Magnesium	100 mL	AA32N-1	AA32N-10X-20ML
2-5% Nitric acid	500 mL	AA32N-5	
Manganese	100 mL	AA33N-1	AA33N-10X-20ML
2-5% Nitric acid	500 mL	AA33N-5	
Mercury	100 mL	AA34N-1	AA34N-10X-20ML
2-5% Nitric acid	500 mL	AA34N-5	
Molybdenum	100 mL	AA35W-1 ▼	AA35W-10X-20ML
Water tr. NH ₄ OH	500 mL	AA35W-5 ▼	
Neodymium	100 mL	AA36N-1	AA36N-10X-20ML
2-5% Nitric acid	500 mL	AA36N-5	
Nickel	100 mL	AA37N-1	AA37N-10X-20ML
2-5% Nitric acid	500 mL	AA37N-5	
Niobium	100 mL	AA38W-1 ▼	AA38W-10X-20ML
Water tr. HF	500 mL	AA38W-5 ▼	
Palladium	100 mL	AA40H-1	AA40H-10X-20ML
10% HCl (min.)	500 mL	AA40H-5	
Phosphorus	100 mL	AA41W-1 ▼	AA41W-10X-20ML
Water	500 mL	AA41W-5 ▼	
Platinum	100 mL	AA42H-1	AA42H-10X-20ML
2% HCl (min.)	500 mL	AA42H-5	
Potassium	100 mL	AA43N-1	AA43N-10X-20ML
2-5% Nitric acid	500 mL	AA43N-5	
Praseodymium	100 mL	AA44N-1	AA44N-10X-20ML
2-5% Nitric acid	500 mL	AA44N-5	
Rhenium	100 mL	AA45W-1 ▼	AA45W-10X-20ML
Water	500 mL	AA45W-5 ▼	
Rhodium	100 mL	AA46H-1	AA46H-10X-20ML
10% HCl (min.)	500 mL	AA46H-5	
Rubidium	100 mL	AA47N-1	AA47N-10X-20ML
2-5% Nitric acid	500 mL	AA47N-5	
Ruthenium	100 mL	AA48H-1	AA48H-10X-20ML
10% HCl	500 mL	AA48H-5	
Samarium	100 mL	AA49N-1	AA49N-10X-20ML
2-5% Nitric acid	500 mL	AA49N-5	
Scandium	100 mL	AA50N-1	AA50N-10X-20ML
2-5% Nitric acid	500 mL	AA50N-5	
Selenium	100 mL	AA51N-1	AA51N-10X-20ML
2-5% Nitric acid	500 mL	AA51N-5	
Silicon	100 mL	AA52W-1 ▼	AA52W-10X-20ML
Water tr. HF	500 mL	AA52W-5 ▼	
Silver	100 mL	AA53N-1	AA53N-10X-20ML
2-5% Nitric acid	500 mL	AA53N-5	
Sodium	100 mL	AA54N-1	AA54N-10X-20ML
2-5% Nitric acid	500 mL	AA54N-5	

▼ Hazardous fee not required.



AA Single Element

- Traceable to NIST Reference Materials
- Certificate of Analysis included
- Actual Concentration Analysis
- High Purity Starting Materials and Acids
- 18 megohm de-ionized Water
- 36 Month Shelf Life

3 Year
Minimum Shelf Life on
Single Element AA
Standards

Save with NoHaz Option on AA Single Elements

- No Hazardous Shipping Fees
- Lower Shipping Costs (less weight)
- Yields More - 200 mL from 20 mL concentrate

Includes empty pre-washed, pre-labeled HDPE 250 mL bottle



Single Element AA			NoHaz 20 mL Size
Element Starting Material Matrix	Unit	1000 µg/mL Cat. No.	20 mL (NoHaz) 10,000 µg/mL Cat. No.
Strontium	100 mL	AA55N-1	AA55N-10X-20ML
2-5% Nitric acid	500 mL	AA55N-5	
Sulfur	100 mL	AA56W-1 ▼	AA56W-10X-20ML
Water	500 mL	AA56W-5 ▼	
Tantalum	100 mL	AA57W-1 ▼	AA57W-10X-20ML
Water tr. HF	500 mL	AA57W-5 ▼	
Tellurium	100 mL	AA58H-1	AA58H-10X-20ML
10% HCl (min.)	500 mL	AA58H-5	
Terbium	100 mL	AA59N-1	AA59N-10X-20ML
2-5% Nitric acid	500 mL	AA59N-5	
Thallium	100 mL	AA60N-1	AA60N-10X-20ML
2-5% Nitric acid	500 mL	AA60N-5	
Thorium	100 mL	AA61N-1	----- --
2-5% Nitric acid	500 mL	AA61N-5	
Thulium	100 mL	AA62N-1	AA62N-10X-20ML
2-5% Nitric acid	500 mL	AA62N-5	
Tin	100 mL	AA63N-1	AA63N-10X-20ML
2-5% Nitric acid tr. HF	500 mL	AA63N-5	
Titanium	100 mL	AA64W-1 ▼	AA64W-10X-20ML
Water tr. HF	500 mL	AA64W-5 ▼	
Tungsten	100 mL	AA65W-1 ▼	AA65W-10X-20ML
Water tr. NH ₄ OH	500 mL	AA65W-5 ▼	
Uranium	100 mL	AA66N-1	----- --
2-5% Nitric acid	500 mL	AA66N-5	
Vanadium	100 mL	AA67N-1	AA67N-10X-20ML
5-10% Nitric acid	500 mL	AA67N-5	
Ytterbium	100 mL	AA68N-1	AA68N-10X-20ML
2-5% Nitric acid	500 mL	AA68N-5	
Yttrium	100 mL	AA69N-1	AA69N-10X-20ML
2-5% Nitric acid	500 mL	AA69N-5	
Zinc	100 mL	AA70N-1	AA70N-10X-20ML
2-5% Nitric acid	500 mL	AA70N-5	
Zirconium	100 mL	AA71N-1	AA71N-10X-20ML
2-5% Nitric acid	500 mL	AA71N-5	

▼ Hazardous fee not required.

AA

Matrix Modifier & Calibration



Matrix Modifier Solutions for Graphite Furnace AA

These Matrix Modifiers enhance sensitivity and suppress background interferences observed in trace metal analysis.

Modifier Description	Modifier Source	Unit	Cat. No.
Ammonium dihydrogen phosphate 40% in H ₂ O	NH ₄ H ₂ PO ₄	50 mL	MOD-02-0.5 ▼
		100 mL	MOD-02-1 ▼
		500 mL	MOD-02-5 ▼
Ammonium nitrate 5% in H ₂ O	NH ₄ NO ₃	50 mL	MOD-03-0.5 ▼
		100 mL	MOD-03-1 ▼
		500 mL	MOD-03-5 ▼
Calcium nitrate 2% Calcium in 5% in HNO ₃	Ca(NO ₃) ₂ • 4H ₂ O	50 mL	MOD-04-0.5
		100 mL	MOD-04-1
		500 mL	MOD-04-5
Lanthanum chloride 5% Lanthanum in 5% HCl	LaCl ₃	50 mL	MOD-05-0.5
		100 mL	MOD-05-1
		500 mL	MOD-05-5
Lanthanum nitrate 5% Lanthanum in 5% HNO ₃	LaNO ₃	50 mL	MOD-06-0.5
		100 mL	MOD-06-1
		500 mL	MOD-06-5
Magnesium nitrate 2% Magnesium in 5% HNO ₃	Mg(NO ₃) ₂	50 mL	MOD-07-0.5
		100 mL	MOD-07-1
		500 mL	MOD-07-5
Nickel nitrate 5% Nickel in 5% HNO ₃	Ni(NO ₃) ₂	50 mL	MOD-08-0.5
		100 mL	MOD-08-1
		500 mL	MOD-08-5
Palladium nitrate 0.2% Palladium in 5% HNO ₃	Pd(NO ₃) ₂	50 mL	MOD-09A-0.5
		100 mL	MOD-09A-1
		500 mL	MOD-09A-5
Palladium nitrate 0.5% Palladium in 5% HNO ₃	Pd(NO ₃) ₂	50 mL	MOD-09B-0.5
		100 mL	MOD-09B-1
		500 mL	MOD-09B-5
Palladium nitrate 1.0% Palladium in 10% HNO ₃	Pd(NO ₃) ₂	50 mL	MOD-09C-0.5
		100 mL	MOD-09C-1
		500 mL	MOD-09C-5

Calibration and Matrix Blanks

Nitric Acid Blank

CLP-BLN-5 500 mL
CLP-BLN-L-VAP 1L
(2 x 500 mL)

5% HNO₃ in 18 Megohm ASTM Type I deionized Water

Hydrochloric Acid Blank

CLP-BLH-5 500 mL
CLP-BLH-L-VAP 1L
(2 x 500 mL)

5% HCl in 18 Megohm ASTM Type I deionized Water

Mixed Acid Blank

CLP-BLMA-5 500 mL
CLP-BLMA-L-VAP 1L
(2 x 500 mL)

5% HCl + 1% HNO₃ in 18 Megohm ASTM Type I deionized Water

Water Blank

CLP-BLW-5 ▼ 500 mL
CLP-BLW-L-VAP ▼ 1L
(2 x 500 mL)

18 Megohm ASTM Type I deionized Water

Multi-Element Graphite Furnace AA Calibration & Spiking Standards

GFAA Instrument Calibration Standard

CLP-CAL-AA 50 mL
At stated conc. (µg/mL) in 5% HNO₃ 6 comps.

Sb (Antimony)	100
As (Arsenic)	50
Cd (Cadmium)	10
Pb (Lead)	50
Se (Selenium)	100
Tl (Thallium)	50

GFAA Predigestion Spike Solution

CLP-SP1-AA 50 mL
At stated conc. (µg/mL) in 5% HNO₃ 6 comps.

Sb (Antimony)	100
As (Arsenic)	40
Cd (Cadmium)	5
Pb (Lead)	20
Se (Selenium)	10
Tl (Thallium)	50

GFAA Initial Calibration Verification

(Meets CLP Second Source Requirements)

CLP-ICV-AA 50 mL
At stated conc. (µg/mL) in 5% HNO₃ 6 comps.

Sb (Antimony)	50
As (Arsenic)	25
Cd (Cadmium)	5
Pb (Lead)	25
Se (Selenium)	50
Tl (Thallium)	25

GFAA Mercury Standard for Calibration or Spiking

CLP-HG-AA 50 mL
100 µg/mL in 5% HNO₃

Hg (Mercury)

GFAA Postdigestion Spike Solution

(2 x CRDL except for Lead)
CLP-SP2-AA 50 mL
At stated conc. (µg/mL) in 5% HNO₃ 6 comps.

Sb (Antimony)	120
As (Arsenic)	20
Cd (Cadmium)	10
Pb (Lead)	20
Se (Selenium)	10
Tl (Thallium)	20

GFAA Set

CLP-AA-SET 5 x 50 mL

CLP-CAL-AA	CLP-HG-AA
CLP-SPI-AA	CLP-SP2-AA
CLP-ICV-AA	

▼ Hazardous fee not required.



- 99.99% High Purity Starting Materials
- 18 Megohm, ASTM type I de-ionized Water
- Packaged in pre-cleaned high quality HDPE bottles.
- Each Standard is Supplied with a Certificate of Analysis, stating traceability to NIST, certified value and expiration date.
- Final Solution is filtered through a 0.2 µm filter to eliminate contaminants (such as suspended solids and microbes). This extends shelf life and protects your column.
- Ready-To-Use Mixes and Calibration Sets.
- Standards may be used for other "Classical or Wet" methods.

Anions

Water Matrix	Unit	100 µg/mL Cat. No.	200 µg/mL Cat. No.	1000 µg/mL Cat. No.
Acetate	100 mL	IC-ACET-1X-1	-----	IC-ACET-10X-1
	500 mL	IC-ACET-1X-5	-----	IC-ACET-10X-5
Br (Bromate)	100 mL	-----	-----	IC-BROM-10X-1
	500 mL	-----	-----	IC-BROM-10X-5
Br (Bromide)	100 mL	IC-BR-1X-1	IC-BR-2X-1	IC-BR-10X-1
	500 mL	IC-BR-1X-5	IC-BR-2X-5	IC-BR-10X-5
Citrate	100 mL	-----	-----	IC-CITR-10X-1
Cl (Chloride)	100 mL	IC-Cl-1X-1	IC-Cl-2X-1	IC-Cl-10X-1
	500 mL	IC-Cl-1X-5	IC-Cl-2X-5	IC-Cl-10X-5
Chromate	100 mL	IC-CHRM-1X-1	-----	IC-CHRM-10X-1
	500 mL	IC-CHRM-1X-5	-----	IC-CHRM-10X-5
F (Fluoride)	100 mL	IC-F-1X-1	IC-F-2X-1	IC-F-10X-1
	500 mL	IC-F-1X-5	IC-F-2X-5	IC-F-10X-5
Formate	100 mL	IC-FORM-1X-1	-----	IC-FORM-10X-1
	500 mL	IC-FORM-1X-5	-----	IC-FORM-10X-5
Glycolate	100 mL	-----	-----	IC-GLYC-10X-1
Iodide	100 mL	-----	-----	IC-I-10X-1
Lactate	100 mL	-----	-----	IC-LACT-10X-1
Malate	100 mL	-----	-----	IC-MALA-10X-1
Maleate	100 mL	-----	-----	IC-MALE-10X-1
NO ₂ (Nitrite)	100 mL	IC-NO2-1X-1	IC-NO2-2X-1	IC-NO2-10X-1
	500 mL	IC-NO2-1X-5	IC-NO2-2X-5	IC-NO2-10X-5
NO ₃ (Nitrate)	100 mL	IC-NO3-1X-1	IC-NO3-2X-1	IC-NO3-10X-1
	500 mL	IC-NO3-1X-5	IC-NO3-2X-5	IC-NO3-10X-5
Oxalate	100 mL	IC-OXAL-1X-1	-----	IC-OXAL-10X-1
	500 mL	IC-OXAL-1X-5	-----	IC-OXAL-10X-5
Perchlorate	100 mL	-----	-----	IC-PER-10X-1
Phthalate	100 mL	-----	-----	IC-PHTH-10X-1
PO ₄ (Phosphate)	100 mL	IC-PO4-1X-1	IC-PO4-2X-1	IC-PO4-10X-1
	500 mL	IC-PO4-1X-5	IC-PO4-2X-5	IC-PO4-10X-5
Propionate	100 mL	-----	-----	IC-PROP-10X-1
Succinate	100 mL	-----	-----	IC-SUCC-10X-1
SO ₄ (Sulfate)	100 mL	IC-SO4-1X-1	IC-SO4-2X-1	IC-SO4-10X-1
	500 mL	IC-SO4-1X-5	IC-SO4-2X-5	IC-SO4-10X-5
Sulfide NEW	20 mL	-----	-----	IC-SULF-10X-20ML
Dilute NaOH, stabilizer	5 x 20 mL	-----	-----	IC-SULF-10X-20ML-VAP
Tartrate	100 mL	-----	-----	IC-TART-10X-1

Anion Kits

IC-AN-1X-1-SET 7 x 100 mL

IC-AN-1X-5-SET 7 x 500 mL

Each at 100 µg/mL in Water

IC-AN-2X-1-SET 7 x 100 mL

IC-AN-2X-5-SET 7 x 500 mL

Each at 200 µg/mL in Water

IC-AN-10X-1-SET 7 x 100 mL

IC-AN-10X-5-SET 7 x 500 mL

Each at 1000 µg/mL in Water

Fluoride	Bromide
Chloride	Phosphate
Nitrite	Sulfate
Nitrate	

Inorganic products containing acid generally require a hazardous shipping fee.

Inorganic products in water generally do not.



Ion Chrom - Ion Singles as the Element

	Unit	100 µg/mL	1000 µg/mL
NO₂-N (Nitrite-Nitrogen)	100 mL	IC-NO2-N-1X-1	IC-NO2-N-10X-1
Water Matrix	500 mL	IC-NO2-N-1X-5	IC-NO2-N-10X-5
NO₃-N (Nitrate-Nitrogen)	100 mL	IC-NO3-N-1X-1	IC-NO3-N-10X-1
Water Matrix	500 mL	IC-NO3-N-1X-5	IC-NO3-N-10X-5
PO₄-P (Phosphate-Phosphorous)	100 mL	IC-PO4-P-1X-1	IC-PO4-P-10X-1
Water Matrix	500 mL	IC-PO4-P-1X-5	IC-PO4-P-10X-5
SO₄-S (Sulfate-Sulfur)	100 mL	IC-SO4-S-1X-1	IC-SO4-S-10X-1
Water Matrix	500 mL	IC-SO4-S-1X-5	IC-SO4-S-10X-5
NH₄-N (Ammonium-Nitrogen)	100 mL	IC-NH4-N-1X-1	IC-NH4-N-10X-1
Water Matrix	500 mL	IC-NH4-N-1X-5	IC-NH4-N-10X-5

Organic Acid Salt Standard

	Unit	100 µg/mL	1000 µg/mL
Formate	100 mL	IC-FORM-1X-1	IC-FORM-10X-1
Water Matrix	500 mL	IC-FORM-1X-5	IC-FORM-10X-5
Acetate	100 mL	IC-ACET-1X-1	IC-ACET-10X-1
Water Matrix	500 mL	IC-ACET-1X-5	IC-ACET-10X-5
Oxalate	100 mL	IC-OXAL-1X-1	IC-OXAL-10X-1
Water Matrix	500 mL	IC-OXAL-1X-5	IC-OXAL-10X-5
Chromate	100 mL	IC-CHRM-1X-1	IC-CHRM-10X-1
Water Matrix	500 mL	IC-CHRM-1X-5	IC-CHRM-10X-5

Anion Single Kits

IC-AN-R-10X-1-SET	7 x 100 mL
IC-AN-R-10X-5-SET	7 x 500 mL
<i>Each at 1000 µg/mL</i>	
Fluoride	Bromide
Chloride	Phosphate-Phosphorous
Nitrite-Nitrogen	Sulfate-Sulfur
Nitrate-Nitrogen	



Method 314.0 Perchlorate in Drinking Water by IC

Perchlorate has become an analyte of environmental interest since being detected in a number of drinking and groundwater supplies located in Midwestern states. EPA method 314.0 was released as an approved method to achieve the required sensitivity.

Perchlorate Standard

IC-PER-10X-1 100 mL
1000 µg/mL in Water
Perchlorate

Conductivity Meter Calibration Standard

M-314.0-CMCS-1 100 mL
1410 µs/cm @ 25°C in Water

Mixed Common Anion Stock

M-314.0-MCA-250X-1 100 mL
25 mg/mL in Water 3 comps.
Chloride
Sulfate
Carbonate

Method 314.0 Perchlorate Calibration Set

M-314.0-SET 100 mL
IC-PER-10X-1
M-314.0-MCA-250X-1
M-314.0-CMCS-1



Ion Chrom Eluents

0.5 M Sodium bicarbonate (100X concentrate)	50 mL	100 mL	5 x 50 mL	5 x 100 mL
	IC-ELU-01-0.5	IC-ELU-01-1	IC-ELU-01-0.5-PAK	IC-ELU-01-1-PAK
0.5 M Sodium carbonate (100X concentrate)	50 mL	100 mL	5 x 50 mL	5 x 100 mL
	IC-ELU-02-0.5	IC-ELU-02-1	IC-ELU-02-0.5-PAK	IC-ELU-02-1-PAK
0.18 M Sodium carbonate/ 0.17 M Sodium bicarbonate (100X concentrate)	50 mL	100 mL	5 x 50 mL	5 x 100 mL
	IC-ELU-03-0.5	IC-ELU-03-1	IC-ELU-03-0.5-PAK	IC-ELU-03-1-PAK

Technical Note

Easy to use, ready to dilute concentrates. Open a fresh bottle and dilute the volume (50 mL to 5 L or 100 mL to 10 L) and be assured of a fresh uncontaminated mobile phase



Anion Mixes

Anion Mix #1

IC-MAN-01-1	100 mL
Water Matrix	5 comps.
F (Fluoride)	20 µg/mL
Cl (Chloride)	30 µg/mL
NO ₃ (Nitrate)	100 µg/mL
PO ₄ (Phosphate)	150 µg/mL
SO ₄ (Sulfate)	150 µg/mL

Anion Mix #2

IC-MAN-02-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	100 µg/mL
Cl (Chloride)	200 µg/mL
Br (Bromide)	400 µg/mL
NO ₃ (Nitrate)	400 µg/mL
PO ₄ (Phosphate)	600 µg/mL
SO ₄ (Sulfate)	400 µg/mL

Anion Mix #3

IC-MAN-03-1	100 mL
Water Matrix	3 comps.
F (Fluoride)	100 µg/mL
Cl (Chloride)	100 µg/mL
SO ₄ (Sulfate)	100 µg/mL

Anion Mix #4

IC-MAN-04-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	100 µg/mL
Cl (Chloride)	100 µg/mL
Br (Bromide)	100 µg/mL
NO ₃ (Nitrate)	100 µg/mL
PO ₄ (Phosphate)	100 µg/mL
SO ₄ (Sulfate)	100 µg/mL

Anion Mix #5

IC-MAN-05-R1-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	10 µg/mL
Cl (Chloride)	20 µg/mL
Br (Bromide)	20 µg/mL
NO ₃ (Nitrate)	20 µg/mL
PO ₄ (Phosphate)	5 µg/mL
SO ₄ (Sulfate)	30 µg/mL

Anion Mix #6

IC-MAN-06-R1-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	1 µg/mL
Cl (Chloride)	5 µg/mL
Br (Bromide)	5 µg/mL
NO ₃ (Nitrate)	5 µg/mL
PO ₄ (Phosphate)	5 µg/mL
SO ₄ (Sulfate)	10 µg/mL

Anion Mix #7

IC-MAN-07-R1-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	1 µg/mL
Cl (Chloride)	10 µg/mL
Br (Bromide)	10 µg/mL
NO ₃ (Nitrate)	10 µg/mL
PO ₄ (Phosphate)	10 µg/mL
SO ₄ (Sulfate)	10 µg/mL

Anion Mix #8

IC-MAN-08-R1-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	10 µg/mL
Cl (Chloride)	20 µg/mL
Br (Bromide)	20 µg/mL
NO ₃ (Nitrate)	20 µg/mL
PO ₄ (Phosphate)	20 µg/mL
SO ₄ (Sulfate)	20 µg/mL

Anion Mix #9

IC-MAN-09-R1-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	20 µg/mL
Cl (Chloride)	40 µg/mL
Br (Bromide)	40 µg/mL
NO ₃ (Nitrate)	40 µg/mL
PO ₄ (Phosphate)	40 µg/mL
SO ₄ (Sulfate)	40 µg/mL

Anion Mix #10

IC-MAN-10-R1-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	25 µg/mL
Cl (Chloride)	50 µg/mL
Br (Bromide)	50 µg/mL
NO ₃ (Nitrate)	50 µg/mL
PO ₄ (Phosphate)	50 µg/mL
SO ₄ (Sulfate)	50 µg/mL

Anion Mix #11

IC-MAN-11-1	100 mL
Water Matrix	5 comps.
Cl (Chloride)	1000 µg/mL
Br (Bromide)	1000 µg/mL
NO ₃ (Nitrate)	1000 µg/mL
PO ₄ (Phosphate)	1000 µg/mL
SO ₄ (Sulfate)	1000 µg/mL

Anion Mix #12

IC-MAN-12-1	100 mL
Water Matrix	5 comps.
Cl (Chloride)	15 µg/mL
Br (Bromide)	15 µg/mL
NO ₃ (Nitrate)	15 µg/mL
PO ₄ (Phosphate)	15 µg/mL
SO ₄ (Sulfate)	15 µg/mL

Anion Mix #13

IC-MAN-13-1	100 mL
Water Matrix	3 comps.
F (Fluoride)	25 µg/mL
Cl (Chloride)	50 µg/mL
SO ₄ (Sulfate)	100 µg/mL

Anion Mix #14

IC-MAN-14-R3-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	20 µg/mL
Cl (Chloride)	30 µg/mL
Br (Bromide)	100 µg/mL
NO ₃ (Nitrate)	100 µg/mL
PO ₄ (Phosphate)	150 µg/mL
SO ₄ (Sulfate)	150 µg/mL

Anion Mix #14 Revised

IC-MAN-14-R2-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	20 µg/mL
Cl (Chloride)	30 µg/mL
Br (Bromide)	100 µg/mL
N-NO ₃ (Nitrogen-Nitrate)	100 µg/mL
P-PO ₄ (Phosphorus-Phosphate)	150 µg/mL
SO ₄ (Sulfate)	150 µg/mL

Anion Mix #14-R2 plus
IC-NO2-N-1X is perfect for
US EPA Method 300.1

Nitrite

IC-NO2-N-1X-1	100 mL
NO ₂ (Nitrite)	100 µg/mL

Dichloroacetate Surrogate Standard

M-300.1-SS	100 mL
0.5 mg/mL Dichloroacetate in Water	

Nitrite

IC-NO2-10X-1	100 mL
NO ₂ (Nitrite)	1000 µg/mL
IC-NO2-1X-1	100 mL
NO ₂ (Nitrite)	100 µg/mL
IC-NO2-0.1X-1	100 mL
NO ₂ (Nitrite)	10 µg/mL

Anion Mix #15

IC-MAN-15-R2-1	100 mL
Water Matrix	3 comps.
F (Fluoride)	100 µg/mL
NO ₃ (Nitrate)	100 µg/mL
PO ₄ (Phosphate)	100 µg/mL

Anion Mix #16

IC-MAN-16-R1-1	100 mL
Water Matrix	6 comps.
F (Fluoride)	100 µg/mL
Cl (Chloride)	200 µg/mL
Br (Bromide)	100 µg/mL
NO ₃ (Nitrate)	20 µg/mL
PO ₄ (Phosphate)	30 µg/mL
SO ₄ (Sulfate)	20 µg/mL

Anion Mix #17

IC-MAN-17-R1-1	100 mL
Water Matrix	5 comps.
Cl (Chloride)	20 µg/mL
Br (Bromide)	12 µg/mL
NO ₃ (Nitrate)	12 µg/mL
PO ₄ (Phosphate)	12 µg/mL
SO ₄ (Sulfate)	80 µg/mL

Anion Mix #18

IC-MAN-18-R1-1	100 mL
Water Matrix	5 comps.
F (Fluoride)	100 µg/mL
Cl (Chloride)	100 µg/mL
NO ₃ (Nitrate)	100 µg/mL
PO ₄ (Phosphate)	100 µg/mL
SO ₄ (Sulfate)	100 µg/mL

Technical Note

To enhance the shelf life and long term stability of our IC products, Nitrite has been removed from mixes that contain Nitrate.

Technical Note

We offer several Nitrite concentrations that can be added just prior to analysis for maximum stability.



Ion Chrom - Cation Singles

Dilute HNO ₃ Matrix	Unit	100 µg/mL Cat. No.	200 µg/mL Cat. No.	1,000 µg/mL Cat. No.
Ca (Calcium)	100 mL	IC-CA-1X-1	IC-CA-2X-1	IC-CA-10X-1
	500 mL	IC-CA-1X-5	IC-CA-2X-5	IC-CA-10X-5
NH ₄ (Ammonium)	100 mL	IC-NH4-1X-1	IC-NH4-2X-1	IC-NH4-10X-1 †
	500 mL	IC-NH4-1X-5	IC-NH4-2X-5	IC-NH4-10X-5 †
Mg (Magnesium)	100 mL	IC-MG-1X-1	IC-MG-2X-1	IC-MG-10X-1
	500 mL	IC-MG-1X-5	IC-MG-2X-5	IC-MG-10X-5
K (Potassium)	100 mL	IC-K-1X-1	IC-K-2X-1	IC-K-10X-1
	500 mL	IC-K-1X-5	IC-K-2X-5	IC-K-10X-5
Na (Sodium)	100 mL	IC-NA-1X-1	IC-NA-2X-1	IC-NA-10X-1
	500 mL	IC-NA-1X-5	IC-NA-2X-5	IC-NA-10X-5
Li (Lithium)	100 mL	IC-LI-1X-1	IC-LI-2X-1	IC-LI-10X-1
	500 mL	IC-LI-1X-5	IC-LI-2X-5	IC-LI-10X-5
Ba (Barium)	100 mL	IC-BA-1X-1	IC-BA-2X-1	IC-BA-10X-1
	500 mL	IC-BA-1X-5	IC-BA-2X-5	IC-BA-10X-5
Sr (Strontium)	100 mL	IC-SR-1X-1	IC-SR-2X-1	IC-SR-10X-1
	500 mL	IC-SR-1X-5	IC-SR-2X-5	IC-SR-10X-5
Sets	8 x 100 mL	IC-CAT-1X-1-SET	IC-CAT-2X-1-SET	IC-CAT-10X-1-SET
Above Listed	8 x 500 mL	IC-CAT-1X-5-SET	IC-CAT-2X-5-SET	IC-CAT-10X-5-SET

† 1,000 µg/mL as Ammonium (NH₄) Other Nitrogen species equivalents are:

NH₃ (Ammonia) = 944 µg/mL N (Nitrogen) = 776 µg/mL

Ion Chrom - Cation Mixes

Cation Mix #1

IC-MCA-01-1	100 mL
Dilute HNO ₃	6 comps.
Ca (Calcium)	1000 µg/mL
NH ₄ (Ammonium)	400 µg/mL
Mg (Magnesium)	200 µg/mL
K (Potassium)	200 µg/mL
Na (Sodium)	200 µg/mL
Li (Lithium)	50 µg/mL

Cation Mix #3

IC-MCA-03-1	100 mL
Dilute HNO ₃	4 comps.
Ca (Calcium)	100 µg/mL
K (Potassium)	100 µg/mL
Na (Sodium)	50 µg/mL
Li (Lithium)	10 µg/mL

Cation Mix #5

IC-MCA-05-1	100 mL
Dilute HNO ₃	4 comps.
NH ₄ (Ammonium)	3 µg/mL
K (Potassium)	6 µg/mL
Na (Sodium)	3 µg/mL
Li (Lithium)	0.5 µg/mL

Cation Mix #6

IC-MCA-06-1	100 mL
Dilute HNO ₃	6 comps.
Ca (Calcium)	2 µg/mL
NH ₄ (Ammonium)	1.5 µg/mL
Mg (Magnesium)	2 µg/mL
K (Potassium)	2.5 µg/mL
Na (Sodium)	1.5 µg/mL
Li (Lithium)	0.2 µg/mL

Cation Mix #2


IC-MCA-02-1	100 mL
Dilute HNO ₃	6 comps.
Ca (Calcium)	100 µg/mL
NH ₄ (Ammonium)	100 µg/mL
Mg (Magnesium)	100 µg/mL
K (Potassium)	100 µg/mL
Na (Sodium)	100 µg/mL
Li (Lithium)	100 µg/mL

Cation Mix #4

IC-MCA-04-1	100 mL
Dilute HNO ₃	4 comps.
Ca (Calcium)	400 µg/mL
Mg (Magnesium)	200 µg/mL
Ba (Barium)	1600 µg/mL
Sr (Strontium)	600 µg/mL

Inorganic products containing acid generally require a hazardous shipping fee.

Inorganic products in water generally do not.





Wet Chemicals

Our Wet Chemical Standards are prepared from the highest quality raw material according to ASTM, EPA or "Standard Methods" ¹ procedures. All balances used for preparation are calibrated regularly against NIST traceable weights. Each batch of finished product is analyzed to verify concentration, against NIST standards when possible. All of our Wet Chemical standards are subjected to the same rigorous quality control procedures as our ICP and IC standards.

¹ Standard Methods for the Examination of Water and Wastewater. American Public Health Association, American Water Works Association, Water Environment Federation

Inorganic Constituents

Many of these methods use classical wet chemical methods to determine the components of either potable or wastewater.

Bromide

IC-BR-10X-1 100 mL
1000 µg/mL Bromide in Water

Method 300.1 Ion Chrom Standard Revised

IC-MAN-14-R2-1 100 mL
Water Matrix 6 comps.

F (Fluoride)	20 µg/mL
Cl (Chloride)	30 µg/mL
Br (Bromide)	100 µg/mL
NO ₃ -N (Nitrate-Nitrogen)	100 µg/mL
PO ₄ -P (Phosphate-Phosphorus)	150 µg/mL
SO ₄ (Sulfate)	150 µg/mL

Technical Note

This product was designed to more closely meet the EPA standard by having the concentrations for the nutrients calculated back to the element rather than the anion.

Dichloroacetate Surrogate Standard

M-300.1-SS 100 mL
0.5 mg/mL Dichloroacetate in Water

Cyanide

WC-CN-10X-1 ▲ 100 mL
1000 µg/mL Cyanide in 2% NaOH

Chloride

IC-CL-10X-1 100 mL
1000 µg/mL Chloride in Water

Total Residual Chlorine

WC-TRC-10X-10ML 10 mL
1000 µg/mL Chlorine in Water

Technical Note

This product is shipped in an amber flame sealed ampule for maximum stability.

Fluoride

IC-F-10X-1 100 mL
1000 µg/mL Fluoride in Water

Iodide

IC-I-10X-1 100 mL
1000 µg/mL Iodide in Water

pH

WC-PH-4-1 100 mL
pH of 4.0 in Water
WC-PH-7-1 100 mL
pH of 7.0 in Water
WC-PH-10-1 100 mL
pH of 10.0 in Water

Phosphorus - Total

IC-PO4-P-10X-1 100 mL
1000 µg/mL Phosphorus in Water

Technical Note

Can also be used for ortho-phosphate analysis.

Technical Note

Nitrogen Species are all calculated back to Nitrogen - Not the Anion or Cation species.

Nitrogen - Ammonium

IC-NH4-N-10X-1 100 mL
1000 µg/mL Ammonium-Nitrogen in Water

Nitrogen - Nitrite

IC-NO2-N-10X-1 100 mL
1000 µg/mL Nitrite-Nitrogen in Water

Nitrogen - Nitrate

IC-NO3-N-10X-1 100 mL
1000 µg/mL Nitrate-Nitrogen in Water

Silica

WC-SIO2-10X-1 100 mL
1000 µg/mL as Silica (SiO₂) in Water tr. HF

Sulfate

IC-SO4-10X-1 100 mL
1000 µg/mL Sulfate (SO₄) in Water

Hexavalent Chromium

WC-HEX-10X-1 100 mL
1000 µg/mL in Water

Standards of Interest

Anions available in additional concentrations, see page 358 for complete product listing.

▲ Hazardous fee require

Physical & Aggregate Properties

These Standards are concerned primarily with measuring actual physical characteristics of a sample as opposed to the chemical characteristics. These analytes are measured frequently in both drinking and waste waters.

Turbidity

WC-TURB-4X-1 100 mL
400 NTU non-ratio Turbidity Standard

A stable solution of microspheres in an aqueous matrix. Can be diluted in turbidity free water to make a calibration curve. Do not shake prior to use.

Alkalinity

WC-ALK-10X-1 100 mL
1000 µg/L CaCO₃ to pH 4.5

Hardness

WC-HARD-10X-1 100 mL
1000 µg/mL equivalent CaCO₃

A combination of Ca and Mg to give an approximate concentration of 1000 µg/mL CaCO₃. Hardness µg/mL equivalent CaCO₃ = 2.497 [Ca µg/mL] + 4.118 [Mg µg/mL]

Conductivity

WC-COND-10X-1 100 mL
1000 µmhos in Water

Solids

WC-SOL sample
2 comps.
1000 ppm TSS (Total Suspended Solids) and 1000 ppm TDS (Total Dissolved Solids) for a 2000 ppm TS (Total Solids).
Dilute to 100 mL. Rinse vial and cap several times to recover all solids.

Methylene Blue Activated Substance (MBAS)

WC-MBAS-R1-10X-1 100 mL
1000 µg/mL in Water



Aggregate Organic

Rather than determining individual organic analytes, these Standards are used to determine organic matter in broad categories, based primarily on how they react.

Biochemical Oxygen Demand (BOD)

WC-BOD-10ML 10 mL
100 µg/mL BOD (After Dilution)

75 mg/L glucose and 75 mg / L glutamic acid provided in a flame sealed ampule. Dilute to 1L immediately before use.

Chemical Oxygen Demand (COD)

WC-COD-5X-10ML 10 mL
500 µg/mL COD in water

Total Organic Carbon (TOC)

WC-TOC-10X-1 100 mL
1000 µg/mL TOC in water, tr. H₂SO₄

Total Inorganic Carbon (TIC)

WC-TIC-10X-1 100 mL
1000 µg/mL Total Inorganic Carbon in Water tr. H₂SO₄

Total Organic Halides (TOX)

WC-TOX-10X-1 1 mL
WC-TOX-10X-1-PAK SAVE 5 x 1 mL
1000 µg/mL in MeOH

Total Organic Nitrogen (TON)

WC-TON-10X-1 100 mL
1000 µg/mL Total Organic Nitrogen in Water

Total Kjeldahl Nitrogen (TKN)

WC-TKN-10X-1 100 mL
1000 µg/mL Total Kjeldahl Nitrogen in Water

Oil and Grease

WC-OILG-10X-1 ▲ 100 mL
1000 µg/mL Total Oil and Grease in n-Propanol

Contains 500 µg/mL vegetable oil and 500 µg/mL of petroleum oil. Shake well before use.

Phenols

WC-PHEN-10X-1 100 mL
1000 µg/mL Phenol in water.

▲ Hazardous fee required



Inorganic products containing acid generally require a hazardous shipping fee.

Inorganic products in water generally do not.





TPH, Oil and Grease

EPA Methods

Method 1664 Oil, Grease & Total Petroleum Hydrocarbon (TPH)

Precision and Recovery (PAR) Spiking Solution

M-1664-5ML		1 x 5 mL
M-1664-5ML-PAK	SAVE	5 x 5 mL
4.0 mg/mL each in Acetone		2 comps.
M-1664-20ML		1 x 20 mL
M-1664-20ML-PAK	SAVE	5 x 20 mL
4.0 mg/mL each in Acetone		2 comps.
Hexadecane	Stearic acid	

Technical Note

This Precision and Recovery (PAR) Spiking Solution was developed for Method 1664. This performance based method was developed to replace previous gravimetric procedures incorporating Freon-113 as the extraction solvent for the determination of Oil and Grease and Total Petroleum Hydrocarbons. Each standard is packaged in a flame sealed ampule conveniently sized for quality control of the analytical batch.

Method 413.2 & 418.1 Total Petroleum Hydrocarbon Analysis by IR

Oil, Grease & Petroleum Hydrocarbon Concentrates Mix

M-418-CON		1 x 1 mL
% by volume		3 comps.
Chlorobenzene (25.0)	Hexadecane (37.5)	
Isooctane (37.5)		

Oil, Grease and Petroleum Hydrocarbon Total Recoverable (IR Method)

M-418		1 x 1 mL
M-418-PAK	SAVE	5 x 1 mL
Total 4.15 mg/mL in Freon 113, (Parts by volume)		3 comps.
Chlorobenzene (10.0)	Isooctane (15.0)	
n-Hexadecane (15.0)		

Method 8440 Total Petroleum Hydrocarbon Analysis

Total Recoverable Petroleum Hydrocarbon Mix

M-8440		1 x 1 mL
M-8440-PAK	SAVE	5 x 1 mL
At stated conc. in Tetrachloroethene		3 comps.
Chlorobenzene (0.10 w/w %)	Isooctane (0.15 w/w %)	
n-Hexadecane (0.15 w/w %)		

Total Petroleum Hydrocarbon Concentrate Mix

M-8440-CON		1 x 1 mL
M-8440-CON-PAK	SAVE	5 x 1 mL
		3 comps.
Chlorobenzene (25.0 vol %)	Isooctane (37.5 vol %)	
n-Hexadecane (37.5 vol %)		

Silica Gel Cleanup Calibration Solution

M-8440-SGC		1 x 1 mL
M-8440-SGC-PAK	SAVE	5 x 1 mL
10.0 mg/mL in Tetrachloroethene		
Corn Oil		

Standards of Interest

AccuStandard offers the widest selection of Petroleum standards. These TPH oil analysis standards are a sample of the hundreds in our Fuel and Hydrocarbon section.

Inorganic Multi-Element Standards

- Traceable to NIST Reference Materials
- Formulated from Ultra High Purity Starting Materials, Water and Acids
- Meets EPA and Customer Applications
- Concentration Verified by Wet Chemical and Instrumental Analysis
- Packaged in Specially prepared Acid Leached Bottles
- Full Documentation



Inorganic products containing acid generally require a hazardous shipping fee.

Inorganic products in water generally do not.

Multi-Element Contents

ICP Multi-Element Standards	366-392
Quality Control and Second Source QC	366
Instrument Check and Screening	367
SDWA Drinking Water	368
Groundwater & Wastewater	369
TCLP	369
MISA Test Group 29 Analysis	370
Contract Laboratory Program (CLP)	371-373
Calibration Check Standards	371
Verification and Spiking	372
Interference Check	373
Detection Limit	373
SW-846, EPA Method 200.7	374-377
Calibration Standards	374
Instrument Performance Check Standards	375
LPCS and LFSS	375
Instrument Fortifying Standards	376
Spiking Standards	377
Interference Check Standards	377
EPA Method 6010	378-379
Merck Multi-Element	381-382
Alternate Source (Agilent/Varian, Perkin Elmer, Jobin Yvon, Teledyne)	383-391
EU Formulation	392
ASTM Methods D5184, D5600	392
ICP/MS Multi-Element Standards	393-396
Calibration Standards and Blanks	393
Tuning Solutions	394
Interference Check	394
Memory Check	394
Spiking Standards for Water & Soil	395
Quality Control	395
Internal Standards	395
EPA Method 200.8	396
EPA Method 6020	396



ICP

Multi-Element QC & Second Source QC

Quality Control Standards

Quality Control Standards can be used for many different applications. AccuTrace QC Standards are ideal for calibration when performing NPDES monitoring requirements and can be used for standard curve checks, inter-element correction methods, interference checks or any other unique application.

QC Standard #1

QCS-01-1 100 mL
QCS-01-5 500 mL
 100 µg/mL each in 5% HNO₃ tr. HF 23 comps.

Sb (Antimony)	Mn (Manganese)
As (Arsenic)	Mo (Molybdenum)
Be (Beryllium)	Ni (Nickel)
Cd (Cadmium)	P (Phosphorus)
Ca (Calcium)	Se (Selenium)
Cr (Chromium)	Sr (Strontium)
Co (Cobalt)	Tl (Thallium)
Cu (Copper)	Sn (Tin)
Fe (Iron)	Ti (Titanium)
Pb (Lead)	V (Vanadium)
Li (Lithium)	Zn (Zinc)
Mg (Magnesium)	

QC Standard #2

QCS-02-1 100 mL
QCS-02-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF 7 comps

Al (Aluminum)	100
Ba (Barium)	100
B (Boron)	100
K (Potassium)	1000
Si (Silicon) †	500
Ag (Silver)	50
Na (Sodium)	100

† 1070 µg/mL as SiO₂

QC Standard #2R

QCS-02-R1-1 100 mL
QCS-02-R1-5 500 mL
 100 µg/mL each in 5% HNO₃ tr. HF 7 comps.

Al (Aluminum)	Si (Silicon) †
Ba (Barium)	Ag (Silver)
B (Boron)	Na (Sodium)
K (Potassium)	† 214 µg/mL as SiO ₂

QC Standard #3

QCS-03-1 100 mL
QCS-03-5 500 mL
 100 µg/mL each in 5% HNO₃ 15 comps.

Al (Aluminum)	Pb (Lead)
Ba (Barium)	Mg (Magnesium)
Cd (Cadmium)	Mn (Manganese)
Ca (Calcium)	Ni (Nickel)
Cr (Chromium)	Na (Sodium)
Co (Cobalt)	Ti (Titanium)
Cu (Copper)	Zn (Zinc)
Fe (Iron)	

QC Standard #4

QCS-04-1 100 mL
 At stated conc. (µg/mL) in 5% HNO₃ 19 comps.

Al (Aluminum)	100
Ba (Barium)	5
Be (Beryllium)	1
Bi (Bismuth)	200
B (Boron)	15
Cd (Cadmium)	20
Cr (Chromium)	25
Co (Cobalt)	20
Cu (Copper)	20
Ga (Gallium)	150
In (Indium)	200
Fe (Iron)	15
Pb (Lead)	200
Mn (Manganese)	5
Ni (Nickel)	50
Ag (Silver)	50
Sr (Strontium)	1
Tl (Thallium)	40
Zn (Zinc)	20

QC Standard #5

QCS-05-1 100 mL
 At stated conc. (µg/mL) in 2% HNO₃ 3 comps.

Li (Lithium)	250
K (Potassium)	10,000
Na (Sodium)	1000

QC Standard #6

QCS-06-1 100 mL
 1000 µg/mL each in 2% HNO₃ 4 comps.

Ba (Barium)	Mg (Magnesium)
Ca (Calcium)	Sr (Strontium)

Quality Control Standards Sets

QCS-1-SET 3 x 100 mL
 QCS-01-1 QCS-02-1 QCS-03-1

QCS-5-SET 3 x 500 mL
 QCS-01-5 QCS-02-5 QCS-03-5

QCS-R1-1-SET 3 x 100 mL
 QCS-01-1 QCS-02-R1-1 QCS-03-1

QCS-R1-5-SET 3 x 500 mL
 QCS-01-5 QCS-02-R1-5 QCS-03-5



Second Source QC Standards

These Alternative Source Line Standards exactly match a formulation from another source you may be already using. Our ASL formulations save you the cost of a custom formulation by providing you with true independent lots.

- NIST Traceable
- Independent Lots
- Exact Match

Second Source QC Standard #1

QCS-ASL-7-1 1 x 100 mL
QCS-ASL-7-5 1 x 500 mL
 At stated conc. (µg/mL) in 2-5% HNO₃ tr. HF 7 comps.

Al (Aluminum)	100
Ba (Barium)	100
B (Boron)	100
K (Potassium)	1000
Si (Silicon)	50
Ag (Silver)	100
Na (Sodium)	100

Second Source QC Standard #2

QCS-ASL-21-1 1 x 100 mL
QCS-ASL-21-5 1 x 500 mL
 100 µg/mL each in 2-5% HNO₃ tr. HF 21 comps.

Sb (Antimony)	Mg (Magnesium)
As (Arsenic)	Mn (Manganese)
Be (Beryllium)	Mo (Molybdenum)
Cd (Cadmium)	Ni (Nickel)
Ca (Calcium)	Se (Selenium)
Cr (Chromium)	Sr (Strontium)
Co (Cobalt)	Tl (Thallium)
Cu (Copper)	Ti (Titanium)
Fe (Iron)	V (Vanadium)
Pb (Lead)	Zn (Zinc)
Li (Lithium)	

Second Source QC Standard #3

QCS-ASL-19-1 1 x 100 mL
QCS-ASL-19-5 1 x 500 mL
 100 µg/mL each in 2-5% HNO₃ tr. HF 19 comps.

Sb (Antimony)	Mg (Magnesium)
As (Arsenic)	Mn (Manganese)
Be (Beryllium)	Mo (Molybdenum)
Cd (Cadmium)	Ni (Nickel)
Ca (Calcium)	Se (Selenium)
Cr (Chromium)	Tl (Thallium)
Co (Cobalt)	Ti (Titanium)
Cu (Copper)	V (Vanadium)
Fe (Iron)	Zn (Zinc)
Pb (Lead)	

Match Other Supplier's Products.
 Use as a **True Second Source.**

ICP

Instrument Check Standards & Screening Standards



Instrument Check Standards

These instrument check standards are used to verify ICP instrumentation performance over specific wavelength ranges from 160 nm to 790 nm. These standards are ideal for method development, technician training and other calibration uses.

Instrument Check Standard #1

ICS-01-1 100 mL
ICS-01-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 9 comps.

Al (Aluminum)	100
Ba (Barium)	10
Be (Beryllium)	10
B (Boron)	100
Ca (Calcium)	10
Ni (Nickel)	100
P (Phosphorus)	1000
Sc (Scandium)	10
Zn (Zinc)	100

Instrument Check Standard #2

ICS-02-1 100 mL
ICS-02-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 7 comps.

Ba (Barium)	50
Be (Beryllium)	20
La (Lanthanum)	20
Mn (Manganese)	20
Ni (Nickel)	20
Sc (Scandium)	20
Zn (Zinc)	20

Instrument Check Standard #3

ICS-03-1 100 mL
ICS-03-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ tr. HF 11 comps.

As (Arsenic)	20
La (Lanthanum)	20
Li (Lithium)	20
Mn (Manganese)	20
Mo (Molybdenum)	20
Ni (Nickel)	20
P (Phosphorus)	100
K (Potassium)	100
Sc (Scandium)	20
Na (Sodium)	20
S (Sulfur)	100

Instrument Check Standard #4

ICS-04-1 100 mL
ICS-04-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 12 comps.

Al (Aluminum)	100
As (Arsenic)	100
Ba (Barium)	10
Cu (Copper)	100
Pb (Lead)	100
Mn (Manganese)	100
Ni (Nickel)	100
P (Phosphorus)	100
K (Potassium)	500
Sc (Scandium)	100
Na (Sodium)	100
Zn (Zinc)	100

Instrument Check Standard #5

ICS-05-1 100 mL
ICS-05-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 15 comps.

Al (Aluminum)	100
As (Arsenic)	100
Cd (Cadmium)	100
Cr (Chromium)	100
Co (Cobalt)	100
Cu (Copper)	100
Fe (Iron)	100
Pb (Lead)	100
Mg (Magnesium)	100
Mn (Manganese)	100
Ni (Nickel)	100
K (Potassium)	100
Na (Sodium)	100
Y (Yttrium)	600
Zn (Zinc)	100

Instrument Check Standard #6

ICS-06-1 100 mL
ICS-06-5 500 mL
50 µg/mL each in 2% HNO₃ 9 comps.

Al (Aluminum)	Pb (Lead)
As (Arsenic)	P (Phosphorus)
Cr (Chromium)	K (Potassium)
Co (Cobalt)	Na (Sodium)
Cu (Copper)	

Instrument Check Standard #7

ICS-07-1 100 mL
ICS-07-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 7 comps.

Al (Aluminum)	50
Ba (Barium)	50
Cd (Cadmium)	50
Cu (Copper)	50
Mn (Manganese)	50
K (Potassium)	500
Zn (Zinc)	50

Screening Standards

These three Qualitative Standards can be combined to scan samples quickly and easily for elements present. They should be used for element identification only. The concentration of each element is approximately 10 µg/mL. To screen for all 68 elements these 3 semi-quantitative standards can be blended together and used immediately.

Semi-Quantitative Standard #1

SQS-01-1 1 x 100 mL
10 µg/mL each in 2-5% HNO₃ tr. HF 33 comps.

Al (Aluminum)	Na (Sodium)
As (Arsenic)	Nd (Neodymium)
Ba (Barium)	P (Phosphorus)
Bi (Bismuth)	Pb (Lead)
Ca (Calcium)	Pr (Praseodymium)
Cd (Cadmium)	Sc (Scandium)
Ce (Cerium)	Se (Selenium)
Dy (Dysprosium)	Sm (Samarium)
Er (Erbium)	Sr (Strontium)
Eu (Europium)	Tb (Terbium)
Ga (Gallium)	Th (Thorium)
Gd (Gadolinium)	Tl (Thallium)
Ho (Holmium)	Tm (Thulium)
In (Indium)	U (Uranium)
La (Lanthanum)	Y (Yttrium)
Lu (Lutetium)	Yb (Ytterbium)
Mg (Magnesium)	

Semi-Quantitative Standard #2

SQS-02-1 1 x 100 mL
10 µg/mL each in 2-5% HNO₃ tr. HCl tr. HF 34 comps.

Au (Gold)	Pd (Palladium)
B (Boron)	Pt (Platinum)
Be (Beryllium)	Rb (Rubidium)
Co (Cobalt)	Re (Rhenium)
Cr (Chromium)	Rh (Rhodium)
Cs (Cesium)	Ru (Ruthenium)
Cu (Copper)	S (Sulfur)
Fe (Iron)	Sb (Antimony)
Ge (Germanium)	Si (Silicon)
Hf (Hafnium)	Sn (Tin)
Ir (Iridium)	Ta (Tantalum)
K (Potassium)	Te (Tellurium)
Li (Lithium)	Ti (Titanium)
Mn (Manganese)	V (Vanadium)
Mo (Molybdenum)	W (Tungsten)
Nb (Niobium)	Zn (Zinc)
Ni (Nickel)	Zr (Zirconium)

Semi-Quantitative Standard #3

SQS-03-1 1 x 100 mL
10 µg/mL each in 2-5% HNO₃ 2 comps.

Hg (Mercury)	Ag (Silver)
--------------	-------------

- 68 Elements in Three Mixes
- Economical ■ Easy to Use

Screening Standard Set

SQS-1-SET 3 x 100 mL
SQS-01-1 SQS-02-1 SQS-03-1

Technical Note

To verify screening results, use single element standards to confirm and quantify the concentration.



ICP

SDWA (Safe Drinking Water Act) Standards

SDWA Standards

For use in SW-846, Method 1310 and U.S. NPDRW 40CFR Part 141. The three Drinking Water Standards are used for monitoring drinking water and/or ground and surface water.

Primary Drinking Water Metals

SDWA-01-1 100 mL
SDWA-01-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 7 comps.

As (Arsenic)	10
Ba (Barium)	100
Cd (Cadmium)	5
Cr (Chromium)	10
Pb (Lead)	10
Se (Selenium)	5
Ag (Silver)	10

Secondary Drinking Water Metals

SDWA-02-1 100 mL
SDWA-02-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 4 comps.

Cu (Copper)	100
Fe (Iron)	30
Mn (Manganese)	5
Zn (Zinc)	500

Mercury Solution

SDWA-03-1 100 mL
SDWA-03-5 500 mL
10 µg/mL in 5% HNO₃

Hg (Mercury)

Drinking Water Sets

SDWA-1-SET 3 x 100 mL
SDWA-01-1 SDWA-02-1 SDWA-03-1

SDWA-5-SET 3 x 500 mL
SDWA-01-5 SDWA-02-5 SDWA-03-5

Standards for Analytes covered in the Safe Drinking Water Act (SDWA)

Primary Metals for Analysis by ICP

Contains all approved elements

SDWA-04-1 100 mL
SDWA-04-5 500 mL
At stated conc. (µg/mL) in 2-5% HNO₃ 9 comps.

As (Arsenic)	100
Ba (Barium)	10
Be (Beryllium)	10
Cd (Cadmium)	10
Ca (Calcium)	100
Cr (Chromium)	10
Cu (Copper)	10
Ni (Nickel)	10
Na (Sodium)	100

Primary Metals for Analysis by ICP-MS

Contains all approved elements

SDWA-06-MS-1 100 mL
SDWA-06-MS-5 500 mL
10 µg/mL each in 2% HNO₃ 11 comps.

Sb (Antimony)	Cu (Copper)
As (Arsenic)	Pb (Lead)
Ba (Barium)	Ni (Nickel)
Be (Beryllium)	Se (Selenium)
Cd (Cadmium)	Tl (Thallium)
Cr (Chromium)	

Primary Metals for Analysis by GFAA/ICP/ICP-MS

SDWA-07-1 100 mL
SDWA-07-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ tr. HF 14 comps.

Sb (Antimony)	100
As (Arsenic)	100
Ba (Barium)	10
Be (Beryllium)	10
Cd (Cadmium)	10
Ca (Calcium)	100
Cr (Chromium)	10
Cu (Copper)	10
Pb (Lead)	10
Ni (Nickel)	10
Se (Selenium)	10
Si (Silicon) †	100
Na (Sodium)	100
Tl (Thallium)	10

† 214 µg/mL as SiO₂

Secondary Metals for Analysis by GFAA/ICP/ICP-MS

SDWA-08-1 100 mL
SDWA-08-5 500 mL
At stated conc. (µg/mL) in 2-5% HNO₃ 5 comps.

Al (Aluminum)	10
Fe (Iron)	100
Mn (Manganese)	10
Ag (Silver)	10
Zn (Zinc)	10

Primary & Secondary Metals for Analysis by GFAA/ICP/ICP-MS

Contains all Primary & Secondary Metals

SDWA-09-1 100 mL
SDWA-09-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 19 comps.

Al (Aluminum)	10
Sb (Antimony)	100
As (Arsenic)	100
Ba (Barium)	10
Be (Beryllium)	10
Cd (Cadmium)	10
Ca (Calcium)	100
Cr (Chromium)	10
Cu (Copper)	10
Fe (Iron)	100
Pb (Lead)	10
Mn (Manganese)	10
Ni (Nickel)	10
Se (Selenium)	10
Si (Silicon) †	100
Ag (Silver)	10
Na (Sodium)	100
Tl (Thallium)	10
Zn (Zinc)	10

† 214 µg/mL as SiO₂

Inorganic products containing acid generally require a hazardous shipping fee.

Inorganic products in water generally do not.



Groundwater & Wastewater Standards

Trace Metals I, II, III

Trace Metals I

WPTM-01-1 100 mL
WPTM-01-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 15 comps.

Al (Aluminum)	500
As (Arsenic)	100
Be (Beryllium)	100
Cd (Cadmium)	25
Cr (Chromium)	100
Co (Cobalt)	100
Cu (Copper)	100
Fe (Iron)	100
Pb (Lead)	100
Mn (Manganese)	100
Hg (Mercury)	5
Ni (Nickel)	100
Se (Selenium)	25
V (Vanadium)	250
Zn (Zinc)	100

Trace Metals II

WPTM-02-1 100 mL
WPTM-02-5 500 mL
 At stated conc. (µg/mL) in 2% HNO₃ 3 comps.

Sb (Antimony)	20
Ag (Silver)	10
Tl (Thallium)	20

Trace Metal Sets

WPTM-1-SET	3 x 100 mL
WPTM-01-1	WPTM-03-1
WPTM-02-1	
WPTM-5-SET	3 x 500 mL
WPTM-01-5	WPTM-03-5
WPTM-02-5	

Trace Metals III

WPTM-03-1 100 mL
WPTM-03-5 500 mL
 At stated conc. (µg/mL) in 2% HNO₃ tr. HF 6 comps.

Ba (Barium)	500
Ca (Calcium)	500
Mg (Magnesium)	100
Mo (Molybdenum)	500
K (Potassium)	100
Na (Sodium)	500

Alternate Metals for Groundwater and Wastewater Analysis

Alternate Metals I

WPAM-01-1 100 mL
WPAM-01-5 500 mL
 At stated conc. (µg/mL) in 2% HNO₃ 11 comps.

Al (Aluminum)	20
Sb (Antimony)	5
Be (Beryllium)	5
Co (Cobalt)	10
Cu (Copper)	10
Fe (Iron)	20
Mn (Manganese)	10
Ni (Nickel)	10
Tl (Thallium)	5
V (Vanadium)	20
Zn (Zinc)	10

Alternate Metals III

WPAM-03-1 100 mL
WPAM-03-5 500 mL
 At stated conc. (µg/mL) in 2% HNO₃ 4 comps.

Ca (Calcium)	500
Mg (Magnesium)	100
K (Potassium)	100
Na (Sodium)	500

Alternate Trace Metal Sets

WPAM-1-SET	2 x 100 mL
WPAM-01-1	WPAM-03-1
WPAM-5-SET	2 x 500 mL
WPAM-01-5	WPAM-03-5

TCLP Multi-Element Calibration Standards

For use in SW-846, Method 1311 Toxicity Characteristic Leaching Procedure

TCLP Standard #1

TCLP-01-1 100 mL
TCLP-01-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 7 comps.

As (Arsenic)	25
Ba (Barium)	500
Cd (Cadmium)	5
Cr (Chromium)	25
Pb (Lead)	25
Se (Selenium)	5
Ag (Silver)	25

TCLP Standard for ICP

TCLP-ICP-1 100 mL
TCLP-ICP-5 500 mL
 At stated conc. (µg/mL) in 2% HNO₃ 4 comps.

Ba (Barium)	500
Cd (Cadmium)	5
Cr (Chromium)	25
Ag (Silver)	25

TCLP Standard for GFAA

TCLP-GFAA-1 100 mL
TCLP-GFAA-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 3 comps.

As (Arsenic)	25
Pb (Lead)	25
Se (Selenium)	5

TCLP Standard #2

For Mercury Analysis using ICP or Cold Vapor AA

TCLP-02-1 100 mL
TCLP-02-5 500 mL
 20 µg/mL in 5% HNO₃

Hg (Mercury)



ICP

MISA Test Group 29

MISA Test Group 29 Analysis Calibration Standards

For use in MISA Test Group 29 Analysis or general use standards. Set of six standards contains 69 elements at 100 µg/mL each. Ideal for the laboratory that wants to analyze for everything.

MISA Standard 1

Rare Earth Metals

MISA-01-1 100 mL
100 µg/mL each in 5% HNO₃ 18 comps.

Ce (Cerium)	Pr (Praseodymium)
Dy (Dysprosium)	Sc (Scandium)
Er (Erbium)	Sm (Samarium)
Eu (Europium)	Tb (Terbium)
Gd (Gadolinium)	Th (Thorium)
Ho (Holmium)	Tm (Thulium)
La (Lanthanum)	U (Uranium)
Lu (Lutetium)	Yb (Ytterbium)
Nd (Neodymium)	Y (Yttrium)

MISA Standard 2

Precious Metals

MISA-02-1 100 mL
100 µg/mL each in 10% HCl 6 comps.

Au (Gold)	Pt (Platinum)
Ir (Iridium)	Rh (Rhodium)
Pd (Palladium)	Ru (Ruthenium)

MISA Standard 3

Tellurium

MISA-03-1 100 mL
100 µg/mL in 10% HCl

Te (Tellurium)

MISA Standard 4

Alkali, Alkaline Earth, Non-Transition Group

MISA-04-1 100 mL
100 µg/mL each in 10% HNO₃ 16 comps.

Al (Aluminum)	Li (Lithium)
As (Arsenic)	Mg (Magnesium)
Ba (Barium)	K (Potassium)
Be (Beryllium)	Rb (Rubidium)
Bi (Bismuth)	Se (Selenium)
Ca (Calcium)	Na (Sodium)
Cs (Cesium)	Sr (Strontium)
Ga (Gallium)	
In (Indium)	

MISA Calibration Set

MISA-1-SET 6 x 100 mL

MISA-01-1	MISA-03-1	MISA-05-1
MISA-02-1	MISA-04-1	MISA-06-1

MISA Standard 5

Fluoride Soluble Group

MISA-05-1 100 mL
100 µg/mL each in 5% HNO₃ tr. HF 15 comps.

Sb (Antimony)	Si (Silicon)
B (Boron)	S (Sulfur)
Ge (Germanium)	Ta (Tantalum)
Hf (Hafnium)	Sn (Tin)
Mo (Molybdenum)	Ti (Titanium)
Nb (Niobium)	W (Tungsten)
P (Phosphorus)	Zr (Zirconium)
Re (Rhenium)	

MISA Standard 6

Transition Metals

MISA-06-1 100 mL
100 µg/mL each in 10% HNO₃ 13 comps.

Cd (Cadmium)	Hg (Mercury)
Co (Cobalt)	Ni (Nickel)
Cu (Copper)	Ag (Silver)
Cr (Chromium)	Tl (Thallium)
Fe (Iron)	V (Vanadium)
Pb (Lead)	Zn (Zinc)
Mn (Manganese)	

Calibration and Matrix Blanks

Nitric Acid Blank

CLP-BLN-5 500 mL
CLP-BLN-L-VAP 1L
(2 x 500 mL)

5% HNO₃ in 18 Megohm ASTM Type I deionized Water

Mixed Acid Blank

CLP-BLMA-5 500 mL
CLP-BLMA-L-VAP 1L
(2 x 500 mL)

5% HCl + 1% HNO₃ in 18 Megohm ASTM Type I deionized Water

Hydrochloric Acid Blank

CLP-BLH-5 500 mL
CLP-BLH-L-VAP 1L
(2 x 500 mL)

5% HCl in 18 Megohm ASTM Type I deionized Water

Water Blank

CLP-BLW-5 ▼ 500 mL
CLP-BLW-L-VAP ▼ 1L
(2 x 500 mL)

18 Megohm ASTM Type I deionized Water



▼ Hazardous fee not required.



Calibration Check Standards

Calibration Standard #1

CLP-CAL-01-1 100 mL
5000 µg/mL each in 5% HNO₃ 4 comps.

Ca (Calcium)	K (Potassium)
Mg (Magnesium)	Na (Sodium)

Calibration Standard #2

CLP-CAL-02-1 100 mL
At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

Cr (Chromium)	100
Mn (Manganese)	150
Ni (Nickel)	400
Ag (Silver)	100
Zn (Zinc)	200

CLP Calibration Standard Set

CLP-CAL-1-SET 7 x 100 mL

CLP-CAL-01	CLP-CAL-04	CLP-CAL-06
CLP-CAL-02	CLP-CAL-05	CLP-CAL-07
CLP-CAL-03		

Calibration Standard #3

CLP-CAL-03-1 100 mL
At stated conc. (µg/mL) in 5% HNO₃ 7 comps.

Al (Aluminum)	2000
Ba (Barium)	2000
Be (Beryllium)	50
Co (Cobalt)	500
Cu (Copper)	250
Fe (Iron)	1000
V (Vanadium)	500

Calibration Standard #4

CLP-CAL-04-1 100 mL
At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

As (Arsenic)	100
Cd (Cadmium)	50
Pb (Lead)	50
Se (Selenium)	50
Tl (Thallium)	100

Calibration Standard #5

CLP-CAL-05-1 100 mL
600 µg/mL in 2% HNO₃

Sb (Antimony)

Calibration Standard #6

CLP-CAL-06-1 100 mL
100 µg/mL in 5% HNO₃

Hg (Mercury)

Calibration Standard #7

CLP-CAL-07-1 100 mL
500 µg/mL each in 5% HNO₃ tr. HF 3 comps.

B (Boron)	Si (Silicon)
Mo (Molybdenum)	





ICP

Contract Laboratory Program (CLP)

Verification Standards

Initial Calibration Verification

CLP-ICV-01-1 100 mL
CLP-ICV-01-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 22 comps.

Al (Aluminum)	200
Sb (Antimony)	60
As (Arsenic)	10
Ba (Barium)	200
Be (Beryllium)	5
Cd (Cadmium)	5
Ca (Calcium)	5000
Cr (Chromium)	10
Co (Cobalt)	50
Cu (Copper)	25
Fe (Iron)	100
Pb (Lead)	5
Mg (Magnesium)	5000
Mn (Manganese)	15
Ni (Nickel)	40
K (Potassium)	5000
Se (Selenium)	5
Ag (Silver)	10
Na (Sodium)	5000
Tl (Thallium)	10
V (Vanadium)	50
Zn (Zinc)	20

Initial Calibration Verification

CLP-ICV-01-R-1 100 mL
CLP-ICV-01-R-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 22 comps.

Al (Aluminum)	200
Sb (Antimony)	60
As (Arsenic)	10
Ba (Barium)	200
Be (Beryllium)	5
Cd (Cadmium)	5
Ca (Calcium)	500
Cr (Chromium)	10
Co (Cobalt)	50
Cu (Copper)	25
Fe (Iron)	100
Pb (Lead)	5
Mg (Magnesium)	500
Mn (Manganese)	15
Ni (Nickel)	40
K (Potassium)	500
Se (Selenium)	5
Ag (Silver)	10
Na (Sodium)	500
Tl (Thallium)	10
V (Vanadium)	50
Zn (Zinc)	20

Continuing Calibration Verification

(Meets CLP QA Second Source Requirements)
CLP-CCV-01-1 100 mL
CLP-CCV-01-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 16 comps.

Al (Aluminum)	1000
Ba (Barium)	1000
Be (Beryllium)	25
Ca (Calcium)	2500
Cr (Chromium)	50
Co (Cobalt)	250
Cu (Copper)	125
Fe (Iron)	500
Mg (Magnesium)	2500
Mn (Manganese)	75
Ni (Nickel)	200
K (Potassium)	2500
Ag (Silver)	50
Na (Sodium)	2500
V (Vanadium)	250
Zn (Zinc)	100

CLP-CCV-02-1 100 mL
CLP-CCV-02-5 500 mL

At stated conc. (µg/mL)s in 5% HNO₃ 5 comps.

As (Arsenic)	50
Cd (Cadmium)	25
Pb (Lead)	25
Se (Selenium)	25
Tl (Thallium)	50

CLP-CCV-03-1 100 mL
CLP-CCV-03-5 500 mL

300 µg/mL in 2% HNO₃

Sb (Antimony)

Technical Note

CLP-ICV-01-R has Ca, Mg, K & Na at 1/10 the concentration of CLP-ICV-01. This improves plasma robustness and often results in superior recoveries.

Initial & Continuing Calibration Verification Sets

CLP-VER-1-SET 4 x 100 mL

CLP-ICV-01-1 CLP-CCV-02-1
CLP-CCV-01-1 CLP-CCV-03-1

CLP-VER-5-SET 4 x 500 mL

CLP-ICV-01-5 CLP-CCV-02-5
CLP-CCV-01-5 CLP-CCV-03-5

Initial & Continuing Calibration Revised Verification Sets

CLP-VER-R-1-SET 4 x 100 mL

CLP-ICV-01-R-1 CLP-CCV-02-1
CLP-CCV-01-1 CLP-CCV-03-1

CLP-VER-R-5-SET 4 x 500 mL

CLP-ICV-01-R-5 CLP-CCV-02-5
CLP-CCV-01-5 CLP-CCV-03-5

Continuing Calibration Verification Standard Sets

CLP-CCV-1-SET 3 x 100 mL

CLP-CCV-01-1 CLP-CCV-03-1
CLP-CCV-02-1

CLP-CCV-5-SET 3 x 500 mL

CLP-CCV-01-5 CLP-CCV-03-5
CLP-CCV-02-5

Spiking Standards

Spiking Solution

CLP-SPS-01-1 100 mL
At stated conc. (µg/mL) in 5% HNO₃ 18 comps.

Al (Aluminum)	200
Sb (Antimony)	50
As (Arsenic)	200
Ba (Barium)	200
Be (Beryllium)	5
Cd (Cadmium)	5
Cr (Chromium)	20
Co (Cobalt)	50
Cu (Copper)	25
Fe (Iron)	100
Pb (Lead)	50
Mn (Manganese)	50
Ni (Nickel)	50
Se (Selenium)	200
Ag (Silver)	5
Tl (Thallium)	200
V (Vanadium)	50
Zn (Zinc)	50

CLP Soil Spiking Solution

CLP-SPS-02-1 100 mL
At stated conc. (µg/mL) in 5% HNO₃ 16 comps.

Sb (Antimony)	100
As (Arsenic)	400
Ba (Barium)	400
Be (Beryllium)	10
Cd (Cadmium)	10
Cr (Chromium)	40
Co (Cobalt)	100
Cu (Copper)	50
Pb (Lead)	100
Mn (Manganese)	100
Ni (Nickel)	100
Se (Selenium)	400
Ag (Silver)	10
Tl (Thallium)	400
V (Vanadium)	100
Zn (Zinc)	100

Technical Note

Spiking solution CLP-SPS-01 can be used for both aqueous and solid samples. An additional spiking solution for soil as outlined in CLP SOW ILM03.0 is also available, CLP-SPS-02.

CLP Spiking Set

CLP-SPS-1-SET 2 x 100 mL

CLP-SPS-01-1 CLP-SPS-02-1



Interference Check & Analyte Standards

The common interferents checked for CLP requirements and their associated analytes are listed in our primary interferent analyte solutions. Occasionally, additional interferents may cause other analytical problems according to CLP SOW ILM03.0. These additional six elements are available with their respective analytes in the alternate interferent/analyte solutions.

Primary Analytes

CLP-PAN-01-1 100 mL
CLP-PAN-01-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 12 comps.

Ag (Silver)	100
Ba (Barium)	50
Be (Beryllium)	50
Cd (Cadmium)	100
Co (Cobalt)	50
Cr (Chromium)	50
Cu (Copper)	50
Mn (Manganese)	50
Ni (Nickel)	100
Pb (Lead)	100
V (Vanadium)	50
Zn (Zinc)	100

Alternate Interferents

CLP-PIN-02-1 100 mL
CLP-PIN-02-5 500 mL
 1000 µg/mL each in 5% HNO₃ 6 comps.

Cr (Chromium)	Ni (Nickel)
Cu (Copper)	Ti (Titanium)
Mn (Manganese)	V (Vanadium)

Alternate Analytes

CLP-PAN-02-1 100 mL
CLP-PAN-02-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF 12 comps.

Al (Aluminum)	100
Sb (Antimony)	100
As (Arsenic)	100
B (Boron)	100
Ca (Calcium)	10
Fe (Iron)	10
Mg (Magnesium)	10
Mo (Molybdenum)	100
Se (Selenium)	100
Si (Silicon)	10
Na (Sodium)	100
Tl (Thallium)	100

Interferent / Analyte Sets

CLP-IA-1-SET 4 x 100 mL

CLP-PIN-01-1	CLP-PIN-02-1
CLP-PAN-01-1	CLP-PAN-02-1

CLP-IA-5-SET 4 x 500 mL

CLP-PIN-01-5	CLP-PIN-02-5
CLP-PAN-01-5	CLP-PAN-02-5

Primary Interferents

CLP-PIN-01-1 100 mL
CLP-PIN-01-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 4 comps.

Al (Aluminum)	5000
Ca (Calcium)	5000
Fe (Iron)	2000
Mg (Magnesium)	5000

Detection Limit Standards

Contract Required Detection Limits (CRDL) Standard Solutions. We offer the flexibility of two convenient solutions:

CLP Detection Limits Standard #1

CLP-CRDL-01-1 100 mL
 At stated conc. (µg/mL) in 5% HNO₃ 15 comps.

Sb (Antimony)	120
As (Arsenic)	120
Be (Beryllium)	10
Cd (Cadmium)	10
Cr (Chromium)	20
Co (Cobalt)	100
Cu (Copper)	50
Pb (Lead)	120
Mn (Manganese)	30
Ni (Nickel)	80
Se (Selenium)	120
Ag (Silver)	20
Tl (Thallium)	120
V (Vanadium)	100
Zn (Zinc)	40

Contract Required Detection Limits (CRDL) Set

CLP-CRDL-1-SET 2 x 100 mL

CLP-CRDL-01	CLP-CRDL-02
-------------	-------------

Technical Note

These standards are prepared to meet the requirements of the CLP protocol; Arsenic (As), Lead (Pb), Selenium (Se) and Thallium (Tl) are at a concentration two times the instrument detection limit (IDL) while the remaining elements are at two times the CRDL.

CLP Detection Limits Standard #2

CLP-CRDL-02-1 100 mL
 At stated conc. (µg/mL) in 5% HNO₃ 15 comps.

Sb (Antimony)	120
As (Arsenic)	20
Be (Beryllium)	10
Cd (Cadmium)	10
Cr (Chromium)	20
Co (Cobalt)	100
Cu (Copper)	50
Pb (Lead)	6
Mn (Manganese)	30
Ni (Nickel)	80
Se (Selenium)	10
Ag (Silver)	20
Tl (Thallium)	20
V (Vanadium)	100
Zn (Zinc)	40

Technical Note

These standards are designed for ICPs equipped with ultrasonic nebulizers and offer the elements of interest at two times the CRDL level.



ICP

EPA Method 200.7

SW-846, Method 200.7 & 6010 Calibration Standards

Mixed Calibration Standard #1

MCS-01-1 100 mL
MCS-01-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 6 comps.

Be (Beryllium)	50
Cd (Cadmium)	150
Pb (Lead)	500
Mn (Manganese)	100
Se (Selenium)	200
Zn (Zinc)	150

Mixed Calibration Standard #2

MCS-02-1 100 mL
MCS-02-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 5 comps.

Ba (Barium)	100
Co (Cobalt)	100
Cu (Copper)	100
Fe (Iron)	10,000
V (Vanadium)	100

Mixed Calibration Standard #3R

MCS-03R-1 100 mL
MCS-03R-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ tr. HF 2 comps.

As (Arsenic)	500
Mo (Molybdenum)	100

Mixed Calibration Standard #4R

MCS-04R-1 100 mL
MCS-04R-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 8 comps.

Al (Aluminum)	200
Ca (Calcium)	1000
Cr (Chromium)	20
Li (Lithium)	100
Ni (Nickel)	20
K (Potassium)	400
Na (Sodium)	200
Sr (Strontium)	10

Mixed Calibration Standard #5R

MCS-05R-1 100 mL
MCS-05R-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 4 comps.

Sb (Antimony)	200
Mg (Magnesium)	1000
Ag (Silver)	50
Tl (Thallium)	200

Mixed Calibration Standard 6R

MCS-06R-1 100 mL
MCS-06R-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃, tr. HF 5 comps.

P (Phosphorus)	200
Sn (Tin)	200
Ti (Titanium)	100
B (Boron)	50
Si (Silicon) †	100

† 214 µg/mL as SiO₂

Complete 6010B, Rev. 2, 1996 Calibration Sets

MCS-1996-1-SET 7 x 100 mL
MCS-01-1 MCS-04R-1 MCS-06R-1
MCS-02-1 MCS-05R-1 TCLP-02-1
MCS-03R-1

MCS-1996-5-SET 7 x 500 mL
MCS-01-5 MCS-04R-5 MCS-06R-5
MCS-02-5 MCS-05R-5 TCLP-02-5
MCS-03R-5

Technical Note

Additional Analyte Calibration Standards.

The use of this Standard Solution (MCS-06R), plus a Mercury Standard (TCLP-02), completes the analyte list for the 1996 Revision 2.

Mercury Standard

Mercury is available in a separate solution due to its incompatibility with other elements.

TCLP-02-1 100 mL
TCLP-02-5 500 mL
20 µg/mL in 5% HNO₃

Hg (Mercury)

Method 200.7 (Revision 4.4, May 1994) Calibration Standards

Mixed Calibration Standard #1

M-200.7-01-1 100 mL
M-200.7-01-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 10 comps.

Sb (Antimony)	50
As (Arsenic)	100
Ba (Barium)	10
B (Boron)	20
Cd (Cadmium)	20
Ca (Calcium)	100
Cu (Copper)	20
Mn (Manganese)	20
Se (Selenium)	50
Ag (Silver)	5

Mixed Calibration Standard #2

M-200.7-02R-1 100 mL
M-200.7-02R-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ tr. HF 6 comps.

Li (Lithium)	50
Mo (Molybdenum)	100
K (Potassium)	200
Na (Sodium)	100
Sr (Strontium)	10
Ti (Titanium)	100

Mixed Calibration Standard #3

M-200.7-03R-1 100 mL
M-200.7-03R-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 4 comps.

Ce (Cerium)	20
Co (Cobalt)	20
P (Phosphorus)	100
V (Vanadium)	20

Mixed Calibration Standard #4

M-200.7-04-1 100 mL
M-200.7-04-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ tr. HF 5 comps.

Al (Aluminum)	100
Cr (Chromium)	50
Si (Silicon) †	100
Sn (Tin)	40
Zn (Zinc)	50

† 214 µg/mL as SiO₂

Mixed Calibration Standard #5

M-200.7-05-1 100 mL
M-200.7-05-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 6 comps.

Be (Beryllium)	10
Fe (Iron)	100
Pb (Lead)	100
Mg (Magnesium)	100
Ni (Nickel)	20
Tl (Thallium)	50

Mixed Calibration Standards Sets

M-200.7-R-1-SET 5 x 100 mL
M-200.7-01-1 M-200.7-04-1
M-200.7-02R-1 M-200.7-05-1
M-200.7-03R-1

M-200.7-5-R-5-SET 5 x 500 mL
M-200.7-01-5 M-200.7-04-5
M-200.7-02R-5 M-200.7-05-5
M-200.7-03-5R



Method 200.7 Instrument Performance Standards

Instrument Performance Check Standard #1

M-200.7-IPC-01-1 100 mL
M-200.7-IPC-01-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 26 comps.

Al (Aluminum)	20	Li (Lithium)	20
As (Arsenic)	20	Mg (Magnesium)	20
Ba (Barium)	20	Mn (Manganese)	20
Be (Beryllium)	20	Ni (Nickel)	20
B (Boron)	20	P (Phosphorus)	100
Cd (Cadmium)	20	K (Potassium)	100
Ca (Calcium)	20	Se (Selenium)	20
Ce (Cerium)	20	Ag (Silver)	2.5
Cr (Chromium)	20	Na (Sodium)	20
Co (Cobalt)	20	Sr (Strontium)	20
Cu (Copper)	20	Tl (Thallium)	20
Fe (Iron)	20	V (Vanadium)	20
Pb (Lead)	20	Zn (Zinc)	20

Instrument Performance Check Standard #2

M-200.7-IPC-02-1 100 mL
M-200.7-IPC-02-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF 5 comps.

Sb (Antimony)	20
Mo (Molybdenum)	20
Si (Silicon) †	100
Sn (Tin)	20
Ti (Titanium)	20

† 214 µg/mL as SiO₂

Method 200.7 Performance Check, Fortifying Solution & Mercury Standard

Laboratory Performance Check Standard

For use in demonstrating the initial and continuing verification of the calibration curves produced by this method.

LPSC-01-1 100 mL
LPSC-01-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF 29 comps.

Al (Aluminum)	20
Sb (Antimony)	20
As (Arsenic)	20
Ba (Barium)	20
Be (Beryllium)	20
B (Boron)	20
Cd (Cadmium)	20
Ca (Calcium)	20
Cr (Chromium)	20
Co (Cobalt)	20
Cu (Copper)	20
Fe (Iron)	20
Pb (Lead)	20
Li (Lithium)	20
Mg (Magnesium)	20
Mn (Manganese)	20
Mo (Molybdenum)	20
Ni (Nickel)	20
P (Phosphorus)	100
K (Potassium)	100
Se (Selenium)	20
Si (Silicon) †	100
Ag (Silver)	5
Na (Sodium)	20
Sr (Strontium)	20
Tl (Thallium)	20
Sn (Tin)	20
V (Vanadium)	20
Zn (Zinc)	20

† 214 µg/mL as SiO₂

Laboratory Fortifying Stock Solution

For use in preparing the laboratory fortified blank and the laboratory fortified sample matrix.

LFSS-01-1 100 mL
LFSS-01-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF 25 comps.

Al (Aluminum)	25
Sb (Antimony)	25
As (Arsenic)	25
Ba (Barium)	25
Be (Beryllium)	5
B (Boron)	25
Cd (Cadmium)	10
Cr (Chromium)	25
Co (Cobalt)	10
Cu (Copper)	25
Fe (Iron)	25
Pb (Lead)	25
Li (Lithium)	25
Mn (Manganese)	25
Mo (Molybdenum)	10
Ni (Nickel)	25
P (Phosphorus)	50
Se (Selenium)	25
Si (Silicon) †	25
Ag (Silver)	2.5
Sr (Strontium)	25
Tl (Thallium)	25
Sn (Tin)	10
V (Vanadium)	10
Zn (Zinc)	25

† 53.5 µg/mL as SiO₂

Technical Note

The analytes Ca, K, Mg, and Na are not included in the stock solution because their concentrations vary widely in environmental samples.

Mercury Standard

Mercury is available in a separate solution due to incompatibility with other elements.

TCLP-02-1 100 mL
TCLP-02-5 500 mL
 20 µg/mL in 5% HNO₃
 Hg (Mercury)

Inorganic products containing acid generally require a hazardous shipping fee.
 Inorganic products in water generally do not.



ICP

EPA Method 200.7

Method 200.7 Fortifying (Spiking & Instrument Performance Standards)

These Standards have been split for stability and ease of use. Choose the Instrument Fortifying Standard you require for Part 1 & use with Part 2 to complete the analyte list. For Lab Fortified Blank use M-200.7-LFSS-01, for Water Samples use M-200.7-LFSS-01W, and for Solid Samples use M-200.7-LFSS-01S.

Part 1

Instrument Fortifying Standard

M-200.7-LFSS-01-1 100 mL
M-200.7-LFSS-01-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 26 comps.

Al (Aluminum)	20
As (Arsenic)	20
Ba (Barium)	20
Be (Beryllium)	20
B (Boron)	20
Cd (Cadmium)	20
Ca (Calcium)	20
Ce (Cerium)	20
Cr (Chromium)	20
Co (Cobalt)	20
Cu (Copper)	20
Fe (Iron)	20
Pb (Lead)	20
Li (Lithium)	20
Mg (Magnesium)	20
Mn (Manganese)	20
Ni (Nickel)	20
P (Phosphorus)	20
K (Potassium)	500
Se (Selenium)	20
Ag (Silver)	7.5
Na (Sodium)	20
Sr (Strontium)	20
Tl (Thallium)	20
V (Vanadium)	20
Zn (Zinc)	20

Instrument Fortifying Standard for Water

M-200.7-LFSS-01W-1 100 mL
M-200.7-LFSS-01W-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 22 comps.

Al (Aluminum)	20
As (Arsenic)	20
Ba (Barium)	20
Be (Beryllium)	20
B (Boron)	20
Cd (Cadmium)	20
Ce (Cerium)	20
Cr (Chromium)	20
Co (Cobalt)	20
Cu (Copper)	20
Fe (Iron)	20
Pb (Lead)	20
Li (Lithium)	20
Mn (Manganese)	20
Ni (Nickel)	20
P (Phosphorus)	20
K (Potassium)	500
Se (Selenium)	20
Ag (Silver)	7.5
Tl (Thallium)	20
V (Vanadium)	20
Zn (Zinc)	20

Instrument Fortifying Standard for Solids

M-200.7-LFSS-01S-1 100 mL
M-200.7-LFSS-01S-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 24 comps.

As (Arsenic)	20
Ba (Barium)	20
Be (Beryllium)	20
B (Boron)	20
Cd (Cadmium)	20
Ca (Calcium)	20
Ce (Cerium)	20
Cr (Chromium)	20
Co (Cobalt)	20
Cu (Copper)	20
Pb (Lead)	20
Li (Lithium)	20
Mg (Magnesium)	20
Mn (Manganese)	20
Ni (Nickel)	20
P (Phosphorus)	20
K (Potassium)	500
Se (Selenium)	20
Ag (Silver)	7.5
Na (Sodium)	20
Sr (Strontium)	20
Tl (Thallium)	20
V (Vanadium)	20
Zn (Zinc)	20

Part 2

Instrument Fortifying Standard #2

M-200.7-LFSS-02-1 100 mL
M-200.7-LFSS-02-5 500 mL
20 µg/mL each in 5% HNO₃ tr. HF 5 comps.

Sb (Antimony)	Sn (Tin)
Mo (Molybdenum)	Ti (Titanium)
Si (Silicon) †	† 42.78 µg/mL as SiO ₂





Method 200.7 Spiking Solutions for Drinking Water

Spiking Standard #1R

M-200.7-SP-01-R ▼ 50 mL
At stated conc. (µg/mL) in H₂O tr. HF 4 comps.

B (Boron)	400
Mo (Molybdenum)	200
Si (Silicon) †	2000
P (Phosphorus)	400

† 4278 µg/mL SiO₂

Spiking Standard #2R

M-200.7-SP-02-R 50 mL
M-200.7-SP-02-R-1 100 mL
M-200.7-SP-02-R-5 500 mL
10,000 µg/mL each in 2% HNO₃ 4 comps.

Ca (Calcium)	K (Potassium)
Mg (Magnesium)	Na (Sodium)

Spiking Standard #3

M-200.7-SP-03 50 mL
At stated conc. (µg/mL) in 5% HNO₃ 12 comps.

Al (Aluminum)	2000
Ba (Barium)	2000
Be (Beryllium)	50
Cr (Chromium)	200
Co (Cobalt)	500
Cu (Copper)	250
Fe (Iron)	1000
Mn (Manganese)	500
Ni (Nickel)	500
Ag (Silver)	50
V (Vanadium)	500
Zn (Zinc)	500

Method 200.7 Spiking Set

M-200.7-SP-R-SET 5 x 50 mL
M-200.7-SP-01-R M-200.7-SP-04-R
M-200.7-SP-02-R M-200.7-SP-05-R
M-200.7-SP-03

Spiking Standard #4R

M-200.7-SP-04-R 50 mL
200 µg/mL in dilute HNO₃
Sb (Antimony)

Spiking Standard #5R

M-200.7-SP-05-R 50 mL
At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

As (Arsenic)	200
Cd (Cadmium)	100
Pb (Lead)	200
Se (Selenium)	400
Tl (Thallium)	400

Method 200.7 Interference Check Standards

For use in testing and verifying the inter-element spectral correction process.

SIC Solution #1

Used to evaluate the spectral interference for the analytes: Al, Sb, Se, Sn, V

SICS-01-1 100 mL
SICS-01-5 500 mL
50 µg/mL in 2% HNO₃ tr. HF

Mo (Molybdenum)

Check Solutions Sets

SIC-1-SET 3 x 100 mL
SICS-01-1 SICS-03-1
SICS-02-1

SIC-5-SET 3 x 500 mL
SICS-01-5 SICS-03-5
SICS-02-5

SIC Solution #2

Used to evaluate the spectral interference for the analytes: Sb, Pb, Zn, Mo, As, Be

SICS-02-1 100 mL
SICS-02-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 5 comps.

Cr (Chromium)	20
Co (Cobalt)	10
Cu (Copper)	40
Mn (Manganese)	20
V (Vanadium)	10

SIC Solution #3

Used to evaluate the spectral interference for the analytes: Sb, Zn, As, Ag, Cr, Mn, V

SICS-03-1 100 mL
SICS-03-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 3 comps.

Al (Aluminum)	30
Fe (Iron)	150
Ni (Nickel)	20

Since everyone experiences different interference problems in their analysis, it is often easiest to design standards to match the "real world" samples. Below is a set of single element standards that can be used for making these standards in your lab.

Spectral Interference Check Set Single Elements

SIC-SING-1-SET 9 x 100 mL
Each at 1,000 µg/mL in HNO₃

Al (Aluminum)	Mn (Manganese)
Cr (Chromium)	Mo (Molybdenum)
Co (Cobalt)	Ni (Nickel)
Cu (Copper)	V (Vanadium)
Fe (Iron)	

▼ Hazardous fee not required.



ICP

EPA Method 6010

Method 6010B (Revision 2, from SW-846) Calibration Standards

Mixed Calibration Standard #1

MCS-01-1 100 mL
MCS-01-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 6 comps.

Be (Beryllium)	50
Cd (Cadmium)	150
Pb (Lead)	500
Mn (Manganese)	100
Se (Selenium)	200
Zn (Zinc)	150

Mixed Calibration Standard #2

MCS-02-1 100 mL
MCS-02-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 5 comps.

Ba (Barium)	100
Co (Cobalt)	100
Cu (Copper)	100
Fe (Iron)	10,000
V (Vanadium)	100

Mixed Calibration Standard #3R

MCS-03R-1 100 mL
MCS-03R-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ tr. HF 2 comps.

As (Arsenic)	500
Mo (Molybdenum)	100

Mixed Calibration Standard #4R

MCS-04R-1 100 mL
MCS-04R-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 8 comps.

Al (Aluminum)	200
Ca (Calcium)	1000
Cr (Chromium)	20
Li (Lithium)	100
Ni (Nickel)	20
K (Potassium)	400
Na (Sodium)	200
Sr (Strontium)	10

Mixed Calibration Standard #5R

MCS-05R-1 100 mL
MCS-05R-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃ 4 comps.

Sb (Antimony)	200
Mg (Magnesium)	1000
Ag (Silver)	50
Tl (Thallium)	200

Mixed Calibration Standard 6R

MCS-06R-1 100 mL
MCS-06R-5 500 mL
At stated conc. (µg/mL) in 2-5% HNO₃, tr. HF 5 comps.

P (Phosphorus)	200
Sn (Tin)	200
Ti (Titanium)	100
B (Boron)	50
Si (Silicon) †	100

† 214 µg/mL as SiO₂

Complete 6010B, Rev. 2, 1996 Calibration Set

MCS-1996-1-SET 7 x 100 mL

MCS-01-1 MCS-04R-1 MCS-06R-1
MCS-02-1 MCS-05R-1 TCLP-02-1
MCS-03R-1

MCS-1996-5-SET 7 x 500 mL

MCS-01-5 MCS-04R-5 MCS-06R-5
MCS-02-5 MCS-05R-5 TCLP-02-5
MCS-03R-5

Technical Note

Additional Analyte Calibration Standards.

The use of this Standard Solution (MCS-06R), plus a Mercury Standard (TCLP-02), completes the analyte list for the 1996 Revision 2.

Mercury Standard

Mercury is available in a separate solution due to its incompatibility with other elements.

TCLP-02-1 100 mL
TCLP-02-5 500 mL

20 µg/mL in 5% HNO₃

Hg (Mercury)

Method 6010B Spiking Standards

Three convenient solutions that can be used for spiking samples pre- or post- digestion.

Spiking Standard #1

QCS-01-1 100 mL
QCS-01-5 500 mL
100 µg/mL each in 5% HNO₃ tr. HF 23 comps.

Sb (Antimony)	Mn (Manganese)
As (Arsenic)	Mo (Molybdenum)
Be (Beryllium)	Ni (Nickel)
Cd (Cadmium)	P (Phosphorus)
Ca (Calcium)	Se (Selenium)
Cr (Chromium)	Sr (Strontium)
Co (Cobalt)	Tl (Thallium)
Cu (Copper)	Sn (Tin)
Fe (Iron)	Ti (Titanium)
Pb (Lead)	V (Vanadium)
Li (Lithium)	Zn (Zinc)
Mg (Magnesium)	

Spiking Standard #2

QCS-02-1 100 mL
QCS-02-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ tr. HF 7 comps.

Al (Aluminum)	100
Ba (Barium)	100
B (Boron)	100
K (Potassium)	1000
Si (Silicon) †	500
Ag (Silver)	50
Na (Sodium)	100

† 1070 µg/mL as SiO₂

QC Standard #2R

QCS-02-R1-1 100 mL
QCS-02-R1-5 500 mL
100 µg/mL each in 5% HNO₃ tr. HF 7 comps.

Al (Aluminum)	Ag (Silver)
Ba (Barium)	Na (Sodium)
B (Boron)	
K (Potassium)	† 214 µg/mL as SiO ₂
Si (Silicon) †	

Mercury Standard

Mercury is available in a separate solution due to incompatibility with other elements.

TCLP-02-1 100 mL
TCLP-02-5 500 mL

20 µg/mL in 5% HNO₃

Hg (Mercury)



Method 6010B (Revision 2 from SW-846, Dec. 1996) Performance and Interference Check Standards

Laboratory Performance Check Standard

LPCS-01R-1 100 mL
LPCS-01R-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃ tr. HF
30 comps.

Al (Aluminum)	20
Sb (Antimony)	20
As (Arsenic)	20
Ba (Barium)	20
Be (Beryllium)	20
B (Boron)	20
Cd (Cadmium)	20
Ca (Calcium)	20
Cr (Chromium)	20
Co (Cobalt)	20
Cu (Copper)	20
Fe (Iron)	20
Pb (Lead)	20
Li (Lithium)	20
Mg (Magnesium)	20
Mn (Manganese)	20
Mo (Molybdenum)	20
Ni (Nickel)	20
P (Phosphorous)	100
K (Potassium)	100
Se (Selenium)	20
Si (Silicon) †	100
Ag (Silver)	5
Na (Sodium)	20
Sr (Strontium)	20
Tl (Thallium)	20
Sn (Tin)	20
Ti (Titanium)	20
V (Vanadium)	20
Zn (Zinc)	20

† 214 µg/mL as SiO₂

Primary Interferents

CLP-PIN-01-1 100 mL
CLP-PIN-01-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 4 comps.

Al (Aluminum)	5000
Ca (Calcium)	5000
Fe (Iron)	2000
Mg (Magnesium)	5000

Alternate Interferents

CLP-PIN-02-1 100 mL
CLP-PIN-02-5 500 mL
1000 µg/mL each in 5% HNO₃ 6 comps.

Cr (Chromium)	Ni (Nickel)
Cu (Copper)	Ti (Titanium)
Mn (Manganese)	V (Vanadium)

Interference Set

SIC-SING-R-1-SET set of 10 x 100 mL
10 individual single element standards at 1,000 µg/mL in HNO₃

Al (Aluminum)	Mg (Magnesium)
Ca (Calcium)	Mn (Manganese)
Cr (Chromium)	Ni (Nickel)
Cu (Copper)	Ti (Titanium)
Fe (Iron)	V (Vanadium)

Technical Note

Because interference problems are dependent on the types of sample matrices encountered - it is often easier to create your own set of matrix matching interference check solutions. Therefore, we are offering a set of single element solutions to be used for that purpose.

If you would like us to formulate an interference check solution to meet your needs, please call our Inorganic Technical Service Department for a custom quotation.

Set-up Solutions

Nebulizer Adjustment Solution

ICP-69N-1 100 mL
1000 µg/mL in HNO₃
Y (Yttrium)

Method 6010 (Revision 0, Sept. 1986) Interference Check Standards

Four standard mixtures are available for interference checks in SW-846, Method 6010 (Rev. 0, Sept. 1986) and Method 200.7.

Interference Check Standard #1

INT-01-1 100 mL
INT-01-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 17 comps.

As (Arsenic)	1000
Ba (Barium)	300
Be (Beryllium)	100
Cd (Cadmium)	300
Cr (Chromium)	300
Co (Cobalt)	300
Cu (Copper)	300
Pb (Lead)	1000
Mn (Manganese)	200
Hg (Mercury)	50
Ni (Nickel)	300
K (Potassium)	20,000
Se (Selenium)	500
Ag (Silver)	300
Tl (Thallium)	1000
V (Vanadium)	300
Zn (Zinc)	300

Interference Check Standard #2

INT-02-1 100 mL
INT-02-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ tr. HF 4 comps.

B (Boron)	500
Mo (Molybdenum)	300
Si (Silicon)	200
Ti (Titanium)	1000

Interference Check Standard #3

INT-03-1 100 mL
INT-03-5 500 mL
500 µg/mL in 2% HNO₃ tr. Tartaric acid

Sb (Antimony)

Interference Check Standard #4

INT-04-1 100 mL
INT-04-5 500 mL
At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

Al (Aluminum)	1200
Ca (Calcium)	6000
Fe (Iron)	5000
Mg (Magnesium)	3000
Na (Sodium)	1000

Interference Check Standards Sets

INT-1986-1-SET 4 x 100 mL

INT-01-1	INT-03-1
INT-02-1	INT-04-1

INT-1986-5-SET 4 x 500 mL

INT-01-5	INT-03-5
INT-02-5	INT-04-5

(20 x 20 inch laminated poster)

Periodic Table of Elements

Tel. 203-786-5290 • www.AccuStandard.com

Acetic acid (1.765 g/mL) (117.11°F)
Acetone (0.791 g/mL) (56.42°F)
Acetonitrile (ACN) (0.786 g/mL) (58.42°F)
Benzoate (1.074 g/mL) (80°F)
2-Butanone (0.806 g/mL) (58°F)
Butyl acetate (0.879 g/mL) (115.11°F)
tert-Butyl methyl ether (MTBE) (0.74 g/mL) (55-56°F)
Carboxy tetrachloride (1.584 g/mL) (76-77°F)
Chloroform (1.489 g/mL) (68°F)
Cyclohexane (0.779 g/mL) (66.7°F)
Cyclopentane (0.751 g/mL) (56°F)
Dichloromethane (1.256 g/mL) (58°F)
N,N-Dimethylformamide (DMF) (0.944 g/mL) (153°F)
1,4-Dioxane (1.036 g/mL) (100-101°F)
Dipropyl ether (0.738 g/mL) (58-59°F)
Ethyl acetate (EC) (0.902 g/mL) (153.11°F)
Ethyl alcohol (EtOH) (0.789 g/mL) (58.9°F)
Ethyl ether (0.706 g/mL) (54.9°F)
n-Heptane (0.684 g/mL) (58°F)
n-Hexane (0.659 g/mL) (58°F)
Isobutene (0.662 g/mL) (58.9°F)
Isopropyl alcohol (IPA) (0.786 g/mL) (82°F)
Methanol (MeOH) (0.791 g/mL) (64.7°F)
Methylene chloride (MCA) (1.250 g/mL) (23.8-48°F)
Methyl sulfoxide (DMSO) (1.193 g/mL) (65.9°F)
n-Pentane (0.626 g/mL) (58°F)
1,2,3-Tetrachloroethane (1.596 g/mL) (147°F)
Tetrahydrofuran (THF) (0.889 g/mL) (54.9°F)
Toluene (0.867 g/mL) (59.9°F)



AccuStandard has received many requests for the following multi-element standards. We offer our own version of these popular mixes offered by Merck. Products are made to the same specifications as other mixes in our product line and subject to the same rigorous quality control.

AccuStandard equivalent of Merck Multi-Element Standards

ICP Multi-Element Standard Solution I

MES-01-1	100 mL
MES-01-5	500 mL
At stated conc. (µg/mL) in 1 mol/L HNO ₃	
19 comps.	
Ag (Silver)	50
Al (Aluminum)	100
B (Boron)	15
Ba (Barium)	5
Be (Beryllium)	1
Bi (Bismuth)	200
Cd (Cadmium)	20
Co (Cobalt)	20
Cr (Chromium)	25
Cu (Copper)	20
Fe (Iron)	15
Ga (Gallium)	150
In (Indium)	200
Mn (Manganese)	5
Ni (Nickel)	50
Pb (Lead)	200
Sr (Strontium)	1
Tl (Thallium)	400
Zn (Zinc)	20

ICP Multi-Element Standard Solution II

MES-02-1	100 mL
MES-02-5	500 mL
At stated conc. (µg/mL) in 1 mol/L HNO ₃	
3 comps.	
Li (Lithium)	250
K (Potassium)	10,000
Na (Sodium)	1000

ICP Multi-Element Standard Solution III

MES-03-1	100 mL
MES-03-5	500 mL
1000 µg/mL each in 1 mol/L HNO ₃	
4 comps.	
Ba (Barium)	Mg (Magnesium)
Ca (Calcium)	Sr (Strontium)

ICP Multi-Element Standard Solution IV

MES-04-1	100 mL
MES-04-5	500 mL
1000 µg/mL each in 1 mol/L HNO ₃	
23 comps.	

Ag (Silver)	In (Indium)
Al (Aluminum)	K (Potassium)
B (Boron)	Li (Lithium)
Ba (Barium)	Mg (Magnesium)
Bi (Bismuth)	Mn (Manganese)
Ca (Calcium)	Na (Sodium)
Cd (Cadmium)	Ni (Nickel)
Co (Cobalt)	Pb (Lead)
Cr (Chromium)	Sr (Strontium)
Cu (Copper)	Tl (Thallium)
Fe (Iron)	Zn (Zinc)
Ga (Gallium)	

ICP Multi-Element Standard Solution V

MES-05-1-SET	2x100 mL
MES-05-5-SET	2x500 mL
At stated conc. (µg/mL) in 2-10% HCl	
26 comps.	
MES-05	
Al (Aluminum)	20
As (Arsenic)	20
B (Boron)	2
Ba (Barium)	2
Be (Beryllium)	1
Ca (Calcium)	10
Cd (Cadmium)	2
Cr (Chromium)	2
Cu (Copper)	2
Fe (Iron)	2
K (Potassium)	100
Li (Lithium)	2
Mg (Magnesium)	1
Mn (Manganese)	1
Na (Sodium)	20
Ni (Nickel)	5
P (Phosphorus)	10
Pb (Lead)	20
Sc (Scandium)	1
Se (Selenium)	20
Sr (Strontium)	1
Te (Tellurium)	20
Ti (Titanium)	2
Y (Yttrium)	1
Zn (Zinc)	2
MES-05-HG	
5% HNO ₃	
Hg (Mercury)	5
Supplied separately for better stability	

ICP Multi-Element Standard Solution VII

MES-07-1 ▼	100 mL
MES-07-5 ▼	500 mL
100 µg/mL each in Water tr. HNO ₃	
9 comps.	

NH ₄ (Ammonium)	Mn (Manganese)
Ba (Barium)	Na (Sodium)
Ca (Calcium)	Sr (Strontium)
K (Potassium)	
Li (Lithium)	
Mg (Magnesium)	

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.

ICP Multi-Element Standard Solution VI for MS

MES-06-1-SET	100 mL
MES-06-5-SET	500 mL
At stated conc. (µg/mL) in 1 mol/L HNO ₃ tr. HF	
30 comps.	
Ag (Silver)	10
Al (Aluminum)	10
As (Arsenic)	100
B (Boron)	100
Ba (Barium)	10
Be (Beryllium)	100
Bi (Bismuth)	10
Ca (Calcium)	1000
Cd (Cadmium)	10
Co (Cobalt)	10
Cr (Chromium)	10
Cu (Copper)	10
Fe (Iron)	100
Ga (Gallium)	10
K (Potassium)	10
Li (Lithium)	10
Mg (Magnesium)	10
Mn (Manganese)	10
Mo (Molybdenum)	10
Na (Sodium)	10
Ni (Nickel)	10
Pb (Lead)	10
Rb (Rubidium)	10
Se (Selenium)	100
Sr (Strontium)	10
Tl (Thallium)	10
U (Uranium)	10
V (Vanadium)	10
Zn (Zinc)	100
MES-06-TE	
Te (Tellurium)	10
supplied separately for better stability in 10% HCl	

ICP Multi-Element Standard Solution VIII

MES-08-1-SET	2x100 mL
MES-08-5-SET	2x500 mL
100 µg/mL each in 1 mol/L HNO ₃	
24 comps.	

MES-08	
Al (Aluminum)	K (Potassium)
B (Boron)	Li (Lithium)
Ba (Barium)	Mg (Magnesium)
Be (Beryllium)	Mn (Manganese)
Bi (Bismuth)	Na (Sodium)
Ca (Calcium)	Ni (Nickel)
Cd (Cadmium)	Pb (Lead)
Co (Cobalt)	Se (Selenium)
Cr (Chromium)	Sr (Strontium)
Cu (Copper)	Tl (Thallium)
Fe (Iron)	Zn (Zinc)
Ga (Gallium)	

MES-08-TE	
10% HCl	
Te (Tellurium)	
supplied separately for better stability	

ICP Multi-Element Standard Solution IX

MES-09-1-SET	2x100 mL
MES-09-5-SET	2x500 mL
100 µg/mL each in 1 mol/L HNO ₃	
8 comps.	

MES-09	
As (Arsenic)	Ni (Nickel)
Be (Beryllium)	Se (Selenium)
Pb (Lead)	Tl (Thallium)
Cr (Chromium)	

MES-09-HG	
Hg (Mercury)	
supplied separately for better stability.	

ICP Multi-Element Standard Solution X

MES-10-1	100 mL
MES-10-5	500 mL
At stated conc. (µg/mL) in 1 mol/L HNO ₃	
23 comps.	

Ca (Calcium)	3500
Mg (Magnesium)	1500
Na (Sodium)	800
K (Potassium)	300
B (Boron)	10
Fe (Iron)	10
Mo (Molybdenum)	10
Sr (Strontium)	10
As (Arsenic)	5
Ba (Barium)	5
Ni (Nickel)	5
V (Vanadium)	5
Zn (Zinc)	5
Mn (Manganese)	3
Co (Cobalt)	2.5
Pb (Lead)	2.5
Be (Beryllium)	2
Cd (Cadmium)	2
Cr (Chromium)	2
Cu (Copper)	2
Bi (Bismuth)	1
Se (Selenium)	1
Tl (Thallium)	1

ICP Multi-Element Standard Solution XI

MES-11-1-SET	2x100 mL
MES-11-5-SET	2x500 mL
At stated conc. (µg/mL) in 1 mol/L HNO ₃	
6 comps.	

MES-11	
Cd (Cadmium)	10
Cr (Chromium)	900
Cu (Copper)	800
Ni (Nickel)	200
Pb (Lead)	900
Zn (Zinc)	2500

MES-11-HG	
Hg (Mercury)	8
supplied separately for better product stability	



AccuStandard equivalent of Merck Multi-Element Standards

ICP Multi-Element

Standard Solution XII

MES-12-1-SET 2x100 mL
MES-12-5-SET 2x500 mL
1000 µg/mL each in 5% HCl tr.
HNO₃ 7 comps.

MES-12-R1

As (Arsenic) Si (Silicon)
Mo (Molybdenum) W (Tungsten)
P (Phosphorus) V (Vanadium)
S (Sulfur)

MES-12-ZR

Zr (Zirconium)
supplied separately for better
product stability

ICP Multi-Element

Standard Solution XIII

MES-13-1-SET 2x100 mL
MES-13-5-SET 2x500 mL
At stated conc. (µg/mL) in 5% HNO₃
15 comps.

MES-13

Al (Aluminum) 500
As (Arsenic) 100
Be (Beryllium) 100
Cd (Cadmium) 25
Co (Cobalt) 100
Cr (Chromium) 100
Cu (Copper) 100
Fe (Iron) 100
Mn (Manganese) 100
Ni (Nickel) 100
Pb (Lead) 100
Se (Selenium) 25
V (Vanadium) 250
Zn (Zinc) 100

MES-13-HG

Hg (Mercury) 5
supplied separately for better
stability

ICP Multi-Element

Standard Solution XIV

MES-14-1 100 mL
MES-14-5 500 mL
At stated conc. (µg/mL) in 2% HCl
tr. HNO₃ 11 comps.

P (Phosphorus) 100
S (Sulfur) 100
K (Potassium) 100
As (Arsenic) 20
La (Lanthanum) 20
Li (Lithium) 20
Mo (Molybdenum) 20
Mn (Manganese) 20
Ni (Nickel) 20
Sc (Scandium) 20
Na (Sodium) 20

ICP Multi-Element

Standard Solution XV

MES-15-1 100 mL
MES-15-5 500 mL
At stated conc. (µg/mL) in 2% HNO₃
8 comps.

Element

Ba (Barium) 1
Ca (Calcium) 1
K (Potassium) 50
La (Lanthanum) 10
Li (Lithium) 10
Mn (Manganese) 10
Na (Sodium) 10
Sr (Strontium) 10

ICP Multi-Element

Standard Solution XVI

MES-16-1 100 mL
MES-16-5 500 mL
100 µg/mL each in 5% HNO₃ tr. HF
21 comps.

Sb (Antimony) Mg (Magnesium)
As (Arsenic) Mn (Manganese)
Be (Beryllium) Mo (Molybdenum)
Cd (Cadmium) Ni (Nickel)
Ca (Calcium) Se (Selenium)
Cr (Chromium) Sr (Strontium)
Co (Cobalt) Tl (Thallium)
Cu (Copper) Ti (Titanium)
Fe (Iron) V (Vanadium)
Pb (Lead) Zn (Zinc)
Li (Lithium)

ICP Multi-Element

Standard Solution XVII

MES-17-1 100 mL
MES-17-5 500 mL
100 µg/mL each in 15% HCl tr.
HNO₃ 7 comps.

Hf (Hafnium) Ta (Tantalum)
Ir (Iridium) Ti (Titanium)
Sb (Antimony) Zr (Zirconium)
Sn (Tin)

ICP Multi-Element

GF AAS

Standard Solution XVIII

MES-18-R1-1 100 mL
MES-18-R1-5 500 mL
At stated conc. (µg/mL) in 5%
HNO₃ 16 comps.

Ag (Silver) 10
Al (Aluminum) 100
As (Arsenic) 100
Ba (Barium) 50
Be (Beryllium) 5
Cd (Cadmium) 5
Co (Cobalt) 50
Cr (Chromium) 20
Cu (Copper) 50
Fe (Iron) 20
Mn (Manganese) 20
Ni (Nickel) 50
Pb (Lead) 100
Sb (Antimony) 100
Se (Selenium) 100
Tl (Thallium) 100

ICP Multi-Element

Standard Solution XIX

for MS
MES-19-1 100 mL
MES-19-5 500 mL
1 µg/mL each in 1% HNO₃
5 comps.

Be (Beryllium) Tl (Thallium)
Co (Cobalt) U (Uranium)
In (Indium)

ICP Multi-Element

Standard Solution XX

for MS
MES-20-1 100 mL
MES-20-5 500 mL
1 µg/mL each in 1% HNO₃ tr. HCl
11 comps.

Mg (Magnesium) Tl (Thallium)
Cu (Copper) Ce (Cerium)
Cd (Cadmium) Ge (Germanium)
Pb (Lead) Tb (Terbium)
Sc (Scandium) Ba (Barium)
Rh (Rhodium)

ICP Multi-Element

Standard Solution XXI

for MS
MES-21-1-SET 2x100 mL
MES-21-5-SET 2x500 mL
10 µg/mL each in 5% HNO₃
30 comps.

MES-21 In (Indium)
Ag (Silver) K (Potassium)
Al (Aluminum) Li (Lithium)
As (Arsenic) Mg (Magnesium)
Ba (Barium) Mn (Manganese)
Be (Beryllium) Na (Sodium)
Bi (Bismuth) Ni (Nickel)
Ca (Calcium) Pb (Lead)
Cd (Cadmium) Rb (Rubidium)
Co (Cobalt) Se (Selenium)
Cr (Chromium) Sr (Strontium)
Cs (Cesium) Tl (Thallium)
Cu (Copper) V (Vanadium)
Fe (Iron) U (Uranium)
Ga (Gallium) Zn (Zinc)

MES-21-HG

Hg (Mercury) 10
supplied separately for better
product stability

ICP Multi-Element

Standard Solution XXII

for MS
MES-22-1 100 mL
MES-22-5 500 mL
2 µg/mL each in 2% HNO₃
5 comps.

Cd (Cadmium) Pb (Lead)
Cu (Copper) Rh (Rhodium)
Mg (Magnesium)

ICP Multi-Element

Standard Solution XXIII

for MS
MES-23-1 100 mL
MES-23-5 500 mL
1 µg/mL each in 5% HNO₃
15 comps.

Ba (Barium) Lu (Lutetium)
B (Boron) Na (Sodium)
Co (Cobalt) Rh (Rhodium)
Fe (Iron) Sc (Scandium)
Ga (Gallium) Tl (Thallium)
In (Indium) U (Uranium)
K (Potassium) Y (Yttrium)
Li (Lithium)

ICP Multi-Element

Standard Solution XXIV

MES-24-1 100 mL
MES-24-5 500 mL
At stated conc. (µg/mL) in 1%
HNO₃ 15 comps.

Al (Aluminum) 50
As (Arsenic) 50
Ba (Barium) 50
Cd (Cadmium) 50
Co (Cobalt) 50
Cr (Chromium) 50
Cu (Copper) 50
K (Potassium) 500
Mn (Manganese) 50
Mo (Molybdenum) 50
Ni (Nickel) 50
Pb (Lead) 50
Se (Selenium) 50
Sr (Strontium) 50
Zn (Zinc) 50

AccuStandard is not affiliated
with the companies and brands
on this page, and the brands and
company names appear for the
purpose of cross reference with
the corresponding AccuStandard
product which is being offered.



AccuStandard offers calibration and testing standards for individual instruments. The Alternate Source Line (ASL) formulations match product from:

■ Agilent / Varian

■ JY (Jobin Yvon)

■ Perkin Elmer (PE)

■ Teledyne (Leeman Labs)

All of these products have been carefully formulated to be used for specific instrument setup and verification.

AccuStandard equivalent of Agilent / Varian Calibration Solutions

ICP-OES Wavelength Calibration Solution

WAVE-CAL-1	100 mL
WAVE-CAL-5	500 mL
WAVE-CAL-10X-1	100 mL
WAVE-CAL-10X-5	500 mL

At stated conc. (µg/mL) in 1% HNO₃ 15 comps.

	CAL	CAL-10X
Al (Aluminum)	5	50
As (Arsenic)	5	50
Ba (Barium)	5	50
Cd (Cadmium)	5	50
Co (Cobalt)	5	50
Cr (Chromium)	5	50
Cu (Copper)	5	50
Mn (Manganese)	5	50
Mo (Molybdenum)	5	50
Ni (Nickel)	5	50
Pb (Lead)	5	50
Se (Selenium)	5	50
Sr (Strontium)	5	50
Zn (Zinc)	5	50
K (Potassium)	50	500

ICP-OES Calibration Solution

WAVE-CAL2-1	100 mL
WAVE-CAL2-5	500 mL
WAVE-CAL2-10X-1	100 mL
WAVE-CAL2-10X-5	500 mL

At stated conc. (µg/mL) in 1% HNO₃ 14 comps.

	CAL2	CAL2-10X
Al (Aluminum)	5	50
As (Arsenic)	5	50
Ba (Barium)	5	50
Cd (Cadmium)	5	50
Co (Cobalt)	5	50
Cr (Chromium)	5	50
Cu (Copper)	5	50
Mo (Molybdenum)	5	50
Ni (Nickel)	5	50
Pb (Lead)	5	50
Sr (Strontium)	5	50
P (Phosphorus)	5	50
Zn (Zinc)	5	50
K (Potassium)	50	500



ICP/MS Tuning Solution

VAR-TUN-ASL-1	100 mL
VAR-TUN-ASL-5	500 mL

10 µg/mL each in 2-5% HNO₃ 8 comps.

Be (Beryllium)	Pb (Lead)
Mg (Magnesium)	Th (Thorium)
Co (Cobalt)	Ba (Barium)
In (Indium)	Ce (Cerium)



AccuStandard equivalent of Agilent / Varian Solutions



7500 Series PA Tuning 1

AG-TUN1-ASL-1	100 mL
AG-TUN1-ASL-5	500 mL

At stated conc. (µg/mL) in 5% HNO₃ 26 comps.

Zn (Zinc)	20
Be (Beryllium)	20
Cd (Cadmium)	20
As (Arsenic)	20
Ni (Nickel)	10
Pb (Lead)	10
Mg (Magnesium)	10
Tl (Thallium)	5
Na (Sodium)	5
Al (Aluminum)	5
U (Uranium)	5
Cu (Copper)	5
Th (Thorium)	5
Ba (Barium)	5
Co (Cobalt)	5
Sr (Strontium)	5
V (Vanadium)	5
Cr (Chromium)	5
Mn (Manganese)	5
Li-6 (Lithium-6)	5
Sc (Scandium)	5
In (Indium)	5
Lu (Lutetium)	5
Bi (Bismuth)	5
Y (Yttrium)	2.5
Yb (Ytterbium)	2.5

7500 Series PA Tuning 2

AG-TUN2-ASL-1	100 mL
AG-TUN2-ASL-5	500 mL

At stated conc. (µg/mL) in 10% HCl, 1% HNO₃ tr. HF 8 comps.

Mo (Molybdenum)	10
Sb (Antimony)	10
Sn (Tin)	10
Ge (Germanium)	10
Ru (Ruthenium)	10
Pd (Palladium)	10
Ti (Titanium)	5
Ir (Iridium)	5

Stock Tuning Solution

AG-TUNSTOCK-ASL-1	100 mL
AG-TUNSTOCK-ASL-5	500 mL

10 µg/mL in 2% HNO₃ 5 comps.

Li (Lithium)
Y (Yttrium)
Ce (Cerium)
Tl (Thallium)
Co (Cobalt)

Stock Tuning Solution

AG-TUNSTOCK1-ASL-1	100 mL
AG-TUNSTOCK1-ASL-5	500 mL

10 µg/mL in 2% HNO₃ 6 comps.

Li (Lithium)
Mg (Magnesium)
Y (Yttrium)
Ce (Cerium)
Tl (Thallium)
Co (Cobalt)

Internal Standard Mix

AG-INTSTD-ASL-1	100 mL
AG-INTSTD-ASL-5	500 mL

100 µg/mL in 10% HNO₃, tr. HCl 8 comps.

Li-6 (Lithium-6)
Sc (Scandium)
Ge (Germanium)
Rh (Rhodium)
In (Indium)
Tb (Terbium)
Lu (Lutetium)
Bi (Bismuth)

Tuning Solution Sets

AG-TUN-ASL-1-SET 2 x 100 mL

AG-TUN1-ASL-1 AG-TUN2-ASL-1

AG-TUN-ASL-5-SET 2 x 500 mL

AG-TUN1-ASL-5 AG-TUN2-ASL-5

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.



AccuStandard equivalent of Agilent

Environmental Calibration Standard

AG-CAL-ASL-1 100 mL
AG-CAL-ASL-5 500 mL
 At stated conc. (µg/mL) in 10% HNO₃ 25 comps.

Ca (Calcium)	1000
Fe (Iron)	1000
K (Potassium)	1000
Mg (Magnesium)	1000
Na (Sodium)	1000
Ag (Silver)	10
Al (Aluminum)	10
As (Arsenic)	10
Ba (Barium)	10
Be (Beryllium)	10
Cd (Cadmium)	10
Co (Cobalt)	10
Cr (Chromium)	10
Cu (Copper)	10
Mn (Manganese)	10
Mo (Molybdenum)	10
Ni (Nickel)	10
Pb (Lead)	10
Sb (Antimony)	10
Se (Selenium)	10
Tl (Thallium)	10
U (Uranium)	10
V (Vanadium)	10
Zn (Zinc)	10
Th (Thorium)	10

Calibration Mix #1

AG-CAL1-ASL-1 100 mL
AG-CAL1-ASL-5 500 mL
 100 µg/mL each in 2% HNO₃ tr.HF 4 comps.

Sb (Antimony)	Sn (Tin)
Mo (Molybdenum)	Tl (Thallium)

Calibration Mix #2

AG-CAL2-ASL-1 100 mL
AG-CAL2-ASL-5 500 mL
 100 µg/mL each in 5% HNO₃ 18 comps.

Ag (Silver)	Mn (Manganese)
Al (Aluminum)	Ni (Nickel)
As (Arsenic)	Pb (Lead)
Ba (Barium)	Se (Selenium)
Be (Beryllium)	Tl (Thallium)
Cd (Cadmium)	Th (Thorium)
Co (Cobalt)	U (Uranium)
Cr (Chromium)	V (Vanadium)
Cu (Copper)	Zn (Zinc)

Calibration Mix Majors

AG-CALMAJOR-ASL-1 100 mL
AG-CALMAJOR-ASL-5 500 mL
 500 µg/mL each in 5% HNO₃ 5 comps.

Ca (Calcium)	Mg (Magnesium)
Fe (Iron)	Na (Sodium)
K (Potassium)	

Interference Check 6020 #1

AG-INTFR-6020-ASL-1 100 mL
AG-INTFR-6020-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF 12 comps.

Cl (Chloride)	20,000
Ca (Calcium)	3000
Fe (Iron)	2500
Na (Sodium)	2500
C (Carbon)	2000
Al (Aluminum)	1000
Mg (Magnesium)	1000
P (Phosphorus)	1000
K (Potassium)	1000
S (Sulfur)	1000
Mo (Molybdenum)	20
Ti (Titanium)	20

Interference Check 6020 #2

AG-INTFR2-6020-ASL-1 / 100 mL
AG-INTFR2-6020-ASL-5 / 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 11 comps.

Cr (Chromium)	20
Co (Cobalt)	20
Cu (Copper)	20
Mn (Manganese)	20
Ni (Nickel)	20
V (Vanadium)	20
As (Arsenic)	10
Cd (Cadmium)	10
Se (Selenium)	10
Zn (Zinc)	10
Ag (Silver)	5

Environmental Internal Standard

AG-INT-ASL-1 100 mL
AG-INT-ASL-5 500 mL
 10 µg/mL each in 5-10% HNO₃ 7 comps.

Bi (Bismuth)	Sc (Scandium)
Ge (Germanium)	Tb (Terbium)
In (Indium)	Y (Yttrium)
Li-6 (Lithium-6)	

ICP Internal Standard

AG-INT2-ASL-1 100 mL
AG-INT2-ASL-5 500 mL
 100 µg/mL each in 5% HNO₃ 6 comps.

Li-6 (Lithium-6)	In (Indium)
Sc (Scandium)	Tb (Terbium)
Y (Yttrium)	Bi (Bismuth)

Multi-Element Calibration Std. #1

AG-MECAL1-ASL-1 100 mL
AG-MECAL1-ASL-5 500 mL
 10 µg/mL each in 5% HNO₃ 17 comps.

Ce (Cerium)	Pr (Praseodymium)
Dy (Dysprosium)	Sc (Scandium)
Er (Erbium)	Sm (Samarium)
Eu (Europium)	Tb (Terbium)
Gd (Gadolinium)	Th (Thorium)
Ho (Holmium)	Tm (Thulium)
La (Lanthanum)	Y (Yttrium)
Lu (Lutetium)	Yb (Ytterbium)
Nd (Neodymium)	

Multi-Element Calibration Std.#2A

AG-MECAL2A-ASL-1 100 mL
AG-MECAL2A-ASL-5 500 mL
 10 µg/mL each in 5% HNO₃ 27 comps.

Ag (Silver)	Li (Lithium)
Al (Aluminum)	Mg (Magnesium)
As (Arsenic)	Mn (Manganese)
Ba (Barium)	Na (Sodium)
Be (Beryllium)	Ni (Nickel)
Ca (Calcium)	Pb (Lead)
Cd (Cadmium)	Rb (Rubidium)
Co (Cobalt)	Se (Selenium)
Cr (Chromium)	Sr (Strontium)
Cs (Cesium)	Tl (Thallium)
Cu (Copper)	U (Uranium)
Fe (Iron)	V (Vanadium)
Ga (Gallium)	Zn (Zinc)
K (Potassium)	

Multi-Element Calibration Std. #3

AG-MECAL3-ASL-1 100 mL
AG-MECAL3-ASL-5 500 mL
 10 µg/mL each in 10% HCl 10 comps.

Au (Gold)	Rh (Rhodium)
Hf (Hafnium)	Ru (Ruthenium)
Ir (Iridium)	Sb (Antimony)
Pd (Palladium)	Sn (Tin)
Pt (Platinum)	Te (Tellurium)

Multi-Element Calibration Std. #4

AG-MECAL4-ASL-1 ▼ 100 mL
AG-MECAL4-ASL-5 ▼ 500 mL
 10 µg/mL each in Water, tr. HF 12 comps.

B (Boron)	S (Sulfur)
Ge (Germanium)	Si (Silicon)
Mo (Molybdenum)	Ta (Tantalum)
Nb (Niobium)	Ti (Titanium)
P (Phosphorus)	W (Tungsten)
Re (Rhenium)	Zr (Zirconium)

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.

▼ Hazardous fee not required.



AccuStandard equivalent of Agilent

Environmental Spike Mix

AG-SPIKE-ASL-R1-1 100 mL
AG-SPIKE-ASL-R1-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF 24 comps.

Ca (Calcium)	1000
Fe (Iron)	1000
K (Potassium)	1000
Mg (Magnesium)	1000
Na (Sodium)	1000
Ag (Silver)	100
Al (Aluminum)	100
As (Arsenic)	100
Ba (Barium)	100
Be (Beryllium)	100
Cd (Cadmium)	100
Co (Cobalt)	100
Cr (Chromium)	100
Cu (Copper)	100
Mn (Manganese)	100
Mo (Molybdenum)	100
Ni (Nickel)	100
Pb (Lead)	100
Sb (Antimony)	100
Se (Selenium)	100
Tl (Thallium)	100
U (Uranium)	100
V (Vanadium)	100
Zn (Zinc)	100

Environmental Initial Calibration Verification

AG-VER1-ASL-R1-1 100 mL
AG-VER1-ASL-R1-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 26 comps.

Ca (Calcium)	1000
Fe (Iron)	1000
K (Potassium)	1000
Mg (Magnesium)	1000
Na (Sodium)	1000
Sr (Strontium)	100
Ag (Silver)	10
Al (Aluminum)	10
As (Arsenic)	10
Ba (Barium)	10
Be (Beryllium)	10
Cd (Cadmium)	10
Co (Cobalt)	10
Cr (Chromium)	10
Cu (Copper)	10
Mn (Manganese)	10
Mo (Molybdenum)	10
Ni (Nickel)	10
Pb (Lead)	10
Sb (Antimony)	10
Se (Selenium)	10
Tl (Thallium)	10
U (Uranium)	10
V (Vanadium)	10
Zn (Zinc)	10
Th (Thorium)	10

AccuStandard equivalent of Jobin Yvon (JY)

Instrument Calibration Standard Heavy Metals

JY-CALHM-ASL-R1-1 100 mL
JY-CALHM-ASL-R1-5 500 mL
 At stated conc. (µg/mL) in 2-5% HNO₃ 5 comps.

As (Arsenic)	100
Tl (Thallium)	100
Cd (Cadmium)	50
Se (Selenium)	50
Pb (Lead)	50

Instrument Calibration Standard

JY-CAL-ASL-1 100 mL
JY-CAL-ASL-5 500 mL
 5000 µg/mL each in 2-5% HNO₃ 4 comps.

Ca (Calcium)	K (Potassium)
Mg (Magnesium)	Na (Sodium)

Instrument Check Standard

JY-CHK-ASL-1 100 mL
JY-CHK-ASL-5 500 mL
 50 µg/mL each in 2-5% HNO₃ 9 comps.

Al (Aluminum)	K (Potassium)
As (Arsenic)	Na (Sodium)
Co (Cobalt)	P (Phosphorus)
Cr (Chromium)	Pb (Lead)
Cu (Copper)	

Instrument Check Standard 1

JY-CHK1-ASL-1 100 mL
JY-CHK1-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

K (Potassium)	1500
Pb (Lead)	1000
Al (Aluminum)	500
Mg (Magnesium)	500
Cd (Cadmium)	100

Quality Control Standard 7

JY-QC7-ASL-1 100 mL
JY-QC7-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 7 comps.

K (Potassium)	1000
Si (Silicon)	500
Al (Aluminum)	100
B (Boron)	100
Ba (Barium)	100
Na (Sodium)	100
Ag (Silver)	50

Quality Control Standard 21

JY-QC21-ASL-1 100 mL
JY-QC21-ASL-5 500 mL
 100 µg/mL each in 2-5% HNO₃ tr. HF 21 comps.

As (Arsenic)	Mo (Molybdenum)
Be (Beryllium)	Ni (Nickel)
Ca (Calcium)	Pb (Lead)
Cd (Cadmium)	Sb (Antimony)
Co (Cobalt)	Se (Selenium)
Cr (Chromium)	Sr (Strontium)
Cu (Copper)	Ti (Titanium)
Fe (Iron)	Tl (Thallium)
Li (Lithium)	V (Vanadium)
Mg (Magnesium)	Zn (Zinc)
Mn (Manganese)	

Quality Control Standard 23

JY-QC23-ASL-1 100 mL
JY-QC23-ASL-5 500 mL
 1000 µg/mL each in 2-5% HNO₃ 23 comps.

Ag (Silver)	In (Indium)
Al (Aluminum)	K (Potassium)
B (Boron)	Li (Lithium)
Ba (Barium)	Mg (Magnesium)
Bi (Bismuth)	Mn (Manganese)
Cd (Cadmium)	Na (Sodium)
Ca (Calcium)	Ni (Nickel)
Cr (Chromium)	Pb (Lead)
Co (Cobalt)	Sr (Strontium)
Cu (Copper)	Tl (Thallium)
Fe (Iron)	Zn (Zinc)
Ga (Gallium)	

Inorganic products containing acid generally require a hazardous shipping fee.

Inorganic products in water generally do not.



AccuStandard equivalent of Perkin Elmer (PE)

Alternate Interferents A

PE-ALTINTA-ASL-1 100 mL
PE-ALTINTA-ASL-5 500 mL
 1000 µg/mL each in 5% HNO₃ 6 comps.

Cr (Chromium)	Ni (Nickel)
Cu (Copper)	Ti (Titanium)
Mn (Manganese)	V (Vanadium)

Analytes B

PE-ANAB-ASL-1 100 mL
PE-ANAB-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF, tr.
 Tartaric acid 14 comps.

Cd (Cadmium)	100
Ni (Nickel)	100
Zn (Zinc)	100
Sb (Antimony)	60
Ba (Barium)	50
Be (Beryllium)	50
Co (Cobalt)	50
Cr (Chromium)	50
Cu (Copper)	50
Mn (Manganese)	50
V (Vanadium)	50
Ag (Silver)	20
As (Arsenic)	10
Tl (Thallium)	10

Alternate Analytes B

PE-ALTB-ASL-1 100 mL
PE-ALTB-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF, tr.
 Tartaric acid 12 comps.

Al (Aluminum)	100
As (Arsenic)	100
B (Boron)	100
Mo (Molybdenum)	100
Na (Sodium)	100
Sb (Antimony)	100
Se (Selenium)	100
Tl (Thallium)	100
Ca (Calcium)	10
Fe (Iron)	10
Mg (Magnesium)	10
Si (Silicon)	10

Instrument Calibration Std. 1

PE-CAL1-ASL-1 100 mL
PE-CAL1-ASL-5 500 mL
 20 µg/mL each in 2% HNO₃ tr. Tartaric acid
 20 comps.

Ag (Silver)	Mo (Molybdenum)
Al (Aluminum)	Ni (Nickel)
As (Arsenic)	Pb (Lead)
Ba (Barium)	Sb (Antimony)
Be (Beryllium)	Se (Selenium)
Cd (Cadmium)	Th (Thorium)
Co (Cobalt)	Tl (Thallium)
Cr (Chromium)	U (Uranium)
Cu (Copper)	V (Vanadium)
Mn (Manganese)	Zn (Zinc)

Instrument Calibration Std. 2

PE-CAL2-ASL-1 100 mL
PE-CAL2-ASL-5 500 mL
 100 µg/mL each in 5% HNO₃ tr. HF, tr. Tartaric acid
 26 comps.

Ag (Silver)	Mn (Manganese)
Al (Aluminum)	Mo (Molybdenum)
As (Arsenic)	Na (Sodium)
Ba (Barium)	Ni (Nickel)
Be (Beryllium)	Pb (Lead)
Ca (Calcium)	Sb (Antimony)
Cd (Cadmium)	Se (Selenium)
Co (Cobalt)	Sn (Tin)
Cr (Chromium)	Sr (Strontium)
Cu (Copper)	Ti (Titanium)
Fe (Iron)	Tl (Thallium)
K (Potassium)	V (Vanadium)
Mg (Magnesium)	Zn (Zinc)

Instrument Calibration Std. 3

PE-CAL3-ASL-1 100 mL
PE-CAL3-ASL-5 500 mL
 1000 µg/mL each in 5% HNO₃ 5 comps.

Fe (Iron)	Na (Sodium)
K (Potassium)	Mg (Magnesium)
Ca (Calcium)	

Initial Calibration Verification Std.

PE-CRDL1-ASL-1 100 mL
PE-CRDL1-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. Tartaric acid
 21 comps.

Ca (Calcium)	5000
Mg (Magnesium)	5000
K (Potassium)	5000
Na (Sodium)	5000
Ba (Barium)	200
Al (Aluminum)	200
Fe (Iron)	100
Sb (Antimony)	60
Co (Cobalt)	50
V (Vanadium)	50
Ni (Nickel)	40
Cu (Copper)	25
Zn (Zinc)	20
Mn (Manganese)	15
As (Arsenic)	10
Cr (Chromium)	10
Ag (Silver)	10
Tl (Thallium)	10
Cd (Cadmium)	5
Se (Selenium)	5
Pb (Lead)	3

Instrument Calibration Std. 1

PE-CAL4-ASL-1 100 mL
PE-CAL4-ASL-5 500 mL
 5000 µg/mL each in 5% HNO₃ 4 comps.

Ca (Calcium)	Na (Sodium)
K (Potassium)	
Mg (Magnesium)	

Instrument Calibration Std. 2

PE-CAL5-ASL-1 100 mL
PE-CAL5-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

Ni (Nickel)	400
Zn (Zinc)	200
Mn (Manganese)	150
Ag (Silver)	100
Cr (Chromium)	100

Instrument Calibration Std. 3

PE-CAL6-ASL-1 100 mL
PE-CAL6-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 7 comps.

Al (Aluminum)	2000
Ba (Barium)	2000
Fe (Iron)	1000
Co (Cobalt)	500
V (Vanadium)	500
Cu (Copper)	250
Be (Beryllium)	50

Instrument Calibration Std. 4

PE-CAL7-ASL-1 100 mL
PE-CAL7-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

As (Arsenic)	100
Tl (Thallium)	100
Cd (Cadmium)	50
Se (Selenium)	50
Pb (Lead)	50

Detection Limit

PE-CRDL2-ASL-1 100 mL
PE-CRDL2-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF tr.
 Tartaric acid 15 comps.

Sb (Antimony)	120
Co (Cobalt)	100
V (Vanadium)	100
Ni (Nickel)	80
Cu (Copper)	50
Zn (Zinc)	40
Mn (Manganese)	30
Ag (Silver)	20
As (Arsenic)	20
Cr (Chromium)	20
Tl (Thallium)	20
Be (Beryllium)	10
Cd (Cadmium)	10
Se (Selenium)	10
P (Phosphorus)	6

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.

Supplied as a 10X concentrate for better stability.



AccuStandard equivalent of Perkin Elmer (PE)

Instrument Check Standard 1

PE-CHK1-ASL-1 100 mL
PE-CHK1-ASL-5 500 mL
 10 µg/mL each in 2% HNO₃ tr. HF, tr. Tartaric acid
 17 comps.

Ag (Silver)	Mn (Manganese)
Al (Aluminum)	Ni (Nickel)
As (Arsenic)	Pb (Lead)
Ba (Barium)	Sb (Antimony)
Be (Beryllium)	Se (Selenium)
Cd (Cadmium)	Tl (Thallium)
Co (Cobalt)	V (Vanadium)
Cr (Chromium)	Zn (Zinc)
Cu (Copper)	

Instrument Check Standard 3

PE-CHK3-ASL-1 100 mL
PE-CHK3-ASL-5 500 mL
 200 µg/mL each in 2% HNO₃ 5 comps.

Ca (Calcium)	Mg (Magnesium)
Fe (Iron)	Na (Sodium)
K (Potassium)	

Instrument Check Standard 4

PE-CHK4-ASL-1 100 mL
PE-CHK4-ASL-5 500 mL
 10 µg/mL each in 2% HNO₃ 3 comps.

Mo (Molybdenum)	U (Uranium)
Th (Thorium)	

Instrument Check Standard 5

PE-CHK5-ASL-1 100 mL
PE-CHK5-ASL-5 500 mL
 10 µg/mL each in 2% HNO₃ tr. HF 4 comps.

Mo (Molybdenum)	Sr (Strontium)
Sn (Tin)	Ti (Titanium)

Detection Limit Standard for use with the ELAN 5000

PE-CRDL3-ASL-1 100 mL
PE-CRDL3-ASL-5 500 mL
 1 µg/mL each in 1% HNO₃ 5 comps.

Be (Beryllium)	Tl (Thallium)
Co (Cobalt)	U (Uranium)
In (Indium)	

Supplied as a 100X concentrate for better stability.

ELAN 6100 Detection Limit Solution

PE-CRDL4-ASL-1 100 mL
PE-CRDL4-ASL-5 500 mL
 10 µg/mL each in 1% HNO₃ 4 comps.

Be (Beryllium)	In (Indium)
Co (Cobalt)	U (Uranium)

Supplied as a 1000X concentrate for better stability.

Interference Check Standard 5

PE-ICSS-ASL-1 100 mL
PE-ICSS-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

Ca (Calcium)	6000
Fe (Iron)	5000
Mg (Magnesium)	3000
Al (Aluminum)	1200
Na (Sodium)	1000

Interference Check Standard 18

PE-ICS18-ASL-1-SET 2 x 100 mL
PE-ICS18-ASL-5-SET 2 x 500 mL

PE-ICS18-ASL
 At stated conc. (µg/mL) in 5% HNO₃ 16 comps.

K (Potassium)	20000
As (Arsenic)	1000
Pb (Lead)	1000
Tl (Thallium)	1000
Se (Selenium)	500
Ag (Silver)	300
Ba (Barium)	300
Cd (Cadmium)	300
Co (Cobalt)	300
Cr (Chromium)	300
Cu (Copper)	300
Ni (Nickel)	300
V (Vanadium)	300
Zn (Zinc)	300
Mn (Manganese)	200
Be (Beryllium)	100

PE-ICS18-HG-ASL
 100 µg/mL in 5% HNO₃

Hg (Mercury)

Supplied separately for better product stability.

Internal Standard Mix

PE-INT-ASL-1 100 mL
PE-INT-ASL-5 500 mL
 10 µg/mL each in 5-10% HNO₃ 7 comps.

Li6 (Lithium)	In (Indium)
Sc (Scandium)	Tb (Terbium)
Ge (Germanium)	Bi (Bismuth)
Y (Yttrium)	

Interferents A

PE-INTA-ASL-1 100 mL
PE-INTA-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 4 comps.

Al (Aluminum)	5000
Ca (Calcium)	5000
Mg (Magnesium)	5000
Fe (Iron)	2000

Interferents Check Solution 1

PE-INTFR1-ASL-1 100 mL
PE-INTFR1-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 12 comps.

Cl (Chloride)	10000
C (Carbon)	2000
Al (Aluminum)	100
Ca (Calcium)	100
Fe (Iron)	100
K (Potassium)	100
Mg (Magnesium)	100
Na (Sodium)	100
P (Phosphorus)	100
S (Sulfur)	100
Mo (Molybdenum)	20
Ti (Titanium)	20

Interference Check Solution 2

PE-INTFR2-ASL-1 100 mL
PE-INTFR2-ASL-5 500 mL
 10 µg/mL each in 2% HNO₃ 9 comps.

Ag (Silver)	Cu (Copper)
As (Arsenic)	Mn (Manganese)
Cd (Cadmium)	Ni (Nickel)
Co (Cobalt)	Zn (Zinc)
Cr (Chromium)	

Interference Check Standard A

PE-INTFRA-ASL-1 100 mL
PE-INTFRA-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF
 12 comps.

Cl (Chloride)	21215
Ca (Calcium)	3000
Na (Sodium)	2500
Fe (Iron)	2500
C (Carbon)	2000
Al (Aluminum)	1000
K (Potassium)	1000
Mg (Magnesium)	1000
P (Phosphorus)	1000
S (Sulfur)	1000
Mo (Molybdenum)	20
Ti (Titanium)	20

Interference Check Standard B

PE-INTFRB-ASL-1 100 mL
PE-INTFRB-ASL-5 500 mL
 At stated conc. (µg/mL) in 2% HNO₃ 11 comps.

Co (Cobalt)	20
Cr (Chromium)	20
Cu (Copper)	20
Mn (Manganese)	20
Ni (Nickel)	20
V (Vanadium)	20
As (Arsenic)	10
Cd (Cadmium)	10
Se (Selenium)	10
Zn (Zinc)	10
Ag (Silver)	5

Interference Check Standard C

PE-INTFRC-ASL-1-SET 2 x 100 mL
PE-INTFRC-ASL-5-SET 2 x 500 mL
 2 µg/mL each in 2% HNO₃ tr. HF tr. Tartaric acid
 16 comps.

Sb (Antimony)	Mn (Manganese)
As (Arsenic)	Ni (Nickel)
Ba (Barium)	Se (Selenium)
Be (Beryllium)	Ag (Silver)
Cd (Cadmium)	Tl (Thallium)
Cr (Chromium)	V (Vanadium)
Co (Cobalt)	Zn (Zinc)
Cu (Copper)	
Pb (Lead)	

PE-INTFRC-HG-ASL

2 µg/mL in 5% HNO₃

Hg (Mercury)

Supplied separately for better product stability.

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.



AccuStandard equivalent of Perkin Elmer (PE)

Mixed Calibration Standard

PE-MCS-ASL-1 100 mL
PE-MCS-ASL-5 500 mL
 At stated conc. (µg/mL) in 2% HNO₃ 10 comps.

As (Arsenic)	50
K (Potassium)	50
La (Lanthanum)	10
Li (Lithium)	10
Mn (Manganese)	10
Ni (Nickel)	10
Sr (Strontium)	10
Zn (Zinc)	10
Ba (Barium)	1
Mg (Magnesium)	1

Mixed Calibration Standard 1

PE-MCS1-ASL-1 100 mL
PE-MCS1-ASL-5 500 mL
 At stated conc. (µg/mL) in 2% HNO₃ 6 comps.

Pb (Lead)	500
Se (Selenium)	200
Cd (Cadmium)	150
Zn (Zinc)	150
Mn (Manganese)	100
Be (Beryllium)	50

Mixed Calibration Standard 2

PE-MCS2-ASL-1 100 mL
PE-MCS2-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

Fe (Iron)	10000
Ba (Barium)	100
Co (Cobalt)	100
Cu (Copper)	100
V (Vanadium)	100

Mixed Calibration Standard 3

PE-MCS3-ASL-1 100 mL
PE-MCS3-ASL-5 500 mL
 at stated conc. (µg/mL) in 2% HNO₃ tr. HF 3 comps.

As (Arsenic)	500
Mo (Molybdenum)	100
Si (Silicon)	100

Mixed Calibration Standard 4

PE-MCS4-ASL-1 100 mL
PE-MCS4-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 6 comps.

Ca (Calcium)	1000
K (Potassium)	400
Al (Aluminum)	200
Na (Sodium)	200
Cr (Chromium)	20
Ni (Nickel)	20

Mixed Calibration Standard 5

PE-MCS5-ASL-1 100 mL
PE-MCS5-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ tr. HF tr.
 Tartaric acid 5 comps.

Mg (Magnesium)	1000
Sb (Antimony)	200
Tl (Thallium)	200
B (Boron)	100
Ag (Silver)	50

Multi-Element Calibration Standard 1

PE-MECAL1-ASL-1 100 mL
PE-MECAL1-ASL-5 500 mL
 10 µg/mL each in 2% HNO₃ 9 comps.

Be (Beryllium)	Mg (Magnesium)
Bi (Bismuth)	Ni (Nickel)
Ce (Cerium)	Pb (Lead)
Co (Cobalt)	U (Uranium)
In (Indium)	

Multi-Element Calibration Standard 2

PE-MECAL2-ASL-1 100 mL
PE-MECAL2-ASL-5 500 mL
 10 µg/mL each in 5% HNO₃ 17 comps.

Ce (Cerium)	Pr (Praseodymium)
Dy (Dysprosium)	Sm (Samarium)
Er (Erbium)	Sc (Scandium)
Eu (Europium)	Tb (Terbium)
Gd (Gadolinium)	Th (Thorium)
Ho (Holmium)	Tm (Thulium)
La (Lanthanum)	Yb (Ytterbium)
Lu (Lutetium)	Y (Yttrium)
Nd (Neodymium)	

Multi-Element Calibration Standard 3

PE-MECAL3-ASL-1-SET 2 x 100 mL
PE-MECAL3-ASL-5-SET 2 x 500 mL
PE-MECAL3-ASL 10 µg/mL each in 5% HNO₃ 29 comps.

Ag (Silver)	K (Potassium)
Al (Aluminum)	Li (Lithium)
As (Arsenic)	Mg (Magnesium)
Ba (Barium)	Mn (Manganese)
Be (Beryllium)	Na (Sodium)
Bi (Bismuth)	Ni (Nickel)
Ca (Calcium)	Pb (Lead)
Cd (Cadmium)	Rb (Rubidium)
Co (Cobalt)	Se (Selenium)
Cr (Chromium)	Sr (Strontium)
Cs (Cesium)	Tl (Thallium)
Cu (Copper)	U (Uranium)
Fe (Iron)	V (Vanadium)
Ga (Gallium)	Zn (Zinc)
In (Indium)	

PE-MECAL3-HG-ASL

10 µg/mL in 5% HNO₃

Hg (Mercury)

Supplied separately for better product stability.

Multi-Element Calibration Standard 4

PE-MECAL4-ASL-1 100 mL
PE-MECAL4-ASL-5 500 mL
 10 µg/mL each in 10% HCl 10 comps.

Au (Gold)	Rh (Rhodium)
Hf (Hafnium)	Ru (Ruthenium)
Ir (Iridium)	Sb (Antimony)
Pd (Palladium)	Sn (Tin)
Pt (Platinum)	Te (Tellurium)

Multi-Element Calibration Standard 5

PE-MECAL5-ASL-1 ▼ 100 mL
PE-MECAL5-ASL-5 ▼ 500 mL
 10 µg/mL each in Water, tr. HF 12 comps.

B (Boron)	S (Sulfur)
Ge (Germanium)	Si (Silicon)
Mo (Molybdenum)	Ta (Tantalum)
Nb (Niobium)	Ti (Titanium)
P (Phosphorus)	W (Tungsten)
Re (Rhenium)	Zr (Zirconium)

Multi-Element Internal Standard

PE-MEINT-ASL-1 100 mL
PE-MEINT-ASL-5 500 mL
 10 µg/mL each in 2% HNO₃ 7 comps.

Bi (Bismuth)	Sc (Scandium)
Ho (Holmium)	Tb (Terbium)
In (Indium)	Y (Yttrium)
Li6 (Lithium)	

Memory Test 1

PE-MEM1-ASL-1 100 mL
PE-MEM1-ASL-5 500 mL
 At stated conc. (µg/mL) in 5% HNO₃ 21 comps.

Al (Aluminum)	1000
Ca (Calcium)	1000
Fe (Iron)	1000
K (Potassium)	1000
Mg (Magnesium)	1000
Na (Sodium)	1000
Ag (Silver)	20
As (Arsenic)	20
Ba (Barium)	20
Be (Beryllium)	20
Cd (Cadmium)	20
Co (Cobalt)	20
Cr (Chromium)	20
Cu (Copper)	20
Mn (Manganese)	20
Ni (Nickel)	20
Pb (Lead)	20
Se (Selenium)	20
Tl (Thallium)	20
V (Vanadium)	20
Zn (Zinc)	20

Memory Test 2

PE-MEM2-ASL-1 ▼ 100 mL
PE-MEM2-ASL-5 ▼ 500 mL
 At stated conc. (µg/mL) in Water, tr. HF 6 comps.

Cl (Chloride)	7200
C (Carbon)	2000
P (Phosphorus)	1000
Mo (Molybdenum)	20
Sb (Antimony)	20
Tl (Thallium)	20

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.

▼ Hazardous fee not required.



AccuStandard equivalent of Perkin Elmer (PE)

QC Standard 7 Elements

PE-QC7-ASL-1 100 mL
PE-QC7-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃ tr. HF
7 comps.

K (Potassium)	1000
Si (Silicon)	500
Al (Aluminum)	100
B (Boron)	100
Ba (Barium)	100
Na (Sodium)	100
Ag (Silver)	50

QC Standard 21 Elements

PE-QC21-ASL-1 100 mL
PE-QC21-ASL-5 500 mL

100 µg/mL each in 5% HNO₃, tr. HF, tr. Tartaric acid
21 comps.

As (Arsenic)	Mo (Molybdenum)
Be (Beryllium)	Ni (Nickel)
Ca (Calcium)	Pb (Lead)
Cd (Cadmium)	Sb (Antimony)
Co (Cobalt)	Se (Selenium)
Cr (Chromium)	Sr (Strontium)
Cu (Copper)	Ti (Titanium)
Fe (Iron)	Tl (Thallium)
Li (Lithium)	V (Vanadium)
Mg (Magnesium)	Zn (Zinc)
Mn (Manganese)	

Primary Drinking Water Metals

PE-SDWA1-ASL-1-SET 2 x 100 mL
PE-SDWA1-ASL-5-SET 2 x 500 mL

PE-SDWA1-ASL

At stated conc. (µg/mL) in 2% HNO₃ 7 comps.

Ba (Barium)	100
Ag (Silver)	10
As (Arsenic)	10
Cr (Chromium)	10
Pb (Lead)	10
Cd (Cadmium)	5
Se (Selenium)	5

PE-SDWA1-HG-ASL

10 µg/mL in 2% HNO₃

Hg (Mercury)

Supplied separately for better product stability.

Secondary Drinking Water Metals

PE-SDWA2-ASL-1 100 mL
PE-SDWA2-ASL-5 500 mL

At stated conc. (µg/mL) in 2% HNO₃ 4 comps.

Zn (Zinc)	500
Cu (Copper)	100
Fe (Iron)	30
Mn (Manganese)	5

ELAN 6100 DRC Sensitivity/ Detection Limit Solution

PE-SENS-ASL-1 100 mL
PE-SENS-ASL-5 500 mL

1 µg/mL each in 2% HNO₃, tr. HCl 13 comps.

Ba (Barium)	Pb (Lead)
Be (Beryllium)	Mg (Magnesium)
Ca (Calcium)	K (Potassium)
Ce (Cerium)	Rh (Rhodium)
Co (Cobalt)	Na (Sodium)
In (Indium)	U (Uranium)
Fe (Iron)	

ELAN 9000/6X00 Dual Detector Calibration Solution

PE-SETUP1-ASL-1 100 mL
PE-SETUP1-ASL-5 500 mL

2 µg/mL each in 1% HNO₃ tr. HCl 5 comps.

Cd (Cadmium)	Mg (Magnesium)
Cu (Copper)	Rh (Rhodium)
Pb (Lead)	

Supplied as a 10X concentrate for better stability.

ELAN 6000/5000 Plasma Setup Solution

PE-SETUP2-ASL-1 100 mL
PE-SETUP2-ASL-5 500 mL

1 µg/mL each in 1% HNO₃ tr. HCl 11 comps.

Ba (Barium)	Mg (Magnesium)
Cd (Cadmium)	Rh (Rhodium)
Ce (Cerium)	Sc (Scandium)
Cu (Copper)	Tb (Terbium)
Ge (Germanium)	Tl (Thallium)
Pb (Lead)	

Supplied as a 100X concentrate for better stability.

SmartTune Solution for ELAN/DRC-e

PE-SMTUNE-ASL-1 100 mL
PE-SMTUNE-ASL-5 500 mL

1 µg/mL each in 2% HNO₃ tr. HCl 9 comps.

Ba (Barium)	Pb (Lead)
Be (Beryllium)	Mg (Magnesium)
Ce (Cerium)	Rh (Rhodium)
Co (Cobalt)	U (Uranium)
In (Indium)	

Supplied as a 100X concentrate for better stability.

SmartTune Solution for DRC/ DRC^{Plus}/DRC II

PE-SMTUNE2-ASL-1 100 mL
PE-SMTUNE2-ASL-5 500 mL

At stated conc. (µg/mL) in 0.5% HNO₃ 10 comps.

Ba (Barium)	10
Be (Beryllium)	1
Ce (Cerium)	1
Co (Cobalt)	1
In (Indium)	1
Fe (Iron)	1
Pb (Lead)	1
Mg (Magnesium)	1
Th (Thorium)	1
U (Uranium)	1

Supplied as a 1000X concentrate for better stability.

Spike Sample Analysis

PE-SPIKE-ASL-1 100 mL
PE-SPIKE-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃, tr. HF, tr. Tartaric acid 18 comps.

Al (Aluminum)	200
As (Arsenic)	200
Ba (Barium)	200
Se (Selenium)	200
Tl (Thallium)	200
Fe (Iron)	100
Co (Cobalt)	50
Mn (Manganese)	50
Ni (Nickel)	50
Pb (Lead)	50
Sb (Antimony)	50
V (Vanadium)	50
Zn (Zinc)	50
Cu (Copper)	25
Cr (Chromium)	20
Ag (Silver)	5
Be (Beryllium)	5
Cd (Cadmium)	5

Spike Sample Standard I (Water)

PE-SPIKE1-ASL-1 100 mL
PE-SPIKE1-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃, tr. HF, tr. Tartaric acid 17 comps.

Fe (Iron)	500
Ba (Barium)	250
Zn (Zinc)	250
Co (Cobalt)	100
Cr (Chromium)	100
Cu (Copper)	100
Mn (Manganese)	100
Ni (Nickel)	100
Sb (Antimony)	100
V (Vanadium)	100
As (Arsenic)	50
Pb (Lead)	50
Ag (Silver)	25
Be (Beryllium)	25
Cd (Cadmium)	25
Se (Selenium)	25
Tl (Thallium)	25

Spike Sample Standard II (Soil)

PE-SPIKE2-ASL-1 100 mL
PE-SPIKE2-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃, tr. HF, tr. Tartaric acid 15 comps.

Ba (Barium)	250
Cr (Chromium)	250
Cu (Copper)	250
Zn (Zinc)	250
V (Vanadium)	150
Ni (Nickel)	125
Co (Cobalt)	100
Pb (Lead)	100
Sb (Antimony)	100
As (Arsenic)	50
Cd (Cadmium)	50
Ag (Silver)	25
Be (Beryllium)	25
Se (Selenium)	25
Tl (Thallium)	25

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.



AccuStandard equivalent of Perkin Elmer (PE)

Spike Sample Standard III (for ILM 05.2)

PE-SPIKE3-ASL-1 100 mL
PE-SPIKE3-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃, tr. HF, tr. Tartaric acid 17 comps.

Al (Aluminum)	200
Ba (Barium)	200
Co (Cobalt)	50
Mn (Manganese)	50
Ni (Nickel)	50
V (Vanadium)	50
Zn (Zinc)	50
Cu (Copper)	25
Cr (Chromium)	20
Sb (Antimony)	10
Be (Beryllium)	5
Cd (Cadmium)	5
Ag (Silver)	5
Tl (Thallium)	5
As (Arsenic)	4
Pb (Lead)	2
Se (Selenium)	1

ELAN 9000/6100 Setup/Stab/Masscal Solution

PE-STAB-ASL-1 100 mL
PE-STAB-ASL-5 500 mL

1 µg/mL each in 1% HNO₃ tr. HCl 9 comps.

Ba (Barium)	Pb (Lead)
Cd (Cadmium)	Mg (Magnesium)
Ce (Cerium)	Rh (Rhodium)
Cu (Copper)	U (Uranium)
In (Indium)	

Supplied as a 100X concentrate for better stability.

Tuning Solution I

PE-TUNSOL-ASL-1 100 mL
PE-TUNSOL-ASL-5 500 mL

10 µg/mL each in 2% HNO₃, tr. HCl 12 comps.

Ba (Barium)	Mg (Magnesium)
Be (Beryllium)	Pb (Lead)
Ce (Cerium)	Rh (Rhodium)
Co (Cobalt)	Tl (Thallium)
In (Indium)	U (Uranium)
Li (Lithium)	Y (Yttrium)

Low UV Standard

PE-UV-ASL-1 100 mL
PE-UV-ASL-5 500 mL

10 µg/mL each in 2% HNO₃ 3 comps.

Al (Aluminum)	S (Sulfur)
P (Phosphorus)	

VIS Wavecal Solution

PE-VISWAVE-ASL-1 100 mL
PE-VISWAVE-ASL-5 500 mL

At stated conc. (µg/mL) in 2% HNO₃ 8 comps.

K (Potassium)	50
La (Lanthanum)	10
Li (Lithium)	10
Mn (Manganese)	10
Na (Sodium)	10
Sr (Strontium)	10
Ba (Barium)	1
Ca (Calcium)	1

UV Wavecal Solution

PE-UVWAVE-ASL-1 100 mL
PE-UVWAVE-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HCl 11 comps.

K (Potassium)	100
P (Phosphorus)	100
S (Sulfur)	100
As (Arsenic)	20
La (Lanthanum)	20
Li (Lithium)	20
Mn (Manganese)	20
Mo (Molybdenum)	20
Na (Sodium)	20
Ni (Nickel)	20
Sc (Scandium)	20

Initial Calibration Verification Standard 1

PE-VER1-ASL-1 100 mL
PE-VER1-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃ tr. Tartaric acid 26 comps.

Fe (Iron)	1000
K (Potassium)	1000
Ca (Calcium)	1000
Na (Sodium)	1000
Mg (Magnesium)	1000
Sr (Strontium)	1000
Ag (Silver)	10
Al (Aluminum)	10
As (Arsenic)	10
Ba (Barium)	10
Be (Beryllium)	10
Cd (Cadmium)	10
Co (Cobalt)	10
Cr (Chromium)	10
Cu (Copper)	10
Mn (Manganese)	10
Mo (Molybdenum)	10
Ni (Nickel)	10
Pb (Lead)	10
Sb (Antimony)	10
Se (Selenium)	10
Tl (Thallium)	10
V (Vanadium)	10
Zn (Zinc)	10
Th (Thorium)	10
U (Uranium)	10

Initial Calibration Verification Standard 2

PE-VER2-ASL-R1-1 100 mL
PE-VER2-ASL-R1-5 500 mL

10 µg/mL each in 2% HNO₃ tr. HF 2 comps.

Sn (Tin)	Ti (Titanium)
----------	---------------

Trace Metals I

PE-WPTM1-ASL-1-SET 2 x 100 mL
PE-WPTM1-ASL-5-SET 2 x 500 mL

PE-WPTM1-ASL

At stated conc. (µg/mL) in 5% HNO₃ 14 comps.

Al (Aluminum)	500
V (Vanadium)	250
As (Arsenic)	100
Be (Beryllium)	100
Co (Cobalt)	100
Cr (Chromium)	100
Cu (Copper)	100
Fe (Iron)	100
Mn (Manganese)	100
Ni (Nickel)	100
Pb (Lead)	100
Zn (Zinc)	100
Cd (Cadmium)	25
Se (Selenium)	25

PE-WPTM1-HG-ASL

10 µg/mL in 5% HNO₃

Hg (Mercury)

Supplied separately for better product stability.

Trace Metals II

PE-WPTM2-ASL-1 100 mL
PE-WPTM2-ASL-5 500 mL

At stated conc. (µg/mL) in 2% HNO₃ 3 comps.

Sb (Antimony)	20
Tl (Thallium)	20
Ag (Silver)	10

Trace Metals III

PE-WPTM3-ASL-1 100 mL
PE-WPTM3-ASL-5 500 mL

At stated conc. (µg/mL) in 2% HNO₃ 6 comps.

Ba (Barium)	500
Ca (Calcium)	500
Mo (Molybdenum)	500
Na (Sodium)	500
K (Potassium)	100
Mg (Magnesium)	100

Alternate Metals 1

PE-WPAM1-ASL-1 100 mL
PE-WPAM1-ASL-5 500 mL

At stated conc. (µg/mL) in 2% HNO₃ 11 comps.

Al (Aluminum)	20
Fe (Iron)	20
V (Vanadium)	20
Co (Cobalt)	10
Cu (Copper)	10
Mn (Manganese)	10
Ni (Nickel)	10
Zn (Zinc)	10
Be (Beryllium)	5
Sb (Antimony)	5
Tl (Thallium)	5

Alternate Metals 3

PE-WPAM3-ASL-1 100 mL
PE-WPAM3-ASL-5 500 mL

At stated conc. (µg/mL) in 2% HNO₃ 4 comps.

Ca (Calcium)	500
Na (Sodium)	500
K (Potassium)	100
Mg (Magnesium)	100

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.



AccuStandard equivalent of Teledyne

Check Mate 1

TELE-CHK1-ASL-1-SET 2 x 100 mL
TELE-CHK1-ASL-5-SET 2 x 500 mL

TELE-CHK1-ASL

At stated conc. (µg/mL) in 5% HCl, 1% HNO₃
 24 comps.

Ca (Calcium)	100
K (Potassium)	100
Mg (Magnesium)	100
Na (Sodium)	100
Al (Aluminum)	10
As (Arsenic)	10
B (Boron)	10
Ba (Barium)	10
Be (Beryllium)	10
Cd (Cadmium)	10
Co (Cobalt)	10
Cr (Chromium)	10
Cu (Copper)	10
Fe (Iron)	10
Mn (Manganese)	10
Mo (Molybdenum)	10
Ni (Nickel)	10
Pb (Lead)	10
Sb (Antimony)	10
Se (Selenium)	10
Si (Silicon)	10
Tl (Thallium)	10
V (Vanadium)	10
Zn (Zinc)	10

TELE-CHK1-AG-ASL

1000 µg/mL in 2% HNO₃

Ag (Silver)

Supplied separately for better product stability.

Check Mate 2

TELE-CHK2-ASL-1-SET 2 x 100 mL
TELE-CHK2-ASL-5-SET 2 x 500 mL

TELE-CHK2-ASL

At stated conc. (µg/mL) in 5% HCl, 1% HNO₃
 17 comps.

Ca (Calcium)	100
K (Potassium)	100
Mg (Magnesium)	100
Na (Sodium)	100
Al (Aluminum)	10
Ba (Barium)	10
Be (Beryllium)	10
Cd (Cadmium)	10
Co (Cobalt)	10
Cr (Chromium)	10
Cu (Copper)	10
Fe (Iron)	10
Mn (Manganese)	10
Ni (Nickel)	10
Sb (Antimony)	10
V (Vanadium)	10
Zn (Zinc)	10

TELE-CHK2-AG-ASL

1000 µg/mL in 2% HNO₃

Ag (Silver)

Supplied separately for better product stability.

Check Mate 3

TELE-CHK3-ASL-1-SET 2 x 100 mL
TELE-CHK3-ASL-5-SET 2 x 500 mL

TELE-CHK3-ASL

At stated conc. (µg/mL) in 5% HCl, 1% HNO₃
 17 comps.

Ca (Calcium)	10
K (Potassium)	10
Mg (Magnesium)	10
Na (Sodium)	10
Al (Aluminum)	1
Ba (Barium)	1
Be (Beryllium)	1
Cd (Cadmium)	1
Co (Cobalt)	1
Cr (Chromium)	1
Cu (Copper)	1
Fe (Iron)	1
Mn (Manganese)	1
Ni (Nickel)	1
Sb (Antimony)	1
V (Vanadium)	1
Zn (Zinc)	1

TELE-CHK3-AG-ASL

1000 µg/mL in 2% HNO₃

Ag (Silver)

Supplied separately for better product stability.

Check Mate 4

TELE-CHK4-ASL-1 100 mL
TELE-CHK4-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃ 22 comps.

Ca (Calcium)	5000
K (Potassium)	5000
Mg (Magnesium)	5000
Na (Sodium)	5000
Ba (Barium)	200
Fe (Iron)	100
Al (Aluminum)	60
Sb (Antimony)	60
Co (Cobalt)	50
V (Vanadium)	50
Ni (Nickel)	40
Cu (Copper)	25
Zn (Zinc)	20
Mn (Manganese)	15
Ag (Silver)	10
As (Arsenic)	10
Cr (Chromium)	10
Tl (Thallium)	10
Be (Beryllium)	5
Cd (Cadmium)	5
Pb (Lead)	5
Se (Selenium)	5

Check Mate 5

TELE-CHK5-ASL-1 100 mL
TELE-CHK5-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃ 16 comps.

Ca (Calcium)	2000
K (Potassium)	2000
Mg (Magnesium)	2000
Na (Sodium)	2000
Al (Aluminum)	1000
Ba (Barium)	1000
Fe (Iron)	1000
Co (Cobalt)	500
Ni (Nickel)	500
V (Vanadium)	500
Cr (Chromium)	200
Cu (Copper)	200
Ag (Silver)	100
Be (Beryllium)	100
Mn (Manganese)	100
Zn (Zinc)	100

Check Mate 6

TELE-CHK6-ASL-1 100 mL
TELE-CHK6-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃ 5 comps.

As (Arsenic)	500
Pb (Lead)	500
Se (Selenium)	500
Tl (Thallium)	500
Cd (Cadmium)	100

Check Mate 7

TELE-CHK7-ASL-1 100 mL
TELE-CHK7-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HCl, 1% HNO₃
 17 comps.

Ca (Calcium)	50
K (Potassium)	50
Mg (Magnesium)	50
Na (Sodium)	50
Al (Aluminum)	5
Ba (Barium)	5
Be (Beryllium)	5
Cd (Cadmium)	5
Co (Cobalt)	5
Cr (Chromium)	5
Cu (Copper)	5
Fe (Iron)	5
Mn (Manganese)	5
Ni (Nickel)	5
Sb (Antimony)	5
V (Vanadium)	5
Zn (Zinc)	5

Check Mate 8

TELE-CHK8-ASL-1 100 mL
TELE-CHK8-ASL-5 500 mL

At stated conc. (µg/mL) in 5% HNO₃ 22 comps.

Ca (Calcium)	5000
K (Potassium)	5000
Na (Sodium)	5000
Mg (Magnesium)	5000
Al (Aluminum)	2000
Ba (Barium)	2000
Fe (Iron)	1000
Sb (Antimony)	600
Co (Cobalt)	500
V (Vanadium)	500
Ni (Nickel)	400
Cu (Copper)	250
Zn (Zinc)	200
Mn (Manganese)	150
Ag (Silver)	100
As (Arsenic)	100
Cr (Chromium)	100
Tl (Thallium)	100
Be (Beryllium)	50
Cd (Cadmium)	50
Pb (Lead)	50
Se (Selenium)	50

AccuStandard is not affiliated with the companies and brands on this page, and the brands and company names appear for the purpose of cross reference with the corresponding AccuStandard product which is being offered.



ICP

Miscellaneous Applications

EU Formulation

Scope: For the determination of 32 elements by ICP (Inductively Coupled Plasma)

DIN EN ISO 11885 - 32 Element

ICP Standard Set

DINENISO-11885-1-SET	2 x 100 mL
DINENISO-11885-5-SET	2 x 500 mL

Part 1

DINENISO-11885A-1 100 mL

DINENISO-11885A-5 500 mL

At stated conc. (µg/mL) in 2-5% HNO₃ tr. HF
24 comps.

Ag (Silver)	20
Al (Aluminum)	40
As (Arsenic)	80
Ba (Barium)	2
Be (Beryllium)	2
Bi (Bismuth)	40
Ca (Calcium)	2
Cd (Cadmium)	10
Co (Cobalt)	10
Cr (Chromium)	10
Cu (Copper)	10
Fe (Iron)	20
K (Potassium)	50
Li (Lithium)	2
Mg (Magnesium)	1
Mn (Manganese)	2
Na (Sodium)	20
Ni (Nickel)	50
Pb (Lead)	50
Sb (Antimony)	50
Se (Selenium)	50
Sr (Strontium)	0.5
V (Vanadium)	10
Zn (Zinc)	5

Part 2

DINENISO-11885B-1 100 mL

DINENISO-11885B-5 500 mL

At stated conc. (µg/mL) in 2-5% HNO₃ tr. HF
8 comps.

B (Boron)	5
Mo (Molybdenum)	30
P (Phosphorus)	50
S (Sulfur)	50
Si (Silicon)	20
Sn (Tin)	50
Ti (Titanium)	5
Zr (Zirconium)	10

ASTM D5184 Aluminum and Silicon in Fuel Oils by Ashing, Fusion, ICP-AES Spectrometry & AA Spectrometry

Tartaric Acid / Hydrochloric Acid Solution

D-5184-91-TA-5 1 x 500 mL

Tartaric acid @ 0.5% w/v in 4% HCl

Silicon Standard Solution

D-5184-91-SI-1 ▼ 1 x 100 mL

D-5184-91-SI-5 ▼ 1 x 500 mL

Silicon @ 1000 µg/mL in water tr. NaOH

Aluminum Standard Solution

D-5184-91-AL-1 1 x 100 mL

D-5184-91-AL-5 1 x 500 mL

Aluminum @ 1000 µg/mL in 5 % HCl

Technical Note

Contact our Technical Service Department for **Ready-to-Aspirate** working level calibration curves designed for your laboratories' specific calibration ranges.

ASTM D5600 Trace Metals in Petroleum Coke by ICP-AES

Multi-Element Calibration Standard

D-5600-01-1 1 x 100 mL

500 µg/mL each in 2-5% HNO₃ tr. HF 12 comps.

Aluminum	Nickel
Barium	Silicon
Calcium	Sodium
Iron	Titanium
Magnesium	Vanadium
Manganese	Zinc

Hydrochloric Acid Diluent

D-5600-BLH-5 1 x 500 mL

D-5600-BLH-L-VAP L (2 x 500 mL)

20% HCl in ASTM Type I water

Lithium Borate Diluent

D-5600-LIB-1 1 x 100 mL

D-5600-LIB-5 1 x 500 mL

2.0% Lithium Borate in 10% HCl

Technical Note

D5600 Ready-to-Aspirate Standards

We have formulated the following stock standards for ASTM Method D5600. We have prepared numerous **Ready-to-Aspirate** ICP multi-element solutions. Should your company want to eliminate the preparation process for Inorganic standards, contact our Inorganic Technical Department for a quote on a **Ready-to-Aspirate** working level Inorganic standards.

▼ Hazardous fee not required.

ICP/MS

Multi-Element Standards



■ Ultra Pure Matrix ■ Special Packaging ■ Traceability to National Reference Materials

AccuStandard's ICP/MS Standards are formulated to meet the needs of this very special instrument. As matrix effect is of utmost concern, each standard is formulated in specially purified 18 megohm de-ionized water and ultra pure acids. After both wet chemical and instrumental analysis, the standards are packaged in acid leached FLPE containers to provide required protection.

Calibration Standards

These five standards encompass the entire range of elements all at 10 ppm.

Calibration Standard 1

ICP-MS-CAL1-1 100 mL
10 µg/mL each in 5% HNO₃ 17 comps.

Element	Isotope
Ce (Cerium)	140
Dy (Dysprosium)	164
Er (Erbium)	166
Eu (Europium)	153
Gd (Gadolinium)	158
Ho (Holmium)	165
La (Lanthanum)	139
Lu (Lutetium)	175
Nd (Neodymium)	143
Pr (Praseodymium)	141
Sm (Samarium)	152
Sc (Scandium)	45
Tb (Terbium)	159
Th (Thorium)	232
Tm (Thulium)	169
Yb (Ytterbium)	174
Y (Yttrium)	89

Calibration Standard 2

ICP-MS-CAL2-1 100 mL
10 µg/mL each in 5% HNO₃ 29 comps.

Element	Isotope
Al (Aluminum)	27
As (Arsenic)	75
Ba (Barium)	138
Be (Beryllium)	9
Bi (Bismuth)	209
Cd (Cadmium)	114
Ca (Calcium)	40
Cs (Cesium)	133
Cr (Chromium)	52
Co (Cobalt)	59
Cu (Copper)	63
Ga (Gallium)	69
In (Indium)	115
Fe (Iron)	56
Pb (Lead)	208
Li (Lithium)	7
Mg (Magnesium)	24
Mn (Manganese)	55
Ni (Nickel)	58
K (Potassium)	39
Rb (Rubidium)	85
Se (Selenium)	80
Ag (Silver)	107
Na (Sodium)	23
Sr (Strontium)	88
Tl (Thallium)	205
U (Uranium)	238
V (Vanadium)	51
Zn (Zinc)	64

Calibration Standard 3

ICP-MS-CAL3-1 100 mL
10 µg/mL each in 10% HCl 10 comps.

Element	Isotope
Sb (Antimony)	121
Au (Gold)	197
Hf (Hafnium)	180
Ir (Iridium)	193
Pd (Palladium)	106
Pt (Platinum)	195
Rh (Rhodium)	103
Ru (Ruthenium)	102
Te (Tellurium)	130
Sn (Tin)	120

Calibration Standard 5

ICP-MS-CAL5-1 100 mL
10 µg/mL in 5% HNO₃

Element	Isotope
Hg (Mercury)	202

Calibration Standard 4

ICP-MS-CAL4-1 ▼ 100 mL
10 µg/mL each in H₂O tr. HF 12 comps.

Element	Isotope
B (Boron)	11
Ge (Germanium)	74
Mo (Molybdenum)	98
Nb (Niobium)	93
P (Phosphorous)	31
Re (Rhenium)	187
Si (Silicon)	28
S (Sulphur)	32
Ta (Tantalum)	181
Ti (Titanium)	48
W (Tungsten)	184
Zr (Zirconium)	90

Calibration Standard Set

ICP-MS-CAL-1-SET 5 x 100 mL

ICP-MS-CAL1-1	ICP-MS-CAL4-1
ICP-MS-CAL2-1	ICP-MS-CAL5-1
ICP-MS-CAL3-1	



Matrix Blanks

Nitric Acid Blank

ICP-MS-BLN-1 100 mL
ICP-MS-BLN-5 500 mL

5% HNO₃ in 18 Megohm ASTM Type I deionized Water

These blanks are prepared from the same water source and acids as your standards and therefore provide a consistent matrix. They are excellent as a blank, preparing a standard curve, or as a diluent for standards and samples.

Hydrochloric Acid Blank

ICP-MS-BLH-1 100 mL
ICP-MS-BLH-5 500 mL

5% HCl in 18 Megohm ASTM Type I deionized Water

Water Blank

ICP-MS-BLW-1 ▼ 100 mL
ICP-MS-BLW-5 ▼ 500 mL

18 Megohm ASTM Type I deionized Water

▼ Hazardous fee not required.



ICP/MS

Multi-Element Standards

Tuning Solutions

We offer two tuning solutions, both range from 7-238 mass units. Choose the one which best suits your needs.

ICP-MS-TUNSOL1-1 100 mL
100 µg/mL each in 2% HNO₃ 8 comps.

Element	Most Abundant Isotope
Ba (Barium)	138
Be (Beryllium)	9
Cu (Copper)	63
In (Indium)	115
Li (Lithium)	7
Mg (Magnesium)	24
Tl (Thallium)	205
U (Uranium)	238

ICP-MS-TUNSOL2-1 100 mL
100 µg/mL each in 2% HNO₃ 13 comps.

Element	Most Abundant Isotope
Ba (Barium)	138
Be (Beryllium)	9
Bi (Bismuth)	209
Ce (Cerium)	140
Cu (Copper)	63
Ho (Holmium)	165
In (Indium)	115
Pb (Lead)	208
Li (Lithium)	7
Mg (Magnesium)	24
Tl (Thallium)	205
U (Uranium)	238
Y (Yttrium)	89

Interference Check Standards

Solution A

ICP-MS-INTA-1 100 mL
At stated conc. (µg/mL) in 1% HNO₃ 12 comps.

Element	µg/mL	Most Abundant Isotope
Al (Aluminum)	1000	27
C (Carbon)	2000	12
Ca (Calcium)	3000	40
Cl (Chloride)	18000	35
Fe (Iron)	2500	56
Mg (Magnesium)	1000	24
Mo (Molybdenum)	20	98
P (Phosphorous)	1000	31
K (Potassium)	1000	39
Na (Sodium)	2500	23
S (Sulphur)	1000	32
Ti (Titanium)	20	48

Solution B

ICP-MS-INTB-1 100 mL
At stated conc. (µg/mL) in 2% HNO₃ 11 comps.

Element	µg/mL	Most Abundant Isotope
As (Arsenic)	10	75
Cd (Cadmium)	10	114
C (Carbon)	20	12
Cr (Chromium)	20	52
Cu (Copper)	20	63
Mn (Manganese)	20	55
Ni (Nickel)	20	58
Se (Selenium)	10	80
Ag (Silver)	20	107
V (Vanadium)	20	51
Zn (Zinc)	10	64

Interference Check Standard Set

ICP-MS-INT-1-SET 2 x 100 mL

ICP-MS-INTA-1 ICP-MS-INTB-1

Memory Check Solution

Memory Check Solution Sets

ICP-MS-MEMCHKA-R1-SET

2 x 100 mL

ICP-MS-MEMCHKA1-R1
ICP-MS-MEMCHKA2-R1

ICP-MS-MEMCHK-R1-SET

3 x 100 mL

ICP-MS-MEMCHKA1-R1
ICP-MS-MEMCHKA2-R1
ICP-MS-MEMCHKB-R1

Solution A

ICP-MS-MEMCHKA1-R1 100 mL
At stated conc. (µg/mL) in 2% HNO₃ 24 comps.

Element	µg/mL	Most Abundant Isotope
Al (Aluminum)	1000	27
Sb (Antimony)	20	121
As (Arsenic)	20	75
Ba (Barium)	20	138
Be (Beryllium)	20	9
Cd (Cadmium)	20	114
Ca (Calcium)	1000	40
C (Carbon)	2000	12
Cr (Chromium)	20	52
Co (Cobalt)	20	59
Cu (Copper)	20	63
Fe (Iron)	1000	56
Pb (Lead)	20	208
Mg (Magnesium)	1000	24
Mo (Molybdenum)	20	98
K (Potassium)	1000	39
Ti (Titanium)	20	48
Mn (Manganese)	20	55
Ni (Nickel)	20	58
Se (Selenium)	20	80
Na (Sodium)	1000	23
Tl (Thallium)	20	205
V (Vanadium)	20	51
Zn (Zinc)	20	64

ICP-MS-MEMCHKA2-R1 100 mL
20 µg/mL In 2% HNO₃

Element	Most Abundant Isotope
Ag (Silver)	107

Solution B

ICP-MS-MEMCHKB-R1 ▼ 100 mL
At stated conc. (µg/mL) in H₂O 3 comps.

Element	µg/mL	Most Abundant Isotope
Cl (Chloride)	7200	35
P (Phosphorous)	1000	31
S (Sulphur)	1000	32

Technical Note

These memory check solutions are not designed to be used as standards. The solutions should be mixed together right before aspiration. Precipitate will form over time - this is normal and will not affect the performance of the solution. The mixture is used only to determine the memory or "carry-over" that occurs after running a "concentrated" solution.

▼ Hazardous fee not required.



ICP/MS

Multi-Element Standards



Spiking Standards

Spiking Standard for Water

ICP-MS-SPIKE-W-1 100 mL

At stated conc. (µg/mL) in 5% HNO₃ 17 comps.

Element	µg/mL	Isotope
Sb (Antimony)	100	121
As (Arsenic)	50	75
Ba (Barium)	250	138
Be (Beryllium)	25	9
Cd (Cadmium)	25	114
Cr (Chromium)	100	52
Co (Cobalt)	100	59
Cu (Copper)	100	63
Fe (Iron)	500	56
Pb (Lead)	50	208
Mn (Manganese)	100	55
Ni (Nickel)	100	58
Se (Selenium)	25	80
Ag (Silver)	25	107
Tl (Thallium)	25	205
V (Vanadium)	100	51
Zn (Zinc)	250	64

Spiking Standard for Soil

ICP-MS-SPIKE-S-1 100 mL

At stated conc. (µg/mL) in 5% HNO₃ 15 comps.

Element	µg/mL	Isotope
Sb (Antimony)	100	121
As (Arsenic)	50	75
Ba (Barium)	250	138
Be (Beryllium)	25	9
Cd (Cadmium)	50	114
Cr (Chromium)	250	52
Co (Cobalt)	100	59
Cu (Copper)	250	63
Pb (Lead)	100	208
Ni (Nickel)	125	58
Se (Selenium)	25	80
Ag (Silver)	25	107
Tl (Thallium)	25	205
V (Vanadium)	150	51
Zn (Zinc)	250	90

Spiking Standard Set

ICP-MS-SPIKE-1-SET 2 x 100 mL

ICP-MS-SPIKE-W-1 ICP-MS-SPIKE-S-1

Quality Control

Sample 1

ICP-MS-QC1-1

10 µg/mL each in 2% HNO₃

100 mL

9 comps.

Element	Isotope
Be (Beryllium)	9
Bi (Bismuth)	209
Ce (Cerium)	140
Co (Cobalt)	59
In (Indium)	115
Pb (Lead)	208
Mg (Magnesium)	24
Ni (Nickel)	58
U (Uranium)	238

Sample 2

ICP-MS-QC2-1

10 µg/mL each in 5% HNO₃

100 mL

25 comps

Element	Isotope
Al (Aluminum)	27
Sb (Antimony)	121
As (Arsenic)	75
Ba (Barium)	138
Be (Beryllium)	9
Cd (Cadmium)	114
Ca (Calcium)	40
Cr (Chromium)	52
Co (Cobalt)	59
Cu (Copper)	63
Fe (Iron)	56
Pb (Lead)	208
Mg (Magnesium)	24
Mn (Manganese)	55
Mo (Molybdenum)	98
Ni (Nickel)	58
K (Potassium)	39
Se (Selenium)	80
Ag (Silver)	107
Na (Sodium)	23
Tl (Thallium)	205
Th (Thorium)	232
U (Uranium)	238
V (Vanadium)	51
Zn (Zinc)	64

Sample 3

ICP-MS-QC3-1

10 µg/mL each in 5% HNO₃ tr. HF

100 mL

21 comps.

Element	Isotope
Sb (Antimony)	121
As (Arsenic)	75
Be (Beryllium)	9
Cd (Cadmium)	114
Ca (Calcium)	40
Cr (Chromium)	52
Co (Cobalt)	59
Cu (Copper)	63
Fe (Iron)	56
Pb (Lead)	208
Li (Lithium)	7
Mg (Magnesium)	24
Mn (Manganese)	55
Mo (Molybdenum)	98
Ni (Nickel)	58
Se (Selenium)	80
Sr (Strontium)	88
Tl (Thallium)	205
Ti (Titanium)	48
V (Vanadium)	51
Zn (Zinc)	64

Internal Standards

Single Internal Standards

For convenience we offer two concentrations - You decide which is most convenient.

Element	Matrix	Unit	10 µg/mL	100 µg/mL
Bismuth	5% HNO ₃	100 mL	ICP-MS-IS-BI-1	ICP-MS-IS-BI-1-10X
Indium	5% HNO ₃	100 mL	ICP-MS-IS-IN-1	ICP-MS-IS-IN-1-10X
Rhodium	10% HCl	100 mL	ICP-MS-IS-RH-1	ICP-MS-IS-RH-1-10X
Scandium	5% HNO ₃	100 mL	ICP-MS-IS-SC-1	ICP-MS-IS-SC-1-10X
Terbium	5% HNO ₃	100 mL	ICP-MS-IS-TB-1	ICP-MS-IS-TB-1-10X
Yttrium	5% HNO ₃	100 mL	ICP-MS-IS-Y-1	ICP-MS-IS-Y-1-10X

Internal Standard Mix

These internal standards have been chosen because they all have nearly 100% abundance of a single isotope and they are not commonly found in routine samples.

ICP-MS-IS-MIX1-1 100 mL
10 µg/mL each in 2% HNO₃ 7 comps.

Element	Isotope
Bi (Bismuth)	209
Ho (Holmium)	165
In (Indium)	115
6-Li (Lithium-6)	6
Sc (Scandium)	45
Tb (Terbium)	159
Y (Yttrium)	89



ICP/MS

EPA Method 200.8 & 6020

Method 200.8 Determination of Trace Elements in Water and Waste by ICP/MS

Calibration Standards

Calibration Standard #1 (1991 Version)

ICP-MS-200.8-CAL1-1

10 µg/mL each in 5% HNO₃ tr. HF 100 mL
18 comps.

Most Abundant	
Element	Isotope
Al (Aluminum)	27
Sb (Antimony)	121
As (Arsenic)	75
Be (Beryllium)	9
Cd (Cadmium)	114
Cr (Chromium)	52
Co (Cobalt)	59
Cu (Copper)	63
Pb (Lead)	208
Mn (Manganese)	55
Mo (Molybdenum)	98
Ni (Nickel)	58
Se (Selenium)	80
Tl (Thallium)	205
Th (Thorium)	232
U (Uranium)	238
V (Vanadium)	51
Zn (Zinc)	64

Calibration Standard #2

ICP-MS-200.8-CAL2-1

10 µg/mL each in 2% HNO₃ tr. HF 100 mL
2 comps.

Most Abundant	
Element	Isotope
Ba (Barium)	138
Ag (Silver)	67

Calibration Standard #1R

(1994 Version)

ICP-MS-200.8-CAL1R-1

At stated conc. (µg/mL) in 2% HNO₃ tr. HF 100 mL
18 comps.

Most Abundant	
Element	Isotope
Al (Aluminum)	27
Sb (Antimony)	121
As (Arsenic)	75
Be (Beryllium)	9
Cd (Cadmium)	114
Cr (Chromium)	52
Co (Cobalt)	59
Cu (Copper)	63
Pb (Lead)	208
Mn (Manganese)	55
Mo (Molybdenum)	98
Ni (Nickel)	58
Se (Selenium)	80
Tl (Thallium)	205
Th (Thorium)	232
U (Uranium)	238
V (Vanadium)	51
Zn (Zinc)	64

Calibration Standard #3

ICP-MS-200.8-CAL3-1

1 component in 5% HNO₃ 100 mL

Most Abundant	
Element	Isotope
Hg (Mercury)	202

Internal Standards

Internal Standard #1

ICP-MS-200.8-IS-1

100 µg/mL each in 2% HNO₃ 100 mL
5 comps.

Most Abundant	
Element	Isotope
Sc (Scandium)	45
Y (Yttrium)	89
In (Indium)	115
Tb (Terbium)	159
Bi (Bismuth)	209

Internal Standard #2

ICP-MS-200.8-IS2-1

100 µg/mL in 2% HNO₃ 100 mL

Most Abundant	
Element	Isotope
Au (Gold)	197

Tuning Standard

ICP-MS-200.8-TUN-1

10 µg/mL each in 2% HNO₃ 100 mL
5 comps.

Most Abundant	
Element	Isotope
Be (Beryllium)	75
Mg (Magnesium)	24
Co (Cobalt)	59
In (Indium)	115
Pb (Lead)	208

Method 6020 Standards for Inductively Coupled Mass Spectrometry

Calibration Standard

ICP-MS-6020-CAL-R-1

10 µg/mL each in 2% HNO₃ 100 mL
22 comps.

Most Abundant	
Element	Isotope
Al (Aluminum)	27
Sb (Antimony)	121
As (Arsenic)	75
Ba (Barium)	138
Be (Beryllium)	9
Cd (Cadmium)	114
Ca (Calcium)	40
Cr (Chromium)	52
Co (Cobalt)	59
Cu (Copper)	63
Fe (Iron)	56
Pb (Lead)	208
Mg (Magnesium)	24
Mn (Manganese)	55
Ni (Nickel)	58
K (Potassium)	39
Se (Selenium)	80
Ag (Silver)	107
Na (Sodium)	23
Tl (Thallium)	205
V (Vanadium)	51
Zn (Zinc)	64

Calibration Standard

ICP-MS-6020-CAL-1

10 µg/mL each in 2% HNO₃ 100 mL
15 comps.

Most Abundant	
Element	Isotope
Al (Aluminum)	27
Sb (Antimony)	121
As (Arsenic)	75
Ba (Barium)	138
Be (Beryllium)	9
Cd (Cadmium)	114
Cr (Chromium)	52
Co (Cobalt)	59
Cu (Copper)	63
Pb (Lead)	208
Mn (Manganese)	55
Ni (Nickel)	58
Ag (Silver)	107
Tl (Thallium)	205
Zn (Zinc)	64

Interference Check Standard #1

ICP-MS-6020-INT1-1

At stated conc. (µg/mL) in 2% HNO₃ 100 mL
12 comps.

Most Abundant	
Element	Isotope
Al (Aluminum)	1000 27
Cl (Chloride)	10000 35
Ca (Calcium)	1000 40
C (Carbon)	2000 12
Fe (Iron)	1000 56
Mg (Magnesium)	1000 24
Mo (Molybdenum)	20 98
P (Phosphorous)	1000 31
K (Potassium)	1000 39
Na (Sodium)	1000 23
S (Sulfur)	1000 32
Ti (Titanium)	20 48

Interference Check Standard #2

ICP-MS-6020-INT2-1

2 µg/mL each in 5% HNO₃ tr. HF 100 mL
9 comps.

Most Abundant	
Element	Isotope
As (Arsenic)	75
Cd (Cadmium)	114
Cr (Chromium)	52
Co (Cobalt)	59
Cu (Copper)	63
Mn (Manganese)	55
Ni (Nickel)	58
Ag (Silver)	107
Zn (Zinc)	64

Tuning Standard

ICP-MS-6020-TUN-1

10 µg/mL each in 2% HNO₃ 100 mL
4 comps.

Most Abundant	
Element	Isotope
Co (Cobalt)	59
In (Indium)	115
Li (Lithium)	7
Tl (Thallium)	205

Organometallic Standards

AA, ICP, DCP & XRF Analysis



These Standards were formulated for the analysis of metals in oils and other organic matrices. These Standards and curves provide a convenient way to analyze for metals (wear metals, additives and contaminants) in lubricating oils, gasolines, residual oils, crude oils, turbine fuels and environmental samples. All standards undergo rigorous quality assurance checks. Major constituents in the final Standard are typically analyzed by both plasma emission and rotrode techniques. Organometallic Standards listed on this page may contain sulfur which can be introduced by possible sulfonate starting materials used to formulate the actual organometallic standard. We developed a Premium Organometallic line for chemists preferring to have organometallic standards with <1 ppm sulfur or phosphorous (see Table of Contents).

- Single & Multi Element Standards
- Formulated from Ultra High Purity Organometallic starting materials & matrices
- Prepared Calibration Curves
- Certificate of Analysis

Single Element Wear Metals

Organometallic ICP/XRF

Element	1000 µg/g in 75 cSt base oil Cat. No. (50 g)	5000 µg/g in 75 cSt base oil Cat. No. (50 g)
Al (Aluminum)	WM-75CST-01	WM-75CST-01-5X
Sb (Antimony)	WM-75CST-02	WM-75CST-02-5X
As (Arsenic)	WM-75CST-03	
Ba (Barium)	WM-75CST-04	WM-75CST-04-5X
Be (Beryllium)	WM-75CST-05	
Bi (Bismuth)	WM-75CST-06	WM-75CST-06-5X
B (Boron)	WM-75CST-07	WM-75CST-07-5X
Cd (Cadmium)	WM-75CST-08	WM-75CST-08-5X
Ca (Calcium)	WM-75CST-09	WM-75CST-09-5X
Cr (Chromium)	WM-75CST-13	WM-75CST-13-5X
Co (Cobalt)	WM-75CST-14	WM-75CST-14-5X
Cu (Copper)	WM-75CST-15	WM-75CST-15-5X
Fe (Iron)	WM-75CST-27	WM-75CST-27-5X
La (Lanthanum)	WM-75CST-28	
Pb (Lead)	WM-75CST-29	WM-75CST-29-5X
Li (Lithium)	WM-75CST-30	WM-75CST-30-5X
Mg (Magnesium)	WM-75CST-32	WM-75CST-32-5X
Mn (Manganese)	WM-75CST-33	WM-75CST-33-5X
Hg (Mercury)	WM-75CST-34	
Mo (Molybdenum)	WM-75CST-35	WM-75CST-35-5X
Ni (Nickel)	WM-75CST-37	WM-75CST-37-5X
P (Phosphorous)	WM-75CST-41	WM-75CST-41-5X
K (Potassium)	WM-75CST-43	WM-75CST-43-5X
Sc (Scandium)	WM-75CST-50	
Se (Selenium)	WM-75CST-51	
Si (Silicon)	WM-75CST-52	WM-75CST-52-5X
Ag (Silver)	WM-75CST-53	WM-75CST-53-5X
Na (Sodium)	WM-75CST-54	WM-75CST-54-5X
Sr (Strontium)	WM-75CST-55	
S (Sulfur)	WM-75CST-56	WM-75CST-56-5X
Tl (Thallium)	WM-75CST-60	
Sn (Tin)	WM-75CST-63	WM-75CST-63-5X
Ti (Titanium)	WM-75CST-64	WM-75CST-64-5X
V (Vanadium)	WM-75CST-67	WM-75CST-67-5X
Y (Yttrium)	WM-75CST-69	WM-75CST-69-5X
Zn (Zinc)	WM-75CST-70	WM-75CST-70-5X
Zn (Zirconium)	WM-75CST-71	WM-75CST-71-5X

Matrix Oil and Stabilizer

75 cSt Oil

MOSOL-75 500 mL

Stabilizer

WM-STAB 1 x 50 g

Technical Note

Used to improve the stability of Organometallic Standards when diluting into solvents such as Kerosene. Add 0.6% by weight.

Metals Additives

Metals Additives Standard

MA-900-100G	100 g
MA-900-200G	200 g
900 µg/g each in Hydrocarbon oil	
MA-1000-100G	100 g
MA-1000-200G	200 g
1000 µg/g each in Hydrocarbon oil	
MA-3000-100G	100 g
MA-3000-200G	200 g
3000 µg/g each in Hydrocarbon oil	
MA-5000-100G	100 g
MA-5000-200G	200 g
5000 µg/g each in Hydrocarbon oil	
	5 comps.

Ba (Barium) P (Phosphorous)
Ca (Calcium) Zn (Zinc)
Mg (Magnesium)

See our BioFuel
Standards in the Petrochemical Section
of this Catalog for Wear Metals in
BioFuels.





Organometallic Standards

AA, ICP, DCP & XRF Analysis

12 Wear Metal Multi-Element

Conc.	Unit	Cat. No.
10 µg/g	100 g	WM-12-1X-100G
	200 g	WM-12-1X-200G
30 µg/g	100 g	WM-12-3X-100G
	200 g	WM-12-3X-200G
50 µg/g	100 g	WM-12-5X-100G
	200 g	WM-12-5X-200G
100 µg/g	100 g	WM-12-10X-100G
	200 g	WM-12-10X-200G
300 µg/g	100 g	WM-12-30X-100G
	200 g	WM-12-30X-200G
500 µg/g	100 g	WM-12-50X-100G
	200 g	WM-12-50X-200G
900 µg/g	100 g	WM-12-90X-100G
	200 g	WM-12-90X-200G
100 gram Set WM-12-100G-SET set of above 7 x 100 g		
200 gram Set WM-12-200G-SET set of above 7 x 200 g		

Each Wear Metal Standard contains 12 elements listed below in hydrocarbon oil at the stated concentration.

Ag (Silver)	Fe (Iron)	Pb (Lead)
Al (Aluminum)	Mg (Magnesium)	Si (Silicon)
Cr (Chromium)	Na (Sodium)	Sn (Tin)
Cu (Copper)	Ni (Nickel)	Ti (Titanium)

21 Wear Metal Multi-Element

Conc.	Unit	Cat. No.
10 µg/g	100 g	WM-21-1X-100G
	200 g	WM-21-1X-200G
30 µg/g	100 g	WM-21-3X-100G
	200 g	WM-21-3X-200G
50 µg/g	100 g	WM-21-5X-100G
	200 g	WM-21-5X-200G
100 µg/g	100 g	WM-21-10X-100G
	200 g	WM-21-10X-200G
300 µg/g	100 g	WM-21-30X-100G
	200 g	WM-21-30X-200G
500 µg/g	100 g	WM-21-50X-100G
	200 g	WM-21-50X-200G
900 µg/g	100 g	WM-21-90X-100G
	200 g	WM-21-90X-200G
100 gram Set WM-21-100G-SET set of above 7 x 100 g		
200 gram Set WM-21-200G-SET set of above 7 x 200 g		

Each Wear Metal Standard contains 21 elements listed below in hydrocarbon oil at the stated concentration.

Ag (Silver)	Cu (Copper)	P (Phosphorus)
Al (Aluminum)	Fe (Iron)	Pb (Lead)
B (Boron)	Mg (Magnesium)	Si (Silicon)
Ba (Barium)	Mn (Manganese)	Sn (Tin)
Ca (Calcium)	Mo (Molybdenum)	Ti (Titanium)
Cd (Cadmium)	Na (Sodium)	V (Vanadium)
Cr (Chromium)	Ni (Nickel)	Zn (Zinc)

22 Wear Metal Multi-Element

Conc.	Unit	Cat. No.
10 µg/g	100 g	WM-22-1X-100G
	200 g	WM-22-1X-200G
30 µg/g	100 g	WM-22-3X-100G
	200 g	WM-22-3X-200G
50 µg/g	100 g	WM-22-5X-100G
	200 g	WM-22-5X-200G
100 µg/g	100 g	WM-22-10X-100G
	200 g	WM-22-10X-200G
300 µg/g	100 g	WM-22-30X-100G
	200 g	WM-22-30X-200G
500 µg/g	100 g	WM-22-50X-100G
	200 g	WM-22-50X-200G
900 µg/g	100 g	WM-22-90X-100G
	200 g	WM-22-90X-200G
100 gram Set WM-22-100G-SET set of above 7 x 100 g		
200 gram Set WM-22-200G-SET set of above 7 x 200 g		

Each Wear Metal Standard contains 22 elements listed below in hydrocarbon oil at the stated concentration.

Ag (Silver)	Fe (Iron)	P (Phosphorus)
Al (Aluminum)	K (Potassium)	Pb (Lead)
B (Boron)	Mg (Magnesium)	Si (Silicon)
Ba (Barium)	Mn (Manganese)	Sn (Tin)
Ca (Calcium)	Mo (Molybdenum)	Ti (Titanium)
Cd (Cadmium)	(Molybdenum)	V (Vanadium)
Cr (Chromium)	Na (Sodium)	Zn (Zinc)
Cu (Copper)	Ni (Nickel)	

23 Wear Metal Multi-Element

Conc.	Unit	Cat. No.
10 µg/g	100 g	WM-23-1X-100G
	200 g	WM-23-1X-200G
30 µg/g	100 g	WM-23-3X-100G
	200 g	WM-23-3X-200G
50 µg/g	100 g	WM-23-5X-100G
	200 g	WM-23-5X-200G
100 µg/g	100 g	WM-23-10X-100G
	200 g	WM-23-10X-200G
300 µg/g	100 g	WM-23-30X-100G
	200 g	WM-23-30X-200G
500 µg/g	100 g	WM-23-50X-100G
	200 g	WM-23-50X-200G
900 µg/g	100 g	WM-23-90X-100G
	200 g	WM-23-90X-200G
100 gram Set WM-23-100G-SET set of above 7 x 100 g		
200 gram Set WM-23-200G-SET set of above 7 x 200 g		

Each Wear Metal Standard contains 23 elements listed below in hydrocarbon oil at the stated concentration.

Ag (Silver)	Fe (Iron)	Pb (Lead)
Al (Aluminum)	K (Potassium)	Sb (Antimony)
B (Boron)	Mg (Magnesium)	Si (Silicon)
Ba (Barium)	Mn (Manganese)	Sn (Tin)
Ca (Calcium)	Mo (Molybdenum)	Ti (Titanium)
Cd (Cadmium)	Na (Sodium)	V (Vanadium)
Cr (Chromium)	Ni (Nickel)	Zn (Zinc)
Cu (Copper)	P (Phosphorus)	



Organometallic Standards

Premium Sulfur-Free



- Highly Concentrated Liquid Organometallic Standards
- Stabilized

- Sulfur/Phosphorous free (< 1 ppm)
- Ready for Use

Organometallic Single Element Stock Standards

Sulfur free Element	@ 1000 µg/g Cat. No.	50 mL	@ 5000 µg/g Cat. No.	50 mL
Al (Aluminum)	WM-NMS-01		WM-NMS-01-5X	
Sb (Antimony)	WM-NMS-02		WM-NMS-02-5X	
As (Arsenic)	WM-NMS-03		WM-NMS-03-5X	
Ba (Barium)	WM-NMS-04		WM-NMS-04-5X	
Be (Beryllium)	WM-NMS-05		WM-NMS-05-5X	
Cd (Cadmium)	WM-NMS-08		WM-NMS-08-5X	
Ca (Calcium)	WM-NMS-09		WM-NMS-09-5X	
Ce (Cerium)	WM-NMS-11		WM-NMS-11-5X	
Cr (Chromium)	WM-NMS-13		WM-NMS-13-5X	
Co (Cobalt)	WM-NMS-14		WM-NMS-14-5X	
Cu (Copper)	WM-NMS-15		WM-NMS-15-5X	
Ga (Gallium)	WM-NMS-20		WM-NMS-20-5X	
Au (Gold)	WM-NMS-22		-----	----
Fe (Iron)	WM-NMS-27		WM-NMS-27-5X	
Pb (Lead)	WM-NMS-29		WM-NMS-29-5X	
Li (Lithium)	WM-NMS-30		WM-NMS-30-5X	
Mg (Magnesium)	WM-NMS-32		WM-NMS-32-5X	
Mn (Manganese)	WM-NMS-33		WM-NMS-33-5X	
Hg (Mercury)	WM-NMS-34		WM-NMS-34-5X	
Mo (Molybdenum)	WM-NMS-35		WM-NMS-35-5X	
Ni (Nickel)	WM-NMS-37		WM-NMS-37-5X	
P (Phosphorous)	WM-NMS-41		WM-NMS-41-5X	
K (Potassium)	WM-NMS-43		WM-NMS-43-5X	
Se (Selenium)	WM-NMS-51		WM-NMS-51-5X	
Si (Silicon)	WM-NMS-52		WM-NMS-52-5X	
Ag (Silver)	WM-NMS-53		WM-NMS-53-5X	
Na (Sodium)	WM-NMS-54		WM-NMS-54-5X	
Sr (Strontium)	WM-NMS-55		WM-NMS-55-5X	
Tl (Thallium)	WM-NMS-60		WM-NMS-60-5X	
Sn (Tin)	WM-NMS-63		WM-NMS-63-5X	
Ti (Titanium)	WM-NMS-64		WM-NMS-64-5X	
V (Vanadium)	WM-NMS-67		WM-NMS-67-5X	
Y (Yttrium)	WM-NMS-69		WM-NMS-69-5X	
Zn (Zinc)	WM-NMS-70		WM-NMS-70-5X	
Zr (Zirconium)	WM-NMS-71		WM-NMS-71-5X	

Stabilization Solutions

The solutions were specifically designed for chelating & solubilizing our line of Sulfur-Free Organometallic Standards. Contact Tech Service for additional information.

Stabilizer Solution A

ASTM-P-0122-0.5	50 mL
ASTM-P-0122-1	100 mL

Stabilizer Solution B

ASTM-P-0123-0.5	50 mL
ASTM-P-0123-1	100 mL

Stabilizer Solution C

ASTM-P-0124-0.5	50 mL
ASTM-P-0124-1	100 mL

Stabilizer Solution D

ASTM-P-0125-0.5	50 mL
ASTM-P-0125-1	100 mL

Premium
Sulfur-Free

Sulfur below
detection limits

Contains
No Metallic
Sulfonates

Organometallic standards do not require a hazardous shipping fee except where noted.

These highly concentrated stabilized solutions can be used for customer specific formulations of additive elements in lubricating oils, iron, nickel and vanadium in residual oil, and wear metals in oils. These organometallic solutions can also be used to prepare single element or multi-element standards for plasma emission (ICP or DCP), rotating disk (rotrode), or atomic absorption spectroscopy (AA). The starting materials are PURE REAGENTS and have been certified against NIST SRM's whenever available. **THEY DO NOT CONTAIN PHOSPHORUS OR SULFUR.**

Organometallic Single Element Concentrates

Element	Conc. (Wt. %)	Cat. No.	25 grams	Cat. No.	50 grams
Al (Aluminum)	3	WM-NMS-01-30X-25G		WM-NMS-01-30X-50G	
Sb (Antimony)	2	WM-NMS-02-20X-25G		WM-NMS-02-20X-50G	
Ba (Barium)	12.5	WM-NMS-04-125X-25G		WM-NMS-04-125X-50G	
Cd (Cadmium)	10	WM-NMS-08-100X-25G		WM-NMS-08-100X-50G	
Ca (Calcium)	5	WM-NMS-09-50X-25G		WM-NMS-09-50X-50G	
Ce (Cerium)	5	WM-NMS-11-50X-25G		WM-NMS-11-50X-50G	
Cr (Chromium)	3.5	WM-NMS-13-35X-25G		WM-NMS-13-35X-50G	
Co (Cobalt)	7.5	WM-NMS-14-75X-25G		WM-NMS-14-75X-50G	
Cu (Copper)	6	WM-NMS-15-60X-25G		WM-NMS-15-60X-50G	
Fe (Iron)	4	WM-NMS-27-40X-25G		WM-NMS-27-40X-50G	
Pb (Lead)	20	WM-NMS-29-200X-25G		WM-NMS-29-200X-50G	
Li (Lithium)	1.5	WM-NMS-30-15X-25G		WM-NMS-30-15X-50G	
Mg (Magnesium)	3	WM-NMS-32-30X-25G		WM-NMS-32-30X-50G	
Mn (Manganese)	6	WM-NMS-33-60X-25G		WM-NMS-33-60X-50G	
Mo (Molybdenum)	5	WM-NMS-35-50X-25G		WM-NMS-35-50X-50G	
Ni (Nickel)	5	WM-NMS-37-50X-25G		WM-NMS-37-50X-50G	
P (Phosphorus)	5	WM-NMS-41-50X-25G		WM-NMS-41-50X-50G	
K (Potassium)	7.5	WM-NMS-43-75X-25G		WM-NMS-43-75X-50G	
Pr (Praseodymium)	3	WM-NMS-44-30X-25G		WM-NMS-44-30X-50G	
Se (Selenium)	3.5	WM-NMS-51-35X-25G		WM-NMS-51-35X-50G	
Si (Silicon)	7.5	WM-NMS-52-75X-25G		WM-NMS-52-75X-50G	
Na (Sodium)	2.5	WM-NMS-54-25X-25G		WM-NMS-54-25X-50G	
Sr (Strontium)	10	WM-NMS-55-100X-25G		WM-NMS-55-100X-50G	
Tl (Thallium)	5	WM-NMS-60-50X-25G		WM-NMS-60-50X-50G	
Sn (Tin)	7.5	WM-NMS-63-75X-25G		WM-NMS-63-75X-50G	
Ti (Titanium)	5	WM-NMS-64-50X-25G		WM-NMS-64-50X-50G	
V (Vanadium)	4	WM-NMS-67-40X-25G		WM-NMS-67-40X-50G	
Y (Yttrium)	2.5	WM-NMS-69-25X-25G		WM-NMS-69-25X-50G	
Zn (Zinc)	6	WM-NMS-70-60X-25G		WM-NMS-70-60X-50G	
Zr (Zirconium)	5	WM-NMS-71-50X-25G		WM-NMS-71-50X-50G	

Technical Note

These standards are **sulfur-free** (below detection limits) **organometallic standards** for X-ray Fluorescence (XRF), plasma emission (ICP or DCP), rotating disk (rotrode), or atomic absorption (AA). They are stabilized for blending together with other analytes to prepare multi-element standards.

Organometallic



Organometallic Standards

Premium Sulfur-Free

11 Wear Metal Multi-Element

Conc.	Unit	Cat. No.
10 µg/g	100 mL	WM-11-NMS-1X-1
30 µg/g	100 mL	WM-11-NMS-3X-1
50 µg/g	100 mL	WM-11-NMS-5X-1
100 µg/g	100 mL	WM-11-NMS-10X-1
300 µg/g	100 mL	WM-11-NMS-30X-1
500 µg/g	100 mL	WM-11-NMS-50X-1
900 µg/g	100 mL	WM-11-NMS-90X-1
100 mL Set		WM-11-NMS-1-SET
set of above 7 x 100 mL		

Each Wear Metal Standard contains 11 elements listed below in hydrocarbon oil at the stated concentration.

Al (Aluminum)	Mg (Magnesium)	Si (Silicon)
Cr (Chromium)	Na (Sodium)	Sn (Tin)
Cu (Copper)	Ni (Nickel)	Ti (Titanium)
Fe (Iron)	Pb (Lead)	

12 Wear Metal Multi-Element

Conc.	Unit	Cat. No.
10 µg/g	100 mL	WM-12-NMS-1X-1
30 µg/g	100 mL	WM-12-NMS-3X-1
50 µg/g	100 mL	WM-12-NMS-5X-1
100 µg/g	100 mL	WM-12-NMS-10X-1
300 µg/g	100 mL	WM-12-NMS-30X-1
500 µg/g	100 mL	WM-12-NMS-50X-1
900 µg/g	100 mL	WM-12-NMS-90X-1
100 mL Set		WM-12-NMS-1-SET
set of above 7 x 100 mL		

Each Wear Metal Standard contains 12 elements listed below in hydrocarbon oil at the stated concentration.

Ag (Silver)	Fe (Iron)	Pb (Lead)
Al (Aluminum)	Mg (Magnesium)	Si (Silicon)
Cr (Chromium)	Na (Sodium)	Sn (Tin)
Cu (Copper)	Ni (Nickel)	Ti (Titanium)

Contains
No Metallic
Sulfonates

Premium
Sulfur-Free

Sulfur below
detection limits

20 Wear Metal Multi-Element

Conc.	Unit	Cat. No.
10 µg/g	100 mL	WM-20-NMS-1X-1
30 µg/g	100 mL	WM-20-NMS-3X-1
50 µg/g	100 mL	WM-20-NMS-5X-1
100 µg/g	100 mL	WM-20-NMS-10X-1
300 µg/g	100 mL	WM-20-NMS-30X-1
500 µg/g	100 mL	WM-20-NMS-50X-1
900 µg/g	100 mL	WM-20-NMS-90X-1
100 mL Set		WM-20-NMS-1-SET
set of above 7 x 100 mL		

Each Wear Metal Standard contains 20 elements listed below in hydrocarbon oil at the stated concentration.

Al (Aluminum)	Fe (Iron)	Pb (Lead)
B (Boron)	Mg (Magnesium)	Si (Silicon)
Ba (Barium)	Mn (Manganese)	Sn (Tin)
Ca (Calcium)	Mo (Molybdenum)	Ti (Titanium)
Cd (Cadmium)	Na (Sodium)	V (Vanadium)
Cr (Chromium)	Ni (Nickel)	Zn (Zinc)
Cu (Copper)	P (Phosphorus)	

21 Wear Metal Multi-Element

Conc.	Unit	Cat. No.
10 µg/g	100 mL	WM-21-NMS-1X-1
30 µg/g	100 mL	WM-21-NMS-3X-1
50 µg/g	100 mL	WM-21-NMS-5X-1
100 µg/g	100 mL	WM-21-NMS-10X-1
300 µg/g	100 mL	WM-21-NMS-30X-1
500 µg/g	100 mL	WM-21-NMS-50X-1
900 µg/g	100 mL	WM-21-NMS-90X-1
100 mL Set		WM-21-NMS-1-SET
set of above 7 x 100 mL		

Each Wear Metal Standard contains 21 elements listed below in hydrocarbon oil at the stated concentration.

Ag (Silver)	Cu (Copper)	P (Phosphorus)
Al (Aluminum)	Fe (Iron)	Pb (Lead)
B (Boron)	Mg (Magnesium)	Si (Silicon)
Ba (Barium)	Mn (Manganese)	Sn (Tin)
Ca (Calcium)	Mo (Molybdenum)	Ti (Titanium)
Cd (Cadmium)	Na (Sodium)	V (Vanadium)
Cr (Chromium)	Ni (Nickel)	Zn (Zinc)

Recommended
Internal Standard

Organometallic (Internal Standard) Sulfur free

	Conc.	Cat. No.	50 mL
Cobalt	1000 µg/g	WM-NMS-14	
	5000 µg/g	WM-NMS-14-5X	



Sulfur and Metals in Oil

Sulfur and Metals in Mineral Oil

ASTM-P-0100-SET				12 x 100 mL 100 mL
Cat. No.	Sulfur (Wt. %)	Nickel (µg/g)	Vanadium (µg/g)	
ASTM-P-0100-01	0.00	0	0	
ASTM-P-0100-02	0.50	10	500	
ASTM-P-0100-03	1.00	100	25	
ASTM-P-0100-04	1.50	80	250	
ASTM-P-0100-05	2.00	40	100	
ASTM-P-0100-06	2.50	5	400	
ASTM-P-0100-07	3.00	60	300	
ASTM-P-0100-08	3.50	0	200	
ASTM-P-0100-09	4.00	100	0	
ASTM-P-0100-10	4.50	50	250	
ASTM-P-0100-11	5.00	20	500	
ASTM-P-0100-12	5.50	100	50	

Sulfur and Metals in Residual Fuel Oil

ASTM-P-0101-SET				12 x 100 mL 100 mL
Cat. No.	Sulfur (Wt. %)	Nickel (µg/g)	Vanadium (µg/g)	
ASTM-P-0101-01	0.00	0	0	
ASTM-P-0101-02	0.50	10	500	
ASTM-P-0101-03	1.00	100	25	
ASTM-P-0101-04	1.50	80	250	
ASTM-P-0101-05	2.00	40	100	
ASTM-P-0101-06	2.50	5	400	
ASTM-P-0101-07	3.00	60	300	
ASTM-P-0101-08	3.50	0	200	
ASTM-P-0101-09	4.00	100	0	
ASTM-P-0101-10	4.50	50	250	
ASTM-P-0101-11	5.00	20	500	
ASTM-P-0101-12	5.50	100	50	

Test Method A - ICP with an Organic Solvent Specimen Solution

Sulfur and Metals in Mineral Oil

ASTM-P-0102-SET				12 x 100 mL 100 mL
Cat. No.	Sulfur (Wt. %)	Iron (µg/g)	Nickel (µg/g)	Vanadium (µg/g)
ASTM-P-0102-01	0.00	0	0	0
ASTM-P-0102-02	0.50	300	10	500
ASTM-P-0102-03	1.00	500	100	25
ASTM-P-0102-04	-----	100	80	250
ASTM-P-0102-05	2.00	200	40	100
ASTM-P-0102-06	2.50	400	5	400
ASTM-P-0102-07	3.00	0	60	300
ASTM-P-0102-08	3.50	500	0	200
ASTM-P-0102-09	-----	100	100	0
ASTM-P-0102-10	4.50	300	50	250
ASTM-P-0102-11	5.00	200	20	500
ASTM-P-0102-12	5.50	50	100	50

Designed for ASTM D5708

Sulfur and Metals in Residual Fuel Oil

ASTM-P-0103-SET				12 x 100 mL 100 mL
Cat. No.	Sulfur (Wt. %)	Iron (µg/g)	Nickel (µg/g)	Vanadium (µg/g)
ASTM-P-0103-01	0.00	0	0	0
ASTM-P-0103-02	0.50	300	10	500
ASTM-P-0103-03	1.00	500	100	25
ASTM-P-0103-04	-----	100	80	250
ASTM-P-0103-05	2.00	200	40	100
ASTM-P-0103-06	2.50	400	5	400
ASTM-P-0103-07	3.00	0	60	300
ASTM-P-0103-08	3.50	500	0	200
ASTM-P-0103-09	-----	100	100	0
ASTM-P-0103-10	4.50	300	50	250
ASTM-P-0103-11	5.00	200	20	500
ASTM-P-0103-12	5.50	50	100	50

Designed for ASTM D5708

Stock Multi-Element in Mineral Oil

D-5708-A-10X 100 mL
100 µg/g in 20 cSt mineral oil 3 comps.
Iron Vanadium
Nickel

After Acid Decomposition of Sample

Working Level Multi-Element Aqueous Standard

D-5708-B-5 500 mL
10 µg/mL each in 2-5% HNO₃ 3 comps.
Iron Vanadium
Nickel

Stock Multi-Element Aqueous Standard

D-5863-95A-10X-1 1 x 100 mL
2-5% HNO₃ 3 comps.
Iron 100 Vanadium 200
Nickel 200

Nitric Acid Blank

CLP-BLN-5 500 mL
CLP-BLN-L-VAP 1 L (2 x 500 mL)
5% HNO₃ in ASTM Type I Water

Stock Multi-Element Aqueous Standard

D-5708-B-10X-1 100 mL
D-5708-B-10X-5 500 mL
100 µg/mL each in 2-5% HNO₃ 3 comps.
Iron Vanadium
Nickel

Stock Multi-Element Standard in Mineral Oil

D-5863-95B-10X 1 x 100 g
At stated conc. (µg/g) in 20 cst Mineral Oil 3 comps.
Sodium 50 Vanadium 150
Nickel 200

ISO/CD 14597 Vanadium and Nickel Standards with Manganese (Internal Standard)

Vanadium Standards - Low Range for ISO/CD 14597 with 0.05% Manganese Internal Standard in Xylene-Mineral Oil

ASTM-P-0104-SET			9 x 100 mL
Cat. No.	Vanadium Conc. (Wt.%)	100 mL	
ASTM-P-0104-01	0.0005		
ASTM-P-0104-02	0.0025		
ASTM-P-0104-03	0.0050		
ASTM-P-0104-04	0.0075		
ASTM-P-0104-05	0.0100		
ASTM-P-0104-06	0.0125		
ASTM-P-0104-07	0.0150		
ASTM-P-0104-08	0.0175		
ASTM-P-0104-09	0.0200		

Vanadium Standards - High Range for ISO/CD 14597 with 0.05% Manganese Internal Standard in Xylene-Mineral Oil

ASTM-P-0105-SET			7 x 100 mL
Cat. No.	Vanadium Conc. (Wt.%)	100 mL	
ASTM-P-0105-01	0.0000		
ASTM-P-0105-02	0.0300		
ASTM-P-0105-03	0.0400		
ASTM-P-0105-04	0.0500		
ASTM-P-0105-05	0.0600		
ASTM-P-0105-06	0.0800		
ASTM-P-0105-07	0.1000		

Nickel Standards for ISO/CD 14597 with 0.05% Manganese Internal Standard in Xylene-Mineral Oil

ASTM-P-0106-SET			7 x 100 mL
Cat. No.	Nickel Conc. (Wt.%)	100 mL	
ASTM-P-0106-01	0.0000		
ASTM-P-0106-02	0.0005		
ASTM-P-0106-03	0.0010		
ASTM-P-0106-04	0.0025		
ASTM-P-0106-05	0.0050		
ASTM-P-0106-06	0.0075		
ASTM-P-0106-07	0.0100		

Internal Standard

ASTM-P-0107-5 500 mL
Manganese @ 0.05 Wt. % in Xylene-Mineral Oil

▲ Hazardous fee required.



Organometallic Standards

AA, ICP, DCP & XRF Analysis

Lubricating Oil Standards

ASTM-P-0108-SET

17 x 100 mL

Designed for ASTM D6481

Cat. No.	Ca	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0108-01	0.600	0.005	0.175	0.060
ASTM-P-0108-02	0.500	0.200	0.050	0.080
ASTM-P-0108-03	0.400	0.150	0.300	0.180
ASTM-P-0108-04	0.260	0.250	0.150	0.120
ASTM-P-0108-05	0.005	0.005	0.450	0.070
ASTM-P-0108-06	0.400	0.025	0.350	0.100
ASTM-P-0108-07	0.300	0.060	0.250	0.120
ASTM-P-0108-08	0.200	0.100	0.450	0.100
ASTM-P-0108-09	0.060	0.080	0.300	0.130
ASTM-P-0108-10	0.060	0.050	0.200	0.050
ASTM-P-0108-11	0.050	0.120	0.100	0.075
ASTM-P-0108-12	0.025	0.150	0.200	0.130
ASTM-P-0108-13	0.005	0.200	0.400	0.150
ASTM-P-0108-14	0.170	0.250	0.550	0.110
ASTM-P-0108-15	0.100	0.100	0.200	0.200
ASTM-P-0108-16	0.010	0.010	0.600	0.250
ASTM-P-0108-17	0.000	0.000	0.000	0.000

ASTM-P-0109-SET

17 x 100 mL

Cat. No.	Ca	Cl	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0109-01	0.600	0.100	0.005	0.175	0.060
ASTM-P-0109-02	0.500	0.000	0.200	0.050	0.080
ASTM-P-0109-03	0.400	0.010	0.150	0.300	0.180
ASTM-P-0109-04	0.260	0.500	0.250	0.150	0.120
ASTM-P-0109-05	0.005	1.000	0.005	0.450	0.070
ASTM-P-0109-06	0.400	0.400	0.025	0.350	0.100
ASTM-P-0109-07	0.300	0.100	0.060	0.250	0.120
ASTM-P-0109-08	0.200	0.010	0.100	0.450	0.100
ASTM-P-0109-09	0.060	0.050	0.080	0.300	0.130
ASTM-P-0109-10	0.060	0.200	0.050	0.200	0.050
ASTM-P-0109-11	0.050	0.500	0.120	0.100	0.075
ASTM-P-0109-12	0.025	0.800	0.150	0.200	0.130
ASTM-P-0109-13	0.005	1.000	0.200	0.400	0.150
ASTM-P-0109-14	0.170	0.600	0.250	0.550	0.110
ASTM-P-0109-15	0.100	0.200	0.100	0.200	0.200
ASTM-P-0109-16	0.010	0.400	0.010	0.600	0.250
ASTM-P-0109-17	0.000	0.000	0.000	0.000	0.000

ASTM-P-0110-SET

17 x 100 mL

Designed for ASTM D4927

Cat. No.	Ba	Ca	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0110-01	0.100	0.600	0.005	0.175	0.060
ASTM-P-0110-02	0.175	0.500	0.200	0.050	0.080
ASTM-P-0110-03	0.000	0.400	0.150	0.300	0.180
ASTM-P-0110-04	0.025	0.260	0.250	0.150	0.120
ASTM-P-0110-05	0.150	0.005	0.005	0.450	0.070
ASTM-P-0110-06	0.000	0.400	0.025	0.350	0.100
ASTM-P-0110-07	0.200	0.300	0.060	0.250	0.120
ASTM-P-0110-08	0.000	0.200	0.100	0.450	0.100
ASTM-P-0110-09	0.100	0.060	0.080	0.300	0.130
ASTM-P-0110-10	0.050	0.060	0.050	0.200	0.050
ASTM-P-0110-11	0.075	0.050	0.120	0.100	0.075
ASTM-P-0110-12	0.010	0.025	0.150	0.200	0.130
ASTM-P-0110-13	0.005	0.005	0.200	0.400	0.150
ASTM-P-0110-14	0.000	0.170	0.250	0.550	0.110
ASTM-P-0110-15	0.000	0.100	0.100	0.200	0.200
ASTM-P-0110-16	0.005	0.010	0.010	0.600	0.250
ASTM-P-0110-17	0.000	0.000	0.000	0.000	0.000

ASTM-P-0111-SET

17 x 100 mL

Designed for ASTM D4927

Cat. No.	Ba	Ca	Cl	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0111-01	0.100	0.600	0.100	0.005	0.175	0.060
ASTM-P-0111-02	0.175	0.500	0.000	0.200	0.050	0.080
ASTM-P-0111-03	0.000	0.400	0.010	0.150	0.300	0.180
ASTM-P-0111-04	0.025	0.260	0.500	0.250	0.150	0.120
ASTM-P-0111-05	0.150	0.005	1.000	0.005	0.450	0.070
ASTM-P-0111-06	0.000	0.400	0.400	0.025	0.350	0.100
ASTM-P-0111-07	0.200	0.300	0.100	0.060	0.250	0.120
ASTM-P-0111-08	0.000	0.200	0.010	0.100	0.450	0.100
ASTM-P-0111-09	0.100	0.060	0.050	0.080	0.300	0.130
ASTM-P-0111-10	0.050	0.060	0.200	0.050	0.200	0.050
ASTM-P-0111-11	0.075	0.050	0.500	0.120	0.100	0.075
ASTM-P-0111-12	0.010	0.025	0.800	0.150	0.200	0.130
ASTM-P-0111-13	0.005	0.005	1.000	0.200	0.400	0.150
ASTM-P-0111-14	0.000	0.170	0.600	0.250	0.550	0.110
ASTM-P-0111-15	0.000	0.100	0.200	0.100	0.200	0.200
ASTM-P-0111-16	0.005	0.010	0.400	0.010	0.600	0.250
ASTM-P-0111-17	0.000	0.000	0.000	0.000	0.000	0.000

ASTM-P-0112-SET

17 x 100 mL

Cat. No.	Ca	Mg	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0112-01	0.600	0.100	0.005	0.175	0.060
ASTM-P-0112-02	0.500	0.150	0.200	0.050	0.080
ASTM-P-0112-03	0.400	0.350	0.150	0.300	0.180
ASTM-P-0112-04	0.260	0.225	0.250	0.150	0.120
ASTM-P-0112-05	0.005	0.450	0.005	0.450	0.070
ASTM-P-0112-06	0.400	0.500	0.025	0.350	0.100
ASTM-P-0112-07	0.300	0.325	0.060	0.250	0.120
ASTM-P-0112-08	0.200	0.250	0.100	0.450	0.100
ASTM-P-0112-09	0.060	0.100	0.080	0.300	0.130
ASTM-P-0112-10	0.060	0.400	0.050	0.200	0.050
ASTM-P-0112-11	0.050	0.300	0.120	0.100	0.075
ASTM-P-0112-12	0.025	0.200	0.150	0.200	0.130
ASTM-P-0112-13	0.005	0.375	0.200	0.400	0.150
ASTM-P-0112-14	0.170	0.175	0.250	0.550	0.110
ASTM-P-0112-15	0.100	0.425	0.100	0.200	0.200
ASTM-P-0112-16	0.010	0.275	0.010	0.600	0.250
ASTM-P-0112-17	0.000	0.000	0.000	0.000	0.000

ASTM-P-0113-SET

17 x 100 mL

Designed for ASTM D4628 & D4927

Cat. No.	Ba	Ca	Mg	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0113-01	0.025	0.600	0.100	0.005	0.175	0.060
ASTM-P-0113-02	0.000	0.500	0.150	0.200	0.050	0.080
ASTM-P-0113-03	0.100	0.400	0.350	0.150	0.300	0.180
ASTM-P-0113-04	0.175	0.260	0.225	0.250	0.150	0.120
ASTM-P-0113-05	0.150	0.005	0.450	0.005	0.450	0.070
ASTM-P-0113-06	0.000	0.400	0.500	0.025	0.350	0.100
ASTM-P-0113-07	0.100	0.300	0.325	0.060	0.250	0.120
ASTM-P-0113-08	0.200	0.200	0.250	0.100	0.450	0.100
ASTM-P-0113-09	0.050	0.060	0.100	0.080	0.300	0.130
ASTM-P-0113-10	0.075	0.060	0.400	0.050	0.200	0.050
ASTM-P-0113-11	0.010	0.050	0.300	0.120	0.100	0.075
ASTM-P-0113-12	0.000	0.025	0.200	0.150	0.200	0.130
ASTM-P-0113-13	0.175	0.005	0.375	0.200	0.400	0.150
ASTM-P-0113-14	0.005	0.170	0.175	0.250	0.550	0.110
ASTM-P-0113-15	0.000	0.100	0.425	0.100	0.200	0.200
ASTM-P-0113-16	0.005	0.010	0.275	0.010	0.600	0.250
ASTM-P-0113-17	0.000	0.000	0.000	0.000	0.000	0.000



Lubricating Oil Standards (Continued)

ASTM-P-0114-SET

17 x 100 mL

EDXRF ASTM Method

Designed for ASTM D6481

Cat. No.	Ca	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0114-01	0.005	0.005	0.050	0.050
ASTM-P-0114-02	0.600	0.000	0.000	0.000
ASTM-P-0114-03	0.000	0.300	0.000	0.000
ASTM-P-0114-04	1.000	0.000	1.000	0.000
ASTM-P-0114-05	0.000	0.000	0.000	0.300
ASTM-P-0114-06	0.005	0.250	0.800	0.300
ASTM-P-0114-07	0.500	0.150	0.500	0.150
ASTM-P-0114-08	0.010	0.200	0.100	0.250
ASTM-P-0114-09	0.050	0.010	0.400	0.075
ASTM-P-0114-10	0.100	0.150	0.200	0.200
ASTM-P-0114-11	0.200	0.200	0.800	0.100
ASTM-P-0114-12	0.400	0.005	0.800	0.300
ASTM-P-0114-13	0.600	0.100	0.500	0.050
ASTM-P-0114-14	0.800	0.010	0.050	0.100
ASTM-P-0114-15	1.000	0.300	1.000	0.150
ASTM-P-0114-16	0.400	0.050	0.600	0.250
ASTM-P-0114-17	0.000	0.000	0.000	0.000

ASTM-P-0115-SET

17 x 100 mL

Cat. No.	Ca	P	S	Zn	Mg
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0115-01	0.005	0.005	0.050	0.050	0.100
ASTM-P-0115-02	0.600	0.000	0.000	0.000	0.150
ASTM-P-0115-03	0.000	0.300	0.000	0.000	0.350
ASTM-P-0115-04	1.000	0.000	1.000	0.000	0.225
ASTM-P-0115-05	0.000	0.000	0.000	0.300	0.450
ASTM-P-0115-06	0.005	0.250	0.800	0.300	0.500
ASTM-P-0115-07	0.500	0.150	0.500	0.150	0.325
ASTM-P-0115-08	0.010	0.200	0.100	0.250	0.250
ASTM-P-0115-09	0.050	0.010	0.400	0.075	0.050
ASTM-P-0115-10	0.100	0.150	0.200	0.200	0.400
ASTM-P-0115-11	0.200	0.200	0.800	0.100	0.300
ASTM-P-0115-12	0.400	0.005	0.800	0.300	0.200
ASTM-P-0115-13	0.600	0.100	0.500	0.050	0.375
ASTM-P-0115-14	0.800	0.010	0.050	0.100	0.175
ASTM-P-0115-15	1.000	0.300	1.000	0.150	0.425
ASTM-P-0115-16	0.400	0.050	0.600	0.250	0.275
ASTM-P-0115-17	0.000	0.000	0.000	0.000	0.000

ASTM-P-0116-SET

11 x 100 mL

Additives

Designed for ASTM D6481

Cat. No.	Ca	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0116-01	0.500	1.000	0.500	0.500
ASTM-P-0116-02	3.000	1.000	3.000	2.000
ASTM-P-0116-03	2.000	1.250	1.000	1.500
ASTM-P-0116-04	5.000	1.500	0.500	1.200
ASTM-P-0116-05	4.000	0.500	1.500	0.750
ASTM-P-0116-06	2.500	0.750	4.000	1.000
ASTM-P-0116-07	4.000	0.500	2.000	1.250
ASTM-P-0116-08	0.500	2.000	5.000	1.000
ASTM-P-0116-09	1.000	0.750	2.000	1.500
ASTM-P-0116-10	2.500	1.200	3.000	0.500
ASTM-P-0116-11	0.000	0.000	0.000	0.000

ASTM-P-0117-SET

10 x 100 mL

WLXRF ASTM Method

Designed for ASTM D6443

Cat. No.	Ca	Cl	Cu	Mg	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0117-01	0.020	0.030	0.010	0.200	0.250	1.000	0.020
ASTM-P-0117-02	0.020	0.020	0.050	0.200	0.020	0.020	0.250
ASTM-P-0117-03	0.020	0.200	0.010	0.040	0.250	0.150	0.250
ASTM-P-0117-04	0.020	0.200	0.050	0.040	0.020	1.000	0.020
ASTM-P-0117-05	0.400	0.020	0.010	0.040	0.020	1.000	0.250
ASTM-P-0117-06	0.400	0.020	0.050	0.040	0.250	0.020	0.020
ASTM-P-0117-07	0.400	0.200	0.010	0.200	0.020	0.020	0.050
ASTM-P-0117-08	0.400	0.200	0.050	0.200	0.250	1.000	0.250
ASTM-P-0117-09	0.200	0.100	0.025	0.080	0.150	0.500	0.100
ASTM-P-0117-10	0.000	0.000	0.000	0.000	0.000	0.000	0.000

ASTM-P-0118-SET

10 x 100 mL

WLXRF ASTM Method

Designed for ASTM D4628, D4927, D4951, D6443

Cat. No.	Ba	Ca	Cl	Cu	Mg	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0118-01	0.020	0.020	0.030	0.010	0.200	0.250	1.000	0.020
ASTM-P-0118-02	0.250	0.020	0.020	0.050	0.200	0.020	0.020	0.250
ASTM-P-0118-03	0.020	0.020	0.200	0.010	0.040	0.250	0.150	0.250
ASTM-P-0118-04	0.250	0.020	0.200	0.050	0.040	0.020	1.000	0.020
ASTM-P-0118-05	0.020	0.400	0.020	0.010	0.040	0.020	1.000	0.250
ASTM-P-0118-06	0.250	0.400	0.020	0.050	0.040	0.250	0.020	0.020
ASTM-P-0118-07	0.020	0.400	0.200	0.010	0.200	0.020	0.020	0.050
ASTM-P-0118-08	0.250	0.400	0.200	0.050	0.200	0.250	1.000	0.250
ASTM-P-0118-09	0.130	0.200	0.100	0.025	0.080	0.150	0.500	0.100
ASTM-P-0118-10	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000



Organometallic Standards

AA, ICP, DCP & XRF Analysis

Lubricating Oil Standards (Continued)

ASTM-P-0119-SET

22 x 100 mL

Designed for ASTM D4628, D4927, D4951, D6443

Cat. No.	Ca	Cl	Cu	Mg	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0119-01	0.300	0.080	0.030	0.060	0.060	0.275	0.060
ASTM-P-0119-02	0.250	0.100	0.000	0.010	0.150	0.000	0.150
ASTM-P-0119-03	0.500	0.000	0.035	0.160	0.150	0.000	0.020
ASTM-P-0119-04	0.350	0.010	0.000	0.120	0.080	0.200	0.000
ASTM-P-0119-05	0.110	0.000	0.015	0.100	0.100	0.300	0.050
ASTM-P-0119-06	0.200	0.100	0.000	0.200	0.050	0.250	0.150
ASTM-P-0119-07	0.000	0.050	0.025	0.000	0.000	0.450	0.020
ASTM-P-0119-08	0.150	0.030	0.000	0.100	0.030	0.400	0.040
ASTM-P-0119-09	0.250	0.150	0.010	0.160	0.000	0.350	0.080
ASTM-P-0119-10	0.110	0.150	0.040	0.005	0.030	0.750	0.150
ASTM-P-0119-11	0.260	0.050	0.000	0.000	0.000	0.750	0.000
ASTM-P-0119-12	0.200	0.000	0.005	0.140	0.080	0.500	0.080
ASTM-P-0119-13	0.000	0.000	0.005	0.020	0.020	0.200	0.020
ASTM-P-0119-14	0.070	0.150	0.020	0.080	0.140	0.650	0.150
ASTM-P-0119-15	0.050	0.000	0.000	0.000	0.150	0.000	0.000
ASTM-P-0119-16	0.400	0.000	0.001	0.080	0.000	0.500	0.020
ASTM-P-0119-17	0.180	0.020	0.020	0.000	0.020	0.600	0.060
ASTM-P-0119-18	0.400	0.010	0.001	0.010	0.020	0.000	0.000
ASTM-P-0119-19	0.010	0.020	0.040	0.010	0.020	0.200	0.100
ASTM-P-0119-20	0.050	0.005	0.050	0.000	0.008	0.000	0.120
ASTM-P-0119-21	0.200	0.050	0.020	0.080	0.050	0.275	0.050
ASTM-P-0119-22	0.000	0.000	0.000	0.000	0.000	0.000	0.000

Standards of Interest

Concentrations for the sets on pages 401-404 are targets. Actual production lots may vary.

ASTM-P-0120-SET

23 x 100 mL

Cat. No.	Ba	Ca	Cl	Cu	Mg	P	S	Zn
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0120-01	0.100	0.300	0.080	0.030	0.060	0.060	0.275	0.060
ASTM-P-0120-02	0.175	0.250	0.100	0.000	0.010	0.150	0.000	0.150
ASTM-P-0120-03	0.040	0.500	0.000	0.035	0.160	0.150	0.000	0.020
ASTM-P-0120-04	0.020	0.350	0.010	0.000	0.120	0.080	0.200	0.000
ASTM-P-0120-05	0.150	0.110	0.000	0.015	0.100	0.100	0.300	0.050
ASTM-P-0120-06	0.000	0.200	0.100	0.000	0.200	0.050	0.250	0.150
ASTM-P-0120-07	0.200	0.000	0.050	0.025	0.000	0.000	0.450	0.020
ASTM-P-0120-08	0.000	0.150	0.030	0.000	0.100	0.030	0.400	0.040
ASTM-P-0120-09	0.000	0.250	0.150	0.010	0.160	0.000	0.350	0.080
ASTM-P-0120-10	0.000	0.110	0.150	0.040	0.005	0.030	0.750	0.150
ASTM-P-0120-11	0.100	0.260	0.050	0.000	0.000	0.000	0.750	0.000
ASTM-P-0120-12	0.050	0.200	0.000	0.005	0.140	0.080	0.500	0.080
ASTM-P-0120-13	0.000	0.000	0.000	0.005	0.020	0.020	0.200	0.020
ASTM-P-0120-14	0.080	0.070	0.150	0.020	0.080	0.140	0.650	0.150
ASTM-P-0120-15	0.010	0.050	0.000	0.000	0.000	0.150	0.000	0.000
ASTM-P-0120-16	0.000	0.400	0.000	0.001	0.080	0.000	0.500	0.020
ASTM-P-0120-17	0.000	0.180	0.020	0.020	0.000	0.020	0.600	0.060
ASTM-P-0120-18	0.000	0.400	0.010	0.001	0.010	0.020	0.000	0.000
ASTM-P-0120-19	0.150	0.010	0.020	0.040	0.010	0.020	0.200	0.100
ASTM-P-0120-20	0.005	0.050	0.005	0.050	0.000	0.008	0.000	0.120
ASTM-P-0120-21	0.100	0.200	0.050	0.020	0.080	0.050	0.275	0.050
ASTM-P-0120-22	0.120	0.200	0.000	0.000	0.000	0.000	0.750	0.000
ASTM-P-0120-23	0.000	0.000	0.000	0.000	0.000	0.000	0.000	0.000

Metal Working Fluids

ASTM-P-0121-SET

13 x 100 mL

Cat. No.	Cl	P	S
Nominal Value	(Wt.%)	(Wt.%)	(Wt.%)
ASTM-P-0121-01	0.000	0.000	0.000
ASTM-P-0121-02	0.750	0.025	0.500
ASTM-P-0121-03	0.050	0.100	3.000
ASTM-P-0121-04	1.000	0.500	2.500
ASTM-P-0121-05	0.100	0.005	2.000
ASTM-P-0121-06	1.500	0.200	1.000
ASTM-P-0121-07	2.000	0.005	3.000
ASTM-P-0121-08	1.000	0.050	0.100
ASTM-P-0121-09	0.500	0.400	0.000
ASTM-P-0121-10	2.000	0.200	1.500
ASTM-P-0121-11	0.000	0.500	1.500
ASTM-P-0121-12	1.250	0.010	0.050
ASTM-P-0121-13	0.050	0.300	0.050

Stabilization Solutions

The solutions were specifically designed for chelating & solubilizing our line of Sulfur-Free Organometallic Standards. Contact Tech Service for additional information.

Stabilizer Solution A

ASTM-P-0122-0.5

50 mL
100 mL

ASTM-P-0122-1

Stabilizer Solution C

ASTM-P-0124-0.5

50 mL
100 mL

ASTM-P-0124-1

Stabilizer Solution B

ASTM-P-0123-0.5

50 mL
100 mL

ASTM-P-0123-1

Stabilizer Solution D

ASTM-P-0125-0.5

50 mL
100 mL

ASTM-P-0125-1

ASTM D3230 Determination of Salts in Crude Oil

Mixed Salt Solution

D-3230-89-1 ▲ 100 mL
D-3230-89-5 ▲ 500 mL

At stated conc. in Alcohol Solution (1-butanol : MeOH) (ratio 63:37) tr. H₂O 3 comps.

Calcium chloride	10 µg/mL	Sodium chloride	70 µg/mL
Magnesium chloride	20 µg/mL		

ASTM D3237 Lead in Gasoline by AA Spectroscopy

Lead Standard Calibration Curve

D-3237-CAL-SET ▲ 4 x 100 mL

Set includes the following Catalog Numbers:

Description	Cat. No.	100 mL
Blank 1% Aliquat 336/MIBK	D-3237-01 ▲	
0.02 g Pb / gal (5.3 mg Pb/ L) in 1% Aliquat 336 / MIBK	D-3237-02 ▲	
0.05 g Pb / gal (13.2 mg Pb/ L) in 1% Aliquat 336 / MIBK	D-3237-03 ▲	
0.10 g Pb / gal (26.4 mg Pb/ L) in 1% Aliquat 336 / MIBK	D-3237-04 ▲	



ASTM D3605 Trace Metals in Gas Turbine Fuels by AA & Flame Emission & Spectroscopy

Trace Metals Standard

D-3605-91-1 1 x 100 mL

At stated conc. in 1,2,3,4-Tetrahydronaphthalene 4 comps.

Na (Sodium)	(250 µg/mL)	Ca (Calcium)	(250 µg/mL)
Pb (Lead)	(250 µg/mL)	V (Vanadium)	(250 µg/mL)

Standards of Interest

See Table of Contents for a complete listing of Wear Metal Standards.

ASTM D3831 Manganese in Gasoline by AA Spectroscopy

Manganese Stock Solution

D-3831-1 ▲ 1 x 100 mL

Manganese @ 1.0 g Mn / gal (264.2 mg Mn / L) in Methyl isobutyl ketone

▲ Hazardous fee required.



Over 40,000 Standards, just a click away

(Over 10,000 Listed and over 30,000 formulated custom standards)

AccuStandard.com



Solid Waste Analysis

Solid Matrix Standards

AccuStandard offers a broad range of trace element standards in Solid Matrices to meet the requirements for “real world” solid waste reference materials. All Standards are furnished with a Certificate of Analysis, with the mean values, standard deviations, confidence intervals and performance intervals. A list of additional elements/analytes, instructions and other pertinent information is also included. Values listed on the Certificate are generally the result of round robin analyses consisting of testing conducted by an average of 20 independent laboratories.

- Ideal for both AA and ICP Analysis
- Certified by EPA Methods and Protocols
- NATURAL Matrix, Not spiked or fortified
- US EPA SW-846, 3rd Edition Method 3050
- Method 1311
- Lead Abatement Program Material

Metals in Sludges

Metal Concentration (in mg/Kg)	Paint Sludge CRM006-050 50 g	Plating Sludge #1 CRM009-100 120 g	Plating Sludge #2 CRM010-100 100 g	Plating Sludge #3 (POTW) CRM011-100 100 g	Raw Sewage CRM018-050 50 g
Al (Aluminum)	73	(890)	692	(20)	22,436
Sb (Antimony)	----	(9)	----	(10)	ND
As (Arsenic)	----	(20)	----	(20)	6.63
Ba (Barium)	9,970	(50)	175	(5)	1,103
Be (Beryllium)	----	----	----	----	0.30
B (Boron)	(53)	(150)	102	(18,000)	(25.8)
Cd (Cadmium)	32	(1)	----	(4)	5.57
Ca (Calcium)	111	(1,100)	563	(180)	49,068
Cr (Chromium)	11	50	79	59,225	40.1
Co (Cobalt)	----	(7)	----	(10)	3.22
Cu (Copper)	(6)	121,415	63,169	108	840
Fe (Iron)	64	(3,800)	2,699	(4,700)	9,903
Pb (Lead)	753	(14,230)	119,344	269	126
Mg (Magnesium)	47	(150)	(80)	(50)	4,302
Mn (Manganese)	(1)	(40)	17	(30)	200
Hg (Mercury)	----	(1)	(2)	(10)	4.78
Mo (Molybdenum)	----	(20)	(32)	----	10.5
Ni (Nickel)	(2)	343	194	41,975	20.4
P (Phosphorus)	----	----	----	----	----
K (Potassium)	8,710	(640)	----	(49,000)	2,657
Se (Selenium)	(3)	----	----	(4)	8.38
Si (Silicon)	----	----	----	----	(609)
Ag (Silver)	----	9	56	(1)	72.1
Na (Sodium)	91	(18,000)	1,576	(23,000)	1,002
Sr (Strontium)	(90)	(30)	----	(1)	420
Tl (Thallium)	(40)	(30)	----	(20)	ND
Sn (Tin)	(100)	(38,000)	----	(120)	----
Ti (Titanium)	(2)	(70)	----	(2)	----
V (Vanadium)	----	(1)	----	(20)	39.2
Zn (Zinc)	737,431	(40)	183	----	1,121

All values expressed in mg/Kg, parts per million (ppm) unless otherwise noted.

TCLP Metals - Method 1311

Metal Concentration (in mg/Kg)	Soil from Superfund Site CRM202-225 225 g	Soil from Superfund Site CRM204-225 225 g	Ash from Incinerator CRM205-225 225 g	Soil from Superfund Site CRM206-225 225 g	Soil from Superfund Site CRM207-225 225 g	TCLP Metals in Soil CRM208-225 225 g
As (Arsenic)	1.44	0.82	89.5	13.9	9.51	3.93
Ba (Barium)	5.85	(0.04)	0.44	0.38	0.40	32.8
Cd (Cadmium)	19.6	14.4	144	8.3	7.45	46.7
Cr (Chromium)	11.1	4.46	12.7	0.13	1.36	0.87
Pb (Lead)	48.5	10.7	155	2.16	2.76	2.14
Hg (Mercury)	5.58	(< 0.0025)	ND	0.65	0.02	0.62
Se (Selenium)	1.38	Trace	ND	20.5	20.8	ND
Ag (Silver)	5.01	(< 0.1)	ND	1.04	0.99	ND

NOTES:

- These values are as extracted into Fluid #1.
- All values expressed in mg/L, parts per million (ppm) unless otherwise noted.
- Values in parenthesis are not certified and are given for information only.
- ND is not detected.

Solid waste standards do not require a hazardous shipping fee.

Solid Waste Analysis

Solid Matrix Standards



Metals in Soil

Metals Conc. (in mg/Kg) Metal	Sewage Sludge Amended Soil CRM005-050 50g	Soil/ Sediment #4 CRM008-050 50g	Dry Soil #2 Certified pH 2.96 CRM020-050 ** 50g	Dry Soil #3 CRM021-100 100g	Dry soil #5 plus Cyanide CRM022-020 20g	Metals in Soil CRM028-050 50g
Al (Aluminum)	15,333	23,906	2,755	2,725	10,060	7,562
Sb (Antimony)	----	(3)	8.38	4,955	(<0.2)	----
As (Arsenic)	7	14	400	25	5	3.83
Ba (Barium)	853	54	24.8	586	109	73.2
Be (Beryllium)	1	1	----	----	0.5	0.38
B (Boron)	----	(27)	----	----	(16)	ND
Cd (Cadmium)	14	(1)	15.4	1	3	0.50
Ca (Calcium)	119,477	2,935	25,584	5,426	27,242	5,883
Cr (Chromium)	41	48	13.6	11	19	19.0
Co (Cobalt)	6	11	4.51	(3)	6	4.30
Cu (Copper)	465	36	729	4,792	12	8.51
Fe (Iron)	12,650	33,042	191,706	6,481	13,555	10,000
Pb (Lead)	89	95	5,111	(144,742)	415	10.4
Mg (Magnesium)	6,706	6,742	2,687	(2,367)	9,524	2,995
Mn (Manganese)	172	261	945	174	318	209
Hg (Mercury)	3	1	1.12	5	(0.02)	----
Mo (Molybdenum)	14	(2)	----	----	(<1)	----
Ni (Nickel)	26	26	16.9	13	16	13.4
P (Phosphorus)	(10,071)	----	----	----	----	----
K (Potassium)	6,229	3,948	(857)	1,006	3,170	2,045
Se (Selenium)	20	(1)	6.57	----	(0.3)	ND
Si (Silicon)	----	(471)	----	----	(80)	----
Ag (Silver)	36	(1)	38.5	7	(<0.5)	----
Na (Sodium)	2,488	8,706	(79.2)	380	268	3.31
Sr (Strontium)	----	42	(24.7)	----	(54)	ND
Tl (Thallium)	(2)	(<1)	5.91	(<1)	(<0.2)	----
Sn (Tin)	----	----	----	(304)	----	----
Ti (Titanium)	----	----	----	----	----	----
V (Vanadium)	109	44	6.47	(8.7)	23	19.2
Zn (Zinc)	625	134	3,011	546	46	75.0
pH	----	----	2.96	----	----	----
CN	----	----	----	----	26.6	----

Continued (in mg/Kg) Metal	Sediment, + CN CRM015-050 50g	Sediment CRM016-050 50g	Soil CRM023-050 50g	Soil CRM025-050 50g	Soil CRM026-050 50g	Soil CRM027-050 50g	Soil CRM024-050 50g
Al (Aluminum)	9,200	8,920	8,472	7,637	17,730	9,149	8,681
Sb (Antimony)	(ND)	(ND)	----	(<3.2)	(<3.2)	----	(1.17)
As (Arsenic)	6.6	6.48	380	339	5.64	12.5	3.42
Ba (Barium)	83.0	79.3	75.5	1,839	214	167	79.6
Be (Beryllium)	0.47	0.49	0.4	0.33	18.0	----	0.43
B (Boron)	(8.6)	(13.0)	(11.2)	(17.2)	(25.4)	2.74	7.22
Cd (Cadmium)	(ND)	0.47	0.9	369	11.7	11.9	2.15
Ca (Calcium)	23,463	22,646	5,425	28,320	6,221	5,963	5,534
Cr (Chromium)	14.3	14.5	31.1	441	27.2	27.0	25.4
Co (Cobalt)	6.04	5.96	8.9	4.07	6.77	4.77	----
Cu (Copper)	16.1	15.5	4.7	7.76	18.8	9.79	8.70
Fe (Iron)	17,070	16,831	10,678	9,439	21,906	11,299	10,196
Pb (Lead)	15.04	14.1	213	1,447	25.6	51.7	15.7
Mg (Magnesium)	13,611	13,246	3,064	4,376	2,837	2,786	2,945
Mn (Manganese)	183.4	180.0	206	173	633	258	199
Hg (Mercury)	0.10	0.11	77.8	99.8	2.42	3.85	0.71
Mo (Molybdenum)	1.16	(0.97)	----	----	----	----	0.58
Ni (Nickel)	17.5	16.7	11.0	12.2	14.4	10.7	15.0
P (Phosphorus)	----	----	----	----	----	----	----
K (Potassium)	2,074	1,958	2,231	1,992	3,600	2,115	2,102
Se (Selenium)	(1.0)	(1.0)	116	518	(1.86)	13.8	(0.54)
Si (Silicon)	(491)	(347)	(352)	(171)	(166)	387	(404)
Ag (Silver)	(ND)	(0.7)	132	(0.57)	5.90	----	13.3
Na (Sodium)	400	292.0	295	313	199	241	287
Sr (Strontium)	(62)	(61)	(32)	(408)	38.4	42.7	35.4
Tl (Thallium)	(ND)	(4.6)	111	(<4.8)	(<4.8)	----	(13.6)
Sn (Tin)	----	----	----	----	----	----	----
Ti (Titanium)	----	----	----	----	----	----	----
V (Vanadium)	22.1	22.5	21.7	19.3	32.0	21.8	20.8
Zn (Zinc)	69.9	69.7	93.8	51.8	140	51.5	37.3
pH	----	----	----	----	----	----	----
CN	6.04	----	----	----	----	----	----

Solid Matrix Standards continued on next page



Solid Waste Analysis

Solid Matrix Standards

Metals in Ashes

Metals Concentration (in mg/Kg) Metal	Ash #1 CRM001-100 100 g	Ash #2 CRM012-100 100 g	Ash #3 CRM019-050 100 g	Ash #4 CRM205-225 225 g
Al (Aluminum)	----	2,160	(2, 759)	----
Sb (Antimony)	----	----	(223)	----
As (Arsenic)	----	----	78	90
Ba (Barium)	428	19	342	0.4
Be (Beryllium)	----	----	(2)	----
B (Boron)	----	----	(336)	----
Cd (Cadmium)	----	362	432	147
Ca (Calcium)	----	2,111	(51,949)	----
Cr (Chromium)	30	161,520	55	13
Co (Cobalt)	----	(22)	(26)	----
Cu (Copper)	40	3,015	279	----
Fe (Iron)	(16,300)	28,664	(12,693)	----
Pb (Lead)	----	120	4,414	155
Mg (Magnesium)	----	1,506	6,307	----
Mn (Manganese)	(300)	202	(480)	----
Hg (Mercury)	----	----	(2)	ND
Mo (Molybdenum)	----	----	(26)	----
Ni (Nickel)	20	13,279	22	----
P (Phosphorus)	----	----	----	----
K (Potassium)	----	73,324	(49,344)	----
Se (Selenium)	----	----	2	ND
Ag (Silver)	----	55	6	ND
Na (Sodium)	----	29,200	(50,522)	----
Sr (Strontium)	(1,016)	(10)	(173)	----
Tl (Thallium)	----	----	(42)	----
Sn (Tin)	----	(2,055)	(410)	----
Ti (Titanium)	(465)	(20)	(2,870)	----
V (Vanadium)	----	(52)	29	----
Zn (Zinc)	----	635	22,217	----

NOTES:

1. All values expressed in mg/Kg, parts per million (ppm) unless otherwise noted.
2. Values in parenthesis are not certified and are given for information only.
3. ND is not detected.

** Values certified by analysis from 260 independent laboratories

Lead Reference Materials

- Natural matrix, not spiked or fortified
- Samples tested & suitable for lead abatement program

Paint Sludge

CRM006-050	50 g
Certified lead value	753 mg/kg

Paint Waste

CRM013-050	50 g
Certified lead value	643 mg/kg

Dust

CRM014-050	50 g
Certified lead value	1,914 mg/kg

Metals in Particulates & Water Treatment Media

Metals Conc. (in mg/Kg) Metal	Activated Charcoal Filter CRM002-100 100 g	Diatomaceous Earth Filter CRM004-100 100 g	Paint Chips CRM013-050 50 g	Baghouse Dust CRM014-050 50 g
Al (Aluminum)	(1,800)	(29,000)	(1,200)	(6,000)
Sb (Antimony)	(2)	(5)	(20)	(20)
As (Arsenic)	(30)	----	(1)	(2)
Ba (Barium)	(80)	1,595	(1,200)	(1,930)
Be (Beryllium)	----	----	----	----
B (Boron)	(80)	(1,160)	12	(50)
Cd (Cadmium)	(1)	2	38	510
Ca (Calcium)	(980)	(28,000)	(1,500)	(3,900)
Cr (Chromium)	36,339	21	618	2,228
Co (Cobalt)	(10)	(140)	(10)	(10)
Cu (Copper)	96,935	(20)	(30)	(110)
Fe (Iron)	(1,150)	----	(1,400)	(5,860)
Pb (Lead)	(5)	11,871	643	1,914
Mg (Magnesium)	(190)	(17,000)	(430)	(1,100)
Mn (Manganese)	(8)	(90)	(120)	(110)
Hg (Mercury)	(5)	(3)	----	(2)
Mo (Molybdenum)	----	(4)	(1)	(30)
Ni (Nickel)	(30)	(140)	(10)	(10)
P (Phosphorus)	----	----	----	----
K (Potassium)	(490)	----	(170)	(340)
Se (Selenium)	(4)	(2,000)	----	(1)
Ag (Silver)	18	(1)	----	(10)
Na (Sodium)	(480)	(4,200)	(310)	(780)
Sr (Strontium)	(110)	(80)	(950)	(1,900)
Tl (Thallium)	(20)	(30)	(4)	(10)
Sn (Tin)	(120)	(580)	(80)	(140)
Ti (Titanium)	(210)	(100)	(70)	(130)
V (Vanadium)	(40)	(28)	(1)	(2)
Zn (Zinc)	----	(8,500)	(2,370)	(3,150)

NOTES:

1. All values expressed in mg/Kg, parts per million (ppm) unless otherwise noted
2. Values in parenthesis are not certified and are given for information only.
3. ND is not detected.

(20 x 20 inch laminated poster)

Periodic Table of Elements



Leader in Analytical Chemical Reference Standards
Tel. 203-786-5290 • www.AccuStandard.com

Technical Reference

Technical Reference

Solvent Miscibility Table, Density and Boiling Point

Acetic acid	(1.049 g/mL)	(117-118°C)
Acetone	(0.791 g/mL)	(56°C)
Acetonitrile (AcCN)	(0.786 g/mL)	(81-82°C)
Benzene	(0.874 g/mL)	(80°C)
2-Butanol	(0.808 g/mL)	(98°C)
Butyl alcohol	(0.81 g/mL)	(116-118°C)
tert-Butylmethyl ether (MtBE)	(0.74 g/mL)	(55-56°C)
Carbon tetrachloride	(1.594 g/mL)	(76-77°C)
Chloroform	(1.492 g/mL)	(60.5-61.5°C)
Cyclohexane	(0.779 g/mL)	(80.7°C)
Cyclopentane	(0.751 g/mL)	(50°C)
Dichloroethane	(1.256 g/mL)	(83°C)
N,N-Dimethylformamide (DMF)	(0.944 g/mL)	(153°C)
1,4-Dioxane	(1.034 g/mL)	(100-102°C)
Dipropyl ether	(0.736 g/mL)	(88-90°C)
Ethyl acetate (EtOAc)	(0.902 g/mL)	(76.5-77.5°C)
Ethyl alcohol (EtOH)	(0.789 g/mL)	(78°C)
Ethyl ether	(0.706 g/mL)	(34.6°C)
n-Heptane	(0.684 g/mL)	(98°C)
n-Hexane	(0.659 g/mL)	(69°C)
Isooctane	(0.692 g/mL)	(98-99°C)
Isopropyl alcohol	(0.785 g/mL)	(82°C)
Methanol (MeOH)	(0.791 g/mL)	(64.7°C)
Methylene chloride (CH ₂ Cl ₂)	(1.325 g/mL)	(39.8-40°C)
Methyl sulfoxide (DMSO)	(1.10 g/mL)	(189°C)
n-Pentane	(0.626 g/mL)	(35-36°C)
1,1,2,2-Tetrachloroethane	(1.586 g/mL)	(147°C)
Tetrahydrofuran (THF)	(0.889 g/mL)	(65-67°C)
Toluene	(0.865 g/mL)	(110-111°C)
Trichloroethane	(1.336 g/mL)	(74-76°C)
Water	(1 g/mL)	(100°C)
Xylene	(0.868 g/mL)	(138-139°C)

☐ Miscible
☒ Immiscible

Read down column
and across for solvent
miscibility

Density@25°C
Boiling Point

Abbreviations

AA	Atomic Absorption	ICP	Inductively Coupled Plasma
ACS	American Chemical Society	IR	Infrared
ANSI	American National Standards Institute	KF	Karl Fischer
AOAC	Association of Official Analytical Chemists	meq	Milliequivalent
APHA	American Public Health Association	NF	National Formulary
ASTM	American Society for Testing Materials	NIST	National Institute of Standards and Technology
BSI	British Standards Institute	OSHA	Occupational Safety and Health Administration
CAS	Chemical Abstracts Service	RFA	Renewable Fuel Association
CI	Color Index	SPE	Solid Phase Extraction
EPA	U.S. Environmental Protection Agency	TLC	Thin Layer Chromatography
GLC	Gas Liquid Chromatography	USDA	U.S. Department of Agriculture
GR	Guaranteed Reagent	USP	U.S. Pharmacopoeia
Instrumentation for Analysis			
GC	Gas Chromatography		
GC/MS	Gas Chromatography/Mass Spectroscopy		
GC/FID	Gas Chromatography/Flame Ionization Detector		
GC/PID	Gas Chromatography/Photoionization Detector		
GC/ECD	Gas Chromatography/Electron Capture Detector		
GC/ELCD	Gas Chromatography/Electrolytic Conductivity Detector		
GC/NPD	Gas Chromatography/Nitrogen phosphorus Detector		
HPLC	High Performance Liquid Chromatography		
NPD/TEA	Nitrogen-Phosphorus Detector/Thermal Energy Analyzer with the reductive Hall detector		
TCD/FID	Thermal Conductivity Detector/Flame Ionization Detector		
NPD/AFD/TSD	Nitrogen-Phosphorus Detector/Alkali-Flame Detector/Thermionic Specific Detector		
NPA/ELCD/FPD	Nitrogen-Phosphorus Detector/Electrolytic Conductivity Detector/Flame Photometric Detector		
HRGC/LRMS	High Resolution Gas Chromatography/Low Resolution Mass Spectrometry		
FTIR	Fourier Transform Infrared		

General Constants

Universal Gas Constant	R = 0.0821 (Atm)(L)/(K)(mole)
Acceleration Due to Gravity	g = 32.17 ft/sec ² , 9.8 m/sec ²
Avogadro's Constant	N = 6.023 x 10 ²³ molecules/mole
Speed of Light	c = 186,282 miles/sec, 3 x 10 ¹⁰ cm/sec
Heat of Fusion (water 1 atm, 0°C)	ΔH _{fus} = 79.7 cal/g
Heat of Vaporization (water 1 atm, 100°C)	ΔH _{vap} = 540 cal/g
Volume of Perfect Gas	22.4 L/mol at 0°C, 760 torr
Faraday's Constant	F = 96485.31 C/mol
Planck's Constant	h = 6.626 x 10 ⁻²⁷ erg/sec
Bragg's Law	2d nλ = 2d sin(θ)
Absolute Zero	-273.15 °C
Temp. of Liquid Nitrogen	-195.8 °C

Unit Conversions

Temperature		
°Fahrenheit	°Celsius	(°F - 32) x 5/9
°Celsius	°Fahrenheit	(°C x 9/5) + 32
°Celsius	°Kelvin	°C + 273
Mass		
1 ounce	=	28.4 grams
0.03527 grams	=	1 ounces
Liquid Volume		Multiply by
Ounces (US)	Milliliters	29.57
Pints (US)	Liters	0.47
Quarts (US)	Liters	0.95
Gallons (US)	Liters	3.8
Milliliters	Ounces (US)	0.034
Liters	Pints (US)	2.1
Liters	Quarts (US)	1.06
Liters	Gallons (US)	0.26
Cubic Feet	Cubic meters	0.03
Cubic Yards	Cubic meters	0.76

1 ppm = 1 µg/mL (wt. /vol.) and
 1 ppm = 1 µg/g (wt./wt.)
 1% = 10,000 ppm (parts per million)
 1 ppm = 1,000 ppb (parts per billion)
 1 ppb = 1,000 ppt (parts per trillion)

1 g = 1,000 mg (milligram)
 1 mg = 1,000 µg (microgram)
 1 µg = 1,000 ng (nanogram)

1 L = 1,000 mL (milliLiter)
 1 mL = 1,000 µL (microLiter)
 1 µL = 1,000 nL (nanoLiter)